

LB121105


245.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

37189



A= 0.0000e+000

B= 2.6899e-004

C= -4.3380e-002

Rho= 0.9998092

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.005	-0.005	142	0.000	142					
0.05	0.050	0.047	-0.003	335	0.0 %	335	0				-6
0.2	0.200	0.188	-0.012	859	0.0 %	859					-6
2.5	2.500	2.433	-0.067	9207	0.0 %	9207					-3
5.0	5.000	5.170	0.170	19380	0.0 %	19380					3
7.5	7.500	7.458	-0.042	27886	0.0 %	27886					1
10.0	10.000	9.960	-0.040	37189	0.0 %	37189					0

MB
7/15/22

LB121105 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 15 Jul 2022 15:43:34

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	142	-	0.0000	245.1	PPB	15 Jul 2022 15:45:24	S	Std
0.05 - 1	50.05	335	-	0.0500	245.1	PPB	15 Jul 2022 15:47:41	S	Std
0.2 - 1	50.2	858	-	0.2000	245.1	PPB	15 Jul 2022 15:49:58	S	Std
2.5 - 1	52.5	9207	-	2.5000	245.1	PPB	15 Jul 2022 15:52:14	S	Std
5.0 - 1	55	19380	-	5.0000	245.1	PPB	15 Jul 2022 15:54:31	S	Std
7.5 - 1	57.5	27886	-	7.5000	245.1	PPB	15 Jul 2022 15:59:24	S	Std
10.0 - 1	59	37189	-	10.0000	245.1	PPB	15 Jul 2022 16:14:29	S	Std
ICV43 - 1	ICV43	15131	4.0267	-245.1		PPB	15 Jul 2022 16:17:10	U	SMPL
ICB43 - 1	ICB43	92	-0.0186	-245.1		PPB	15 Jul 2022 16:19:25	U	SMPL
CCV47 - 1	CCV47	19225	5.1280	-245.1		PPB	15 Jul 2022 16:21:40	U	SMPL
CCB47 - 1	CCB47	72	-0.0240	-245.1		PPB	15 Jul 2022 16:23:54	U	SMPL
CRA - 1	CRA	330	0.0454	-245.1		PPB	15 Jul 2022 16:26:10	U	SMPL
HighStd - 1	HighStd	37170	9.9550	-245.1		PPB	15 Jul 2022 16:28:25	U	SMPL
ChkStd - 1	ChkStd	26294	7.0295	-245.1		PPB	15 Jul 2022 16:30:40	U	SMPL
PB146263BL - 1	PBW	14	-0.0396	-245.1		PPB	15 Jul 2022 16:32:58	U	SMPL
PB146263BS - 1	LCSW	15624	4.1593	-245.1		PPB	15 Jul 2022 16:35:16	U	SMPL
N3214-10 - 1	LOD-MDL-WATER-04-QT2-2022	309	0.0397	-245.1		PPB	15 Jul 2022 16:37:32	U	SMPL
N3214-10RE - 1	LOD-MDL-WATER-04-QT2-2022RE	714	0.1487	-245.1		PPB	15 Jul 2022 16:39:49	U	SMPL
N3214-12 - 1	MDL-WATER-06-QT2-2022	326	0.0443	-245.1		PPB	15 Jul 2022 16:42:05	U	SMPL
CCV48 - 1	CCV48	19543	5.2135	-245.1		PPB	15 Jul 2022 16:44:22	U	SMPL
CCB48 - 1	CCB48	86	-0.0202	-245.1		PPB	15 Jul 2022 16:46:38	U	SMPL