



Analytical Summary Report

Analysis Method: 9045D

Parameter: Corrosivity

Run Number: LB121108

BalanceID: WC SC-6

Analyst By : JIGNESH

Supervisor Review By : ketankumar

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	07/15/2022	15:00
2	CAL2	1	Water	NA	NA	20.8	7.00	07/15/2022	15:01
3	CAL3	1	Water	NA	NA	20.8	10.03	07/15/2022	15:02
4	ICV	1	Water	NA	NA	20.7	6.98	07/15/2022	15:05
5	CCV1	1	Water	NA	NA	20.7	2.01	07/15/2022	15:10
6	N3730-02	1	Solid	20.02	20	22.7	8.72	07/15/2022	15:15
7	N3730-04	1	Solid	20.01	20	23.3	8.86	07/15/2022	15:16
8	N3730-06	1	Solid	20.03	20	22.5	8.66	07/15/2022	15:20
9	N3730-08	1	Solid	20.04	20	21.8	6.52	07/15/2022	15:22
10	N3731-02	1	Solid	20.03	20	23.3	8.91	07/15/2022	15:26
11	N3731-04	1	Solid	20.04	20	23.7	8.83	07/15/2022	15:30
12	N3731-04DUP	1	Solid	20.03	20	23.8	8.84	07/15/2022	15:32
13	CCV2	1	Water	NA	NA	20.8	12.03	07/15/2022	15:35

WORKLIST(Hardcopy Internal Chain)

132-1108

WorkList Name : corrosivity n3731 WorkList ID : 161413 Department : Wet-Chemistry Date : 07-15-2022 10:39:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/13/2022	Solid	N3730-02	Corrosivity	Cool 4 deg C	LIRO01	G11	SB-01-COMP	07/13/2022	9045D
07/13/2022	Solid	N3730-04	Corrosivity	Cool 4 deg C	LIRO01	G11	SB-02-COMP	07/13/2022	9045D
07/13/2022	Solid	N3730-06	Corrosivity	Cool 4 deg C	LIRO01	G11	SB-03-COMP	07/13/2022	9045D
07/13/2022	Solid	N3730-08	Corrosivity	Cool 4 deg C	LIRO01	G11	SB-04-COMP	07/13/2022	9045D
07/14/2022	Solid	N3731-02	Corrosivity	Cool 4 deg C	PSEG03	G11	TP-1	07/14/2022	9045D
07/14/2022	Solid	N3731-04	Corrosivity	Cool 4 deg C	PSEG03	G11	TP-2	07/14/2022	9045D

Date/Time 07/15/22 12:30
 Raw Sample Received by: JDC/SM
 Raw Sample Relinquished by: JDC/SM

Date/Time 07/15/22 16:30
 Raw Sample Received by: JDC/SM
 Raw Sample Relinquished by: JDC/SM