



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB121125

Analyst By : JIGNESH

Supervisor Review By : mohammad

Slope : 99.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.9	4.01	07/19/2022	08:11
2	CAL2	1	Water	NA	NA	20.9	7.01	07/19/2022	08:12
3	CAL3	1	Water	NA	NA	20.9	10.03	07/19/2022	08:13
4	ICV	1	Water	NA	NA	20.9	6.98	07/19/2022	08:15
5	CCV1	1	Water	NA	NA	20.8	2.01	07/19/2022	08:20
6	N3760-01	1	Water	NA	NA	25.8	5.87	07/19/2022	08:25
7	N3760-02	1	Water	NA	NA	25.7	6.03	07/19/2022	08:26
8	N3760-04	1	Water	NA	NA	26.0	8.66	07/19/2022	08:30
9	N3760-05	1	Water	NA	NA	25.9	5.58	07/19/2022	08:33
10	N3760-07	1	Water	NA	NA	26.4	7.96	07/19/2022	08:40
11	N3763-02	1	Water	NA	NA	26.3	6.70	07/19/2022	08:43
12	N3763-02DUP	1	Water	NA	NA	26.4	6.71	07/19/2022	08:44
13	CCV2	1	Water	NA	NA	20.9	12.02	07/19/2022	08:45

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 3760 w WorkList ID : 161459 Department : Wet-Chemistry Date : 07-19-2022 08:03:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/18/2022	Water	N3760-01	pH C	Cool 4 deg C	PSEG03	G12	AUD-1821	07/18/2022	9040C
07/18/2022	Water	N3760-02	pH C	Cool 4 deg C	PSEG03	G12	AUD-1822	07/18/2022	9040C
07/18/2022	Water	N3760-04	pH A	Cool 4 deg C	PSEG03	G12	AUD-1640	07/18/2022	9040C
07/18/2022	Water	N3760-05	pH A	Cool 4 deg C	PSEG03	G12	AUD-1645	07/18/2022	9040C
07/18/2022	Water	N3760-07	pH A	Cool 4 deg C	PSEG03	G12	AUD-1651	07/18/2022	9040C
07/18/2022	Water	N3763-02	pH C	Cool 4 deg C	PSEG03	G12	TRE-1279-WATER	07/18/2022	9040C

Date/Time 07-19-22 08:10
 Raw Sample Received by: JN WCI
 Raw Sample Relinquished by: CP SAN

Date/Time 07-19-22 10:00
 Raw Sample Received by: CP SAN
 Raw Sample Relinquished by: JN WCI