

LB121126


7470A

INSTRUMENT ID: CV2

Linear

 μ Abs.:

37764



A= 0.0000e+000

B= 2.5783e-004

C= 5.4024e-002

Rho= 0.9984306

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.057	0.057	10	0.000	10					-
0.2	0.200	0.189	-0.011	524	0.0 %	524					-6
2.5	2.500	2.211	-0.289	8365	0.0 %	8365					-12
5.0	5.000	5.334	0.334	20477	0.0 %	20477					7
7.5	7.500	7.619	0.119	29341	0.0 %	29341					2
10.0	10.000	9.791	-0.209	37764	0.0 %	37764					-2

LB121126 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Jul 2022 10:34:26

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	10	-	0.0000	7470A	PPB	19 Jul 2022 10:44:42	S	Std
0.2 - 1	50.2	524	-	0.2000	7470A	PPB	19 Jul 2022 10:47:00	S	Std
2.5 - 1	52.5	8365	-	2.5000	7470A	PPB	19 Jul 2022 10:49:17	S	Std
5.0 - 1	55	20477	-	5.0000	7470A	PPB	19 Jul 2022 10:51:33	S	Std
7.5 - 1	57.5	29341	-	7.5000	7470A	PPB	19 Jul 2022 10:56:20	S	Std
10.0 - 1	510	37764	-	10.0000	7470A	PPB	19 Jul 2022 11:03:53	S	Std
ICV44 - 1	ICV44	15715	4.1058	-7470A	PPB	19 Jul 2022 11:06:37	U	SMPL	
ICB44 - 1	ICB44	-42	0.0432	-7470A	PPB	19 Jul 2022 11:08:52	U	SMPL	
CCV49 - 1	CCV49	20174	5.2555	-7470A	PPB	19 Jul 2022 11:11:07	U	SMPL	
CCB49 - 1	CCB49	-43	0.0429	-7470A	PPB	19 Jul 2022 11:13:22	U	SMPL	
CRA - 1	CRA	530	0.1907	-7470A	PPB	19 Jul 2022 11:15:36	U	SMPL	
HighStd - 1	HighStd	37916	9.8300	-7470A	PPB	19 Jul 2022 11:17:51	U	SMPL	
ChkStd - 1	ChkStd	27830	7.2295	-7470A	PPB	19 Jul 2022 11:20:06	U	SMPL	
PB146278BL - 1	PBW	-35	0.0450	-7470A	PPB	19 Jul 2022 11:22:23	U	SMPL	
PB146278BS - 1	LCSW	14951	3.9089	-7470A	PPB	19 Jul 2022 11:24:40	U	SMPL	
N3730-02 - 1	SB-01-COMP	11	0.0569	-7470A	PPB	19 Jul 2022 11:26:56	U	SMPL	
N3730-04 - 1	SB-02-COMP	53	0.0677	-7470A	PPB	19 Jul 2022 11:29:12	U	SMPL	
N3730-06 - 1	SB-03-COMP	70	0.0721	-7470A	PPB	19 Jul 2022 11:31:29	U	SMPL	
N3730-08 - 1	SB-04-COMP	21	0.0594	-7470A	PPB	19 Jul 2022 11:33:46	U	SMPL	
N3731-02 - 1	TP-1	88	0.0767	-7470A	PPB	19 Jul 2022 11:36:01	U	SMPL	
CCV50 - 1	CCV50	20253	5.2759	-7470A	PPB	19 Jul 2022 11:38:16	U	SMPL	
CCB50 - 1	CCB50	-48	0.0416	-7470A	PPB	19 Jul 2022 11:40:31	U	SMPL	
N3731-04 - 1	TP-2	60	0.0695	-7470A	PPB	19 Jul 2022 11:42:46	U	SMPL	
N3731-04DUP - 1	TP-2DUP	99	0.0795	-7470A	PPB	19 Jul 2022 11:45:02	U	SMPL	
N3731-04MS - 1	TP-2MS	15411	4.0275	-7470A	PPB	19 Jul 2022 11:47:17	U	SMPL	
N3731-04MSD - 1	TP-2MSD	14582	3.8137	-7470A	PPB	19 Jul 2022 11:49:33	U	SMPL	
PB146253TB - 1	PB146253TB	6	0.0556	-7470A	PPB	19 Jul 2022 11:51:49	U	SMPL	
N3731-04LX5 - 1		-49	0.0414	-7470A	PPB	19 Jul 2022 11:54:05	U	SMPL	
N3731-04A - 1		14027	3.6706	-7470A	PPB	19 Jul 2022 11:56:22	U	SMPL	
CCV51 - 1	CCV51	17821	4.6488	-7470A	PPB	19 Jul 2022 11:58:39	U	SMPL	
CCB51 - 1	CCB51	-57	0.0393	-7470A	PPB	19 Jul 2022 12:00:55	U	SMPL	