



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Guerda

Parameter: Flash Point

Supervisor Review By: mohammad

Run Number: LB121131

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID:

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	07/19/2022	15:00
2	N3760-03		1	12	100.00	>212.0	>212.0	07/19/2022	15:40
3	N3760-03DUP		1	17	100.00	>212.0	>212.0	07/19/2022	16:02
4	N3760-04		1	18	100.00	>212.0	>212.0	07/19/2022	16:22
5	N3760-05		1	17	100.00	>212.0	>212.0	07/19/2022	16:48
6	N3760-07		1	17	100.00	>212.0	>212.0	07/19/2022	17:05
7	N3763-02		1	18	100.00	>212.0	>212.0	07/19/2022	17:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB121131

WorkList Name : FLASHPOINT-07-18-22 WorkList ID : 161457 Department : Wet-Chemistry Date : 07-18-2022 15:52:45

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/21/2022	Water	N3760-03	Flash Point	Cool 4 deg C	PSEG03	G12	AUD-1821-1822	07/18/2022	1010B
07/21/2022	Water	N3760-04	Flash Point	Cool 4 deg C	PSEG03	G12	AUD-1640	07/18/2022	1010B
07/21/2022	Water	N3760-05	Flash Point	Cool 4 deg C	PSEG03	G12	AUD-1645	07/18/2022	1010B
07/21/2022	Water	N3760-07	Flash Point	Cool 4 deg C	PSEG03	G12	AUD-1651	07/18/2022	1010B
07/21/2022	Water	N3763-02	Flash Point	Cool 4 deg C	PSEG03	G12	TRE-1279-WATER	07/18/2022	1010B

Date/Time 7/19/2022 1430
 Raw Sample Received by: AK (wsc)
 Raw Sample Relinquished by: KS (wsc)

Date/Time 7/19/2022 17:45
 Raw Sample Received by: KS (wsc)
 Raw Sample Relinquished by: AK (wsc)