


LB121132

7471B

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 1.7909e-004

C= 4.1605e-002

Rho= 0.9998539

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.041	0.041	-6	0.000	-6					
0.2	0.200	0.227	0.027	1034	0.0 %	1034					14
2.5	2.500	2.503	0.003	13742	0.0 %	13742					0
5.0	5.000	4.890	-0.110	27070	0.0 %	27070					-2
7.5	7.500	7.455	-0.045	41393	0.0 %	41393					-1
10.0	10.000	10.085	0.085	56081	0.0 %	56081					1

LB121132 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 19 Jul 2022 14:40:01

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-6	-	0.0000	7471B	PPB	19 Jul 2022 15:29:29	S	Std
0.2 - 1	50.2	1034	-	0.2000	7471B	PPB	19 Jul 2022 15:31:46	S	Std
2.5 - 1	52.5	13742	-	2.5000	7471B	PPB	19 Jul 2022 15:34:03	S	Std
5.0 - 1	55	27070	-	5.0000	7471B	PPB	19 Jul 2022 15:36:20	S	Std
7.5 - 1	57.5	41393	-	7.5000	7471B	PPB	19 Jul 2022 15:41:32	S	Std
10.0 - 1	510	56081	-	10.0000	7471B	PPB	19 Jul 2022 15:46:22	S	Std
ICV45 - 1	ICV45	21802	3.9462	-	7471B	PPB	19 Jul 2022 15:49:06	U	SMPL
ICB45 - 1	ICB45	-71	0.0289	-	7471B	PPB	19 Jul 2022 15:51:21	U	SMPL
CCV52 - 1	CCV52	25162	4.5480	-	7471B	PPB	19 Jul 2022 15:53:36	U	SMPL
CCB52 - 1	CCB52	-105	0.0228	-	7471B	PPB	19 Jul 2022 15:55:51	U	SMPL
CRA - 1	CRA	960	0.2135	-	7471B	PPB	19 Jul 2022 15:58:08	U	SMPL
HighStd - 1	HighStd	55925	10.0575	-	7471B	PPB	19 Jul 2022 16:00:23	U	SMPL
ChkStd - 1	ChkStd	38323	6.9050	-	7471B	PPB	19 Jul 2022 16:02:39	U	SMPL
PB146292BL - 1	PBS	-138	0.0169	-	7471B	PPB	19 Jul 2022 16:04:56	U	SMPL
PB146292BS - 1	LCSS	19435	3.5223	-	7471B	PPB	19 Jul 2022 16:07:14	U	SMPL
N3726-01 - 1	HL-12-12.5	376	0.1089	-	7471B	PPB	19 Jul 2022 16:09:30	U	SMPL
N3726-01DUP - 1	HL-12-12.5DUP	429	0.1184	-	7471B	PPB	19 Jul 2022 16:14:42	U	SMPL
N3726-01MS - 1	HL-12-12.5MS	19773	3.5828	-	7471B	PPB	19 Jul 2022 16:16:58	U	SMPL
N3726-01MSD - 1	HL-12-12.5MSD	19731	3.5753	-	7471B	PPB	19 Jul 2022 16:19:15	U	SMPL
N3726-02 - 1	OVS-12-12.5	746	0.1752	-	7471B	PPB	19 Jul 2022 16:21:32	U	SMPL
CCV53 - 1	CCV53	26589	4.8035	-	7471B	PPB	19 Jul 2022 16:23:49	U	SMPL
CCB53 - 1	CCB53	-101	0.0235	-	7471B	PPB	19 Jul 2022 16:26:04	U	SMPL
N3726-03 - 1	AWOM-12-12.5	518	0.1344	-	7471B	PPB	19 Jul 2022 16:28:20	U	SMPL
N3726-04 - 1	SEPTIC-12-12.5	548	0.1397	-	7471B	PPB	19 Jul 2022 16:30:35	U	SMPL
N3746-01 - 1	CL-02-071522	4022	0.7619	-	7471B	PPB	19 Jul 2022 16:32:51	U	SMPL
N3747-01 - 1	OK-01-071522	724	0.1713	-	7471B	PPB	19 Jul 2022 16:35:08	U	SMPL
N3726-01LX5 - 1		8	0.0430	-	7471B	PPB	19 Jul 2022 16:37:23	U	SMPL
N3726-01A - 1		21392	3.8728	-	7471B	PPB	19 Jul 2022 16:39:40	U	SMPL
CCV54 - 1	CCV54	26528	4.7926	-	7471B	PPB	19 Jul 2022 16:41:57	U	SMPL
CCB54 - 1	CCB54	-102	0.0233	-	7471B	PPB	19 Jul 2022 16:44:15	U	SMPL