



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : mohammad

Run Number: LB121135

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	07/20/2022	08:20
2	CAL2	1	Water	NA	NA	20.8	7.00	07/20/2022	08:21
3	CAL3	1	Water	NA	NA	20.8	10.03	07/20/2022	08:22
4	ICV	1	Water	NA	NA	20.9	6.98	07/20/2022	08:25
5	CCV1	1	Water	NA	NA	20.8	2.01	07/20/2022	08:26
6	N3774-01	1	Solid	20.03	20	27.0	6.58	07/20/2022	08:30
7	N3774-01DUP	1	Solid	20.04	20	27.1	6.59	07/20/2022	08:31
8	CCV2	1	Water	NA	NA	20.9	12.03	07/20/2022	08:33

WB 121135

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph s n3778 WorkList ID : 161488 Department : Wet-Chemistry Date : 07-20-2022 08:13:46

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/19/2022	Solid	N3774-01	pH	Cool 4 deg C	PSEG03	G12	34067-69	07/19/2022	9045D

Date/Time 07/20/22 08:15
Raw Sample Received by: J.P. (GOL)
Raw Sample Relinquished by: J.P. (GOL)

Date/Time 07/20/22 11:00
Raw Sample Received by: J.P. (GOL)
Raw Sample Relinquished by: J.P. (GOL)