



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Guerda

Parameter: Flash Point

Supervisor Review By: mohammad

Run Number: LB121143

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID:

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	07/20/2022	09:30
2	N3770-01		1	16	100.00	>212.0	>212.0	07/20/2022	10:20
3	N3770-01DUP		1	17	100.00	>212.0	>212.0	07/20/2022	10:50
4	N3770-03		1	16	100.00	>212.0	>212.0	07/20/2022	11:45
5	N3775-01		1	15	100.00	>212.0	>212.0	07/20/2022	12:20
6	N3775-03		1	11	100.00	>212.0	>212.0	07/20/2022	12:45
7	N3779-01		1	17	100.00	>212.0	>212.0	07/20/2022	01:20
8	N3779-03		1	12	100.00	>212.0	>212.0	07/20/2022	01:45
9	N3779-04		1	13	100.00	>212.0	>212.0	07/20/2022	02:25
10	N3779-05		1	14	100.00	>212.0	>212.0	07/20/2022	02:45
11	N3779-06		1	19	100.00	>212.0	>212.0	07/20/2022	03:15
12	N3779-07		1	15	100.00	>212.0	>212.0	07/20/2022	03:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB121143

Worklist Name : FLASHPOINT-07-20-22

Worklist ID : 161498

Department : Wet-Chemistry

Date : 07-20-2022 09:18:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/26/2022	Water	N3770-01	Flash Point	Cool 4 deg C	PSEG03	G11	33763	07/19/2022	1010B
07/26/2022	Water	N3770-03	Flash Point	Cool 4 deg C	PSEG03	G11	33744	07/19/2022	1010B
07/21/2022	Water	N3775-01	Flash Point	Cool 4 deg C	PSEG03	G12	M-R-1	07/19/2022	1010B
07/21/2022	Water	N3775-03	Flash Point	Cool 4 deg C	PSEG03	G12	M-R-2	07/19/2022	1010B
07/26/2022	Water	N3779-01	Flash Point	Cool 4 deg C	PSEG03	G12	FRAC-TANIK-40811	07/19/2022	1010B
07/26/2022	Water	N3779-03	Flash Point	Cool 4 deg C	PSEG03	G12	070622-A	07/19/2022	1010B
07/26/2022	Water	N3779-04	Flash Point	Cool 4 deg C	PSEG03	G12	070622-B	07/19/2022	1010B
07/26/2022	Water	N3779-05	Flash Point	Cool 4 deg C	PSEG03	G12	070622-C	07/19/2022	1010B
07/26/2022	Water	N3779-06	Flash Point	Cool 4 deg C	PSEG03	G12	070622-D	07/19/2022	1010B
07/26/2022	Water	N3779-07	Flash Point	Cool 4 deg C	PSEG03	G12	070622-E	07/19/2022	1010B

Date/Time 7/20/22 9:20 AM
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: KS (WC)

Date/Time 7/20/22 4:00 PM
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: AM (WC)