

Analytical Summary Report

Analysis Method: 365.3

Parameter: Phosphorus-Total

Run Number: LB121144

ANALYST: Iwona

SUPERVISOR REVIEW BY: JIGNESH

Reagent/Standard	Lot/Log #
calibration std. phosphate 1 ppm	WP98003
calibration std. phosphate 0.5 ppm	WP98002
calibration std. phosphate 0.3 ppm	WP98001
calibration std. phosphate 0.1 ppm	WP98000
calibration std. phosphate 0.05 ppm	WP97999
calibration std. 0 ppm	WP97998
phosphate CCV std.	WP98006
Combined reagent	WP98005
Phenolphthalein indicator	WP96861
Sodium hydroxide, 1N	WP95405
Phosphate ICV-LCS Std	WP98007

Intercept: -0.0018 Slope: 0.6548 Regression: 0.999906

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	%D	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	0.003		07/20/2022	14:10
2	CAL2	0.05	1	50	50	0.032	0.052	-4	07/20/2022	14:10
3	CAL3	0.10	1	50	50	0.066	0.104	-4	07/20/2022	14:11
4	CAL4	0.30	1	50	50	0.188	0.29	3.3	07/20/2022	14:11
5	CAL5	0.50	1	50	50	0.325	0.499	0.2	07/20/2022	14:12
6	CAL6	1.00	1	50	50	0.655	1.003	-0.3	07/20/2022	14:12



Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: JIGNESH

Run Number: LB121144

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.329	0.505	07/20/2022	14:13
2	ICB		1	50	50	0.002	0.006	07/20/2022	14:13
3	CCV1	0.50	1	50	50	0.321	0.493	07/20/2022	14:14
4	CCB1		1	50	50	0.003	0.007	07/20/2022	14:15
5	PB146340BL		1	50	50	0.002	0.006	07/20/2022	14:15
6	PB146340BS	0.50	1	50	50	0.327	0.502	07/20/2022	14:16
7	N3717-05		1	50	50	0.824	1.261	07/20/2022	14:16
8	N3717-05DUP		1	50	50	0.826	1.264	07/20/2022	14:17
9	N3717-05MS	0.50	1	50	50	0.984	1.505	07/20/2022	14:17
10	N3717-05MSD	0.50	1	50	50	0.981	1.501	07/20/2022	14:18
11	N3717-05		5	50	50	0.550	0.843	07/20/2022	14:18
12	N3717-05DUP		5	50	50	0.551	0.844	07/20/2022	14:19
13	N3717-05MS	0.50	5	50	50	0.627	0.960	07/20/2022	14:19
14	N3717-05MSD	0.50	5	50	50	0.626	0.959	07/20/2022	14:20
15	CCV2	0.50	1	50	50	0.326	0.501	07/20/2022	14:20
16	CCB2		1	50	50	0.003	0.007	07/20/2022	14:21