



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB121151

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.9	4.01	07/20/2022	17:30
2	CAL2	1	Water	NA	NA	20.9	7.00	07/20/2022	17:31
3	CAL3	1	Water	NA	NA	20.9	10.03	07/20/2022	17:32
4	ICV	1	Water	NA	NA	20.8	6.99	07/20/2022	17:35
5	CCV1	1	Water	NA	NA	20.8	2.01	07/20/2022	17:36
6	N3806-02	1	Water	NA	NA	27.6	7.09	07/20/2022	17:39
7	N3813-05	1	Water	NA	NA	23.1	7.08	07/20/2022	17:42
8	N3813-05DUP	1	Water	NA	NA	23.2	7.09	07/20/2022	17:43
9	CCV2	1	Water	NA	NA	20.8	12.03	07/20/2022	17:45

✓ 121151

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph n3815 WorkList ID : 161524 Department : Wet-Chemistry Date : 07-20-2022 17:20:54

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample		Collect Date	Method
						Storage	Location		
07/20/2022	Water	N3806-02	pH	Cool 4 deg C	PSEG03		FRAC-TANK-N48964	07/20/2022	9040C
07/20/2022	Water	N3813-05	pH	Cool 4 deg C	PARS02	N11	QUEENS54A-WC-20220720	07/20/2022	9040C

Date/Time 07/20/22 17:22
Raw Sample Received by: 20(wc)
Raw Sample Relinquished by: KS(wc)

Date/Time 07/20/22 18:30
Raw Sample Received by: KS(wc)
Raw Sample Relinquished by: 1P(wc)