



PERCENT SOLID

Supervisor: mohammad

Analyst: JIGNESH

Date: 7/22/2022

OVENTEMP IN Celsius(°C): 107

Time IN: 14:00

In Date: 07/21/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 104

Time OUT: 08:15

Out Date: 07/22/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB121158

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N3790-01	DBUJ6	1	1.00	1.00	2.00	2.00	100.0	P.E.SAMPLE
N3790-02	DBUJ7	2	1.00	1.00	2.00	2.00	100.0	P.E.SAMPLE
N3790-03	DBUJ8	3	1.00	1.00	2.00	2.00	100.0	P.E.SAMPLE
N3790-04	DBUJ9	4	0.91	8.76	9.67	8.59	87.7	
N3790-05	DBUK0	5	0.96	8.81	9.77	8.11	81.2	
N3790-06	DBUK1	6	0.9	8.85	9.75	8.42	85.0	
N3790-07	DBUK3	7	0.92	8.85	9.77	8.89	90.1	
N3790-08	DBUK4	8	0.93	8.82	9.75	9.12	92.9	
N3790-09	DBUK4MS	9	0.93	8.82	9.75	9.12	92.9	
N3790-10	DBUK4MSD	10	0.93	8.82	9.75	9.12	92.9	
N3790-11	DBUK5	11	0.92	8.74	9.66	9.28	95.7	
N3790-12	DBUK6	12	0.91	9.05	9.96	9.33	93.0	
N3790-13	DBUK7	13	0.91	8.74	9.65	8.63	88.3	
N3790-14	DBUK8	14	0.93	8.75	9.68	7.48	74.9	
N3790-15	DBUK9	15	0.95	9.03	9.98	7.48	72.3	
N3790-16	DBUL1	16	0.9	8.71	9.61	7.72	78.3	
N3790-17	DBUL3	17	0.93	9.02	9.95	7.97	78.0	
N3790-18	DBUL4	18	0.91	9.06	9.97	8.63	85.2	
N3790-19	DBUL5	19	0.96	8.73	9.69	7.66	76.7	
N3790-20	DBUL6	20	0.91	8.67	9.58	7.78	79.2	
N3790-21	DBUL7	21	0.95	9.00	9.95	8.00	78.3	
N3790-22	DBUL8	22	0.89	8.86	9.75	8.19	82.4	
N3790-23	DBUM1	23	0.91	9.06	9.97	8.36	82.2	
N3790-24	DBUM3	24	0.91	8.64	9.55	9.54	99.9	
N3790-25	VHBLK001	25	1.00	1.00	2.00	2.00	100.0	VHBLK

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$