



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: JIGNESH

Run Number: LB121162

Reagent/Standard	Lot/Log #
Turbidity Calibration std, 0NTU	WP98046
Turbidity Calibration std, 20NTU	WP98041
Turbidity Calibration std, 40NTU	WP98040
Turbidity Calibration std, 1NTU	WP98044
Turbidity Calibration std, 5NTU	WP98043
Turbidity Calibration - CCV std, 10 NTU	WP98042
10 NTU Standard 500 ml	W2752

Intercept: 0.0512

Slope: 0.9314

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.150	0.11		07/21/2022	09:45
2	CAL2	1	1	1.150	1.18	-18	07/21/2022	09:49
3	CAL3	5	1	4.420	4.69	6.2	07/21/2022	09:52
4	CAL4	10	1	9.470	10.11	-1.1	07/21/2022	09:55
5	CAL5	20	1	18.500	19.81	1	07/21/2022	09:58
6	CAL6	40	1	37.400	40.10	-0.3	07/21/2022	10:02



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: JIGNESH

Run Number: LB121162

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.620	10.274	07/21/2022	10:06
2	ICB		1	0.138	0.093	07/21/2022	10:10
3	CCV1	10	1	9.520	10.166	07/21/2022	10:14
4	CCB1		1	0.148	0.104	07/21/2022	10:18
5	LB121162BL		1	0.146	0.102	07/21/2022	10:22
6	N3734-02		1	4.060	4.304	07/21/2022	10:26
7	N3734-02DUP		1	4.110	4.358	07/21/2022	10:30
8	CCV2	10	1	9.420	10.059	07/21/2022	10:34
9	CCB2		1	0.152	0.108	07/21/2022	10:38