

LB121163


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

53459



A= 0.0000e+000

B= 1.9139e-004

C= -2.0624e-002

Rho= 0.9984858

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.012	0.012	168	0.000	168					7
0.2	0.200	0.215	0.015	1231	0.0 %	1231					2
2.5	2.500	2.545	0.045	13403	0.0 %	13403					3
5.0	5.000	5.142	0.142	26976	0.0 %	26976					-6
7.5	7.500	7.076	-0.424	37077	0.0 %	37077					2
10.0	10.000	10.211	0.211	53459	0.0 %	53459					

LB121163 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 21 Jul 2022 10:10:29					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	168	-	0.0000	7471B	PPB	21 Jul 2022 10:16:53	S	Std
0.2 - 1	50.2	1231	-	0.2000	7471B	PPB	21 Jul 2022 10:19:10	S	Std
2.5 - 1	52.5	13403	-	2.5000	7471B	PPB	21 Jul 2022 10:21:27	S	Std
5.0 - 1	55	26976	-	5.0000	7471B	PPB	21 Jul 2022 10:23:44	S	Std
7.5 - 1	57.5	37077	-	7.5000	7471B	PPB	21 Jul 2022 10:33:37	S	Std
10.0 - 1	510	53459	-	10.0000	7471B	PPB	21 Jul 2022 10:35:53	S	Std
ICV47 - 1	ICV47	21705	4.1335	-	7471B	PPB	21 Jul 2022 10:41:13	U	SMPL
ICB47 - 1	ICB47	-62	-0.0325	-	7471B	PPB	21 Jul 2022 10:43:28	U	SMPL
CCV61 - 1	CCV61	26541	5.0591	-	7471B	PPB	21 Jul 2022 10:45:43	U	SMPL
CCB61 - 1	CCB61	-36	-0.0275	-	7471B	PPB	21 Jul 2022 10:47:58	U	SMPL
CRA - 1	CRA	1168	0.2029	-	7471B	PPB	21 Jul 2022 10:50:14	U	SMPL
HighStd - 1	HighStd	53707	10.2584	-	7471B	PPB	21 Jul 2022 10:52:29	U	SMPL
ChkStd - 1	ChkStd	34562	6.5942	-	7471B	PPB	21 Jul 2022 10:54:44	U	SMPL
PB146358BL - 1	PBS	-33	-0.0269	-	7471B	PPB	21 Jul 2022 10:59:33	U	SMPL
PB146358BS - 1	LCSS	20495	3.9019	-	7471B	PPB	21 Jul 2022 11:01:48	U	SMPL
N3756-01 - 1	TP-1	23189	4.4175	-	7471B	PPB	21 Jul 2022 11:04:04	U	SMPL
N3756-01DUP - 1	TP-1DUP	23172	4.4143	-	7471B	PPB	21 Jul 2022 11:06:22	U	SMPL
N3756-01MS - 1	TP-1MS	46123	8.8069	-	7471B	PPB	21 Jul 2022 11:11:14	U	SMPL
N3756-01MSD - 1	TP-1MSD	45723	8.7304	-	7471B	PPB	21 Jul 2022 11:13:32	U	SMPL
N3760-06 - 1	AUD-1606	349	0.0462	-	7471B	PPB	21 Jul 2022 11:15:49	U	SMPL
CCV62 - 1	CCV62	25726	4.9031	-	7471B	PPB	21 Jul 2022 11:18:06	U	SMPL
CCB62 - 1	CCB62	-67	-0.0334	-	7471B	PPB	21 Jul 2022 11:20:22	U	SMPL
N3763-01 - 1	TRE-1279-SLUDGE	16644	3.1649	-	7471B	PPB	21 Jul 2022 11:22:38	U	SMPL
N3756-01LX5 - 1		4783	0.8948	-	7471B	PPB	21 Jul 2022 11:24:53	U	SMPL
N3756-01A - 1		44847	8.5627	-	7471B	PPB	21 Jul 2022 11:27:09	U	SMPL
CCV63 - 1	CCV63	26263	5.0059	-	7471B	PPB	21 Jul 2022 11:29:24	U	SMPL
CCB63 - 1	CCB63	-78	-0.0356	-	7471B	PPB	21 Jul 2022 11:31:42	U	SMPL