



PERCENT SOLID

Supervisor: mohammad

Analyst: JIGNESH

Date: 7/22/2022

OVENTEMP IN Celsius(°C): 108

Time IN: 17:00

In Date: 07/21/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 104

Time OUT: 08:25

Out Date: 07/22/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB121166

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N3789-01	IDW-VB1-071922	13	0.91	9.05	9.96	8.18	80.3	
N3789-02	IDW-VB1-071922	14	0.91	9.05	9.96	8.18	80.3	
N3811-01	S1-03-B	1	0.9	8.89	9.79	9.34	94.9	
N3811-02	S1-04-B	2	0.92	9.06	9.98	9.48	94.5	
N3811-03	S1-05-B	3	0.93	8.62	9.55	9.14	95.2	
N3812-01	S6-02-52	4	0.93	8.65	9.58	8.95	92.7	
N3812-02	S6-03-52	5	0.91	9.03	9.94	9.56	95.8	
N3814-01	ENV-DE-4 (0-5)	15	0.91	8.95	9.86	8.46	84.4	
N3814-02	ENV-DE-4 (5-10)	16	0.92	9.03	9.95	9.00	89.5	
N3814-03	ENV-DE-4 (10-15)	17	0.92	9.05	9.97	9.55	95.4	
N3814-04	ENV-DE-3 (0-5)	18	0.94	9.00	9.94	8.63	85.4	
N3814-05	ENV-DE-3 (5-10)	19	0.95	8.91	9.86	8.37	83.3	
N3814-06	ENV-DE-3 (10-15)	20	0.91	9.04	9.95	9.24	92.1	
N3814-07	ENV-DE-2 (0-5)	21	0.95	9.02	9.97	8.82	87.3	
N3814-08	ENV-DE-2 (5-10)	22	0.91	9.02	9.93	8.43	83.4	
N3814-09	ENV-DE-2 (10-15)	23	0.98	9.00	9.98	8.77	86.6	
N3814-10	ENV-DE-1 (0-5)	24	0.91	9.00	9.91	8.74	87.0	
N3817-01	S17-01-B	6	0.88	8.75	9.63	8.66	88.9	
N3817-02	S17-02-B	7	0.95	9.01	9.96	8.84	87.6	
N3817-03	S17-03-B	8	0.94	8.83	9.77	8.77	88.7	
N3817-04	DUP-12	9	0.91	9.02	9.93	9.02	89.9	
N3817-05	S17-08-S	10	0.9	8.92	9.82	9.28	93.9	
N3817-06	S17-09-S	11	0.88	8.64	9.52	8.74	91.0	
N3817-07	S17-10-S	12	0.91	9.07	9.98	8.89	88.0	
N3822-11	SITE-1-SS01-20220720	25	0.92	8.61	9.53	9.25	96.7	
N3822-12	SITE-1-SS02-20220720	26	0.91	8.64	9.55	9.32	97.3	
N3822-13	SITE-1-SS03-20220720	27	0.97	9.01	9.98	9.46	94.2	
N3822-14	SITE-1-SS04-20220720	28	0.81	8.74	9.55	9.21	96.1	



PERCENT SOLID

Supervisor: mohammad

Analyst: JIGNESH

Date: 7/22/2022

OVENTEMP IN Celsius(°C): 108

Time IN: 17:00

In Date: 07/21/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 104

Time OUT: 08:25

Out Date: 07/22/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB121166

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N3822-15	SITE-1-SS05-20220720	29	0.98	8.95	9.93	9.66	97.0	
N3822-16	SITE-1-SS06-20220720	30	0.87	8.98	9.85	9.73	98.7	
N3822-17	SITE-1-SS07-20220720	31	0.92	8.73	9.65	9.55	98.9	
N3822-18	SITE-1-SS08-20220720	32	0.93	9.03	9.96	9.72	97.3	
N3830-01	RBR-20-0010	33	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
N3830-03	HS-SOIL-0721	34	0.91	8.70	9.61	9.27	96.1	
N3831-01	VNJ-237	35	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
N3832-01	MMC-DRUMS-0721	36	0.93	9.02	9.95	7.62	74.2	
N3833-01	OR-03-072122	37	0.89	8.94	9.83	9.00	90.7	
N3833-02	OR-03-072122	38	0.89	8.94	9.83	9.00	90.7	
N3834-01	SU-04-072122	39	0.91	9.03	9.94	9.45	94.6	
N3834-02	SU-04-072122	40	0.91	9.03	9.94	9.45	94.6	
N3835-01	MH-191E	41	0.93	9.04	9.97	9.09	90.3	
N3835-02	MH-191E-EPH-2	42	0.93	9.04	9.97	9.09	90.3	
N3837-01	PX-1	43	0.92	9.04	9.96	8.42	83.0	
N3837-02	PX-2	44	0.87	9.06	9.93	7.64	74.7	
N3837-03	PX-3	45	0.91	8.77	9.68	8.81	90.1	
N3837-04	PX-4	46	0.89	8.80	9.69	7.7	77.4	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$