

LB121171


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

56908



A= 0.0000e+000

B= 1.7304e-004

C= -2.4422e-003

Rho= 0.9996279

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.003	-0.003	-2	0.000	-2					
0.2	0.200	0.166	-0.034	973	0.0 %	973					-1.7
2.5	2.500	2.464	-0.036	14256	0.0 %	14256					-1
5.0	5.000	5.048	0.048	29188	0.0 %	29188					1
7.5	7.500	7.679	0.179	44390	0.0 %	44390					2
10.0	10.000	9.845	-0.155	56908	0.0 %	56908					-2

LB121171 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 21 Jul 2022 13:27:58

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-2	-	0.00007471B	PPB		21 Jul 2022 13:30:30	S	Std
0.2 - 1	50.2	973	-	0.20007471B	PPB		21 Jul 2022 13:32:46	S	Std
2.5 - 1	52.5	14256	-	2.50007471B	PPB		21 Jul 2022 13:35:03	S	Std
5.0 - 1	55	29188	-	5.00007471B	PPB		21 Jul 2022 13:37:20	S	Std
7.5 - 1	57.5	44390	-	7.50007471B	PPB		21 Jul 2022 13:42:36	S	Std
10.0 - 1	59	56908	-	10.00007471B	PPB		21 Jul 2022 13:44:51	S	Std
ICV48 - 1	ICV48	22378	3.8699	-7471B	PPB		21 Jul 2022 13:47:52	U	SMPL
ICB48 - 1	ICB48	-71	-0.0147	-7471B	PPB		21 Jul 2022 13:50:07	U	SMPL
CCV64 - 1	CCV64	29168	5.0449	-7471B	PPB		21 Jul 2022 13:52:23	U	SMPL
CCB64 - 1	CCB64	-68	-0.0142	-7471B	PPB		21 Jul 2022 13:54:37	U	SMPL
CRA - 1	CRA	1108	0.1893	-7471B	PPB		21 Jul 2022 13:56:54	U	SMPL
HighStd - 1	HighStd	56820	9.8299	-7471B	PPB		21 Jul 2022 13:59:09	U	SMPL
ChkStd - 1	ChkStd	43649	7.5507	-7471B	PPB		21 Jul 2022 14:04:53	U	SMPL
PB146362BL - 1	PBS	-96	-0.0191	-7471B	PPB		21 Jul 2022 14:07:09	U	SMPL
PB146362BS - 1	LCSS	21611	3.7372	-7471B	PPB		21 Jul 2022 14:09:26	U	SMPL
N3776-01 - 1	72-11966	2104	0.3616	-7471B	PPB		21 Jul 2022 14:11:43	U	SMPL
N3777-01 - 1	TR-06-071922	21309	3.6849	-7471B	PPB		21 Jul 2022 14:14:00	U	SMPL
N3778-01 - 1	BU-03-071922	3651	0.6293	-7471B	PPB		21 Jul 2022 14:16:17	U	SMPL
N3786-01 - 1	0718-WC-TP-12EA	2528	0.4350	-7471B	PPB		21 Jul 2022 14:24:01	U	SMPL
N3786-01DUP - 1	0718-WC-TP-12EADUP	2907	0.5006	-7471B	PPB		21 Jul 2022 14:28:56	U	SMPL
CCV65 - 1	CCV65	29363	5.0786	-7471B	PPB		21 Jul 2022 14:31:11	U	SMPL
CCB65 - 1	CCB65	-83	-0.0168	-7471B	PPB		21 Jul 2022 14:33:27	U	SMPL
N3786-01MS - 1	0718-WC-TP-12EAMS	23160	4.0052	-7471B	PPB		21 Jul 2022 14:35:43	U	SMPL
N3786-01MSD - 1	0718-WC-TP-12EAMSD	23967	4.1449	-7471B	PPB		21 Jul 2022 14:37:59	U	SMPL
N3786-03 - 1	0718-WC-TP-12EB	359	0.0597	-7471B	PPB		21 Jul 2022 14:40:15	U	SMPL
N3786-05 - 1	0718-WC-TP-WC3B	7962	1.3753	-7471B	PPB		21 Jul 2022 14:42:33	U	SMPL
N3786-07 - 1	0718-WC-TP-WC3A	470	0.0789	-7471B	PPB		21 Jul 2022 14:44:49	U	SMPL
N3786-09 - 1	0718-WC-TP-WC3EB	1450	0.2485	-7471B	PPB		21 Jul 2022 14:47:05	U	SMPL
N3786-11 - 1	0718-WC-TP-WC3EA	9670	1.6709	-7471B	PPB		21 Jul 2022 14:49:21	U	SMPL
N3786-13 - 1	0718-WC-TP-WC5A	4309	0.7432	-7471B	PPB		21 Jul 2022 14:51:38	U	SMPL
N3786-15 - 1	0718-WC-TP-WC5B	1105	0.1888	-7471B	PPB		21 Jul 2022 14:53:55	U	SMPL
N3786-17 - 1	0718-WC-TP-WC17A	207	0.0334	-7471B	PPB		21 Jul 2022 14:56:11	U	SMPL
CCV66 - 1	CCV66	28832	4.9867	-7471B	PPB		21 Jul 2022 14:58:26	U	SMPL
CCB66 - 1	CCB66	-74	-0.0152	-7471B	PPB		21 Jul 2022 15:00:42	U	SMPL
N3786-19 - 1	0718-WC-TP-WC17B	1383	0.2369	-7471B	PPB		21 Jul 2022 15:02:59	U	SMPL
N3810-01 - 1	COMP-1	42796	7.4031	-7471B	PPB		21 Jul 2022 15:05:14	U	SMPL
N3810-02 - 1	COMP-2	32952	5.6997	-7471B	PPB		21 Jul 2022 15:07:30	U	SMPL
N3810-03 - 1	COMP-3	13994	2.4191	-7471B	PPB		21 Jul 2022 15:09:48	U	SMPL
N3815-01 - 1	MH-190E	5225	0.9017	-7471B	PPB		21 Jul 2022 15:12:05	U	SMPL
N3816-01 - 1	TR-05-07202022	2663	0.4584	-7471B	PPB		21 Jul 2022 15:14:21	U	SMPL
N3786-01LX5 - 1		590	0.0997	-7471B	PPB		21 Jul 2022 15:16:38	U	SMPL
N3786-01A - 1		26738	4.6244	-7471B	PPB		21 Jul 2022 15:18:55	U	SMPL
CCV67 - 1	CCV67	28343	4.9021	-7471B	PPB		21 Jul 2022 15:21:12	U	SMPL
CCB67 - 1	CCB67	-84	-0.0170	-7471B	PPB		21 Jul 2022 15:23:28	U	SMPL