

Clear table

Instrument ID: IC-2 Analyst : IZ Method: 300.0 / 9056A													
ident	concentration F-	concentration CL-	concentration on NO2	concentration on BR-	concentration on NO3	concentration on HPO4	concentration on SO4	file name	date time	Initial wt/ Final Vol	Analyst		
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2022-07-20_	7/20/22 11:28		IZ/AP		
STD2	0.4440	0.6240	0.6780	2.2120	0.5410	0.9610	3.3840	2022-07-20_	7/20/22 11:58		IZ/AP		
STD3	0.7650	1.1880	1.2040	3.9750	1.0160	1.8290	6.0140	2022-07-20_	7/20/22 12:34		IZ/AP		
STD4	0.9920	1.4780	1.4820	5.0350	1.2340	2.4870	7.4130	2022-07-20_	7/20/22 13:04		IZ/AP		
STD5	1.9830	3.0350	2.9220	9.9560	2.4600	5.3080	14.6610	2022-07-20_	7/20/22 13:37		IZ/AP		
STD6	4.0330	5.9320	5.9430	19.8010	4.9360	10.0890	29.6160	2022-07-20_	7/20/22 14:07		IZ/AP		
STD7	4.9830	7.5430	7.5710	25.1750	6.3120	12.3250	37.9120	2022-07-20_	7/20/22 14:37		IZ/AP		
ICV	1.9340	2.8690	2.8520	9.6350	2.4120	5.1520	14.2250	2022-07-20_	7/20/22 15:45		IZ/AP		
ICB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2022-07-20_	7/20/22 16:15		IZ/AP		
CCV	1.8940	2.8900	2.8950	9.7210	2.4120	5.4480	14.2720	2022-07-21_	7/21/22 9:27		IZ/AP		
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2022-07-21_	7/21/22 9:57		IZ/AP		
LB121175BLW	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2022-07-21_	7/21/22 10:27		IZ/AP		
LB121175BSW	1.8150	2.7510	2.7860	9.2950	2.2910	5.1690	13.6620	2022-07-21_	7/21/22 10:57		IZ/AP		
N3806-02	0.5260	285.6730	0.0000	0.8630	0.0000	-0.0940	65.1020	2022-07-21_	7/21/22 11:27		IZ/AP		
N3806-02MS	3.0960	276.8510	3.0600	10.7410	2.5400	10.3850	79.6610	2022-07-21_	7/21/22 11:56		IZ/AP		
N3806-02MSD	2.9560	276.4120	3.0680	10.8160	2.5750	10.5190	79.8490	2022-07-21_	7/21/22 12:26		IZ/AP		
N3796-01X10	0.0000	157.2210	0.0000	1.0610	0.0000	0.1550	1.2490	2022-07-21_	7/21/22 12:56		IZ/AP		
N3796-01DLX300	0.0000	3.9640	0.0000	0.0000	0.0000	0.0000	0.0000	2022-07-21_	7/21/22 13:25		IZ/AP		
CCV	1.8440	2.8970	2.8860	9.5900	2.3830	5.4870	14.1720	2022-07-21_	7/21/22 15:30		IZ/AP		
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2022-07-21_	7/21/22 16:02		IZ/AP		

Report date: 7/22/2022 9:16:16 AM
Printed by: wet


Ident: CCV
Analysis from: 7/21/2022 9:27:43 AM
File: _2022-07-21_

Last save: 7/21/2022 9:54:43 AM

Method: AnionsIC2-072022.mtw
Run operator: wet
Analysis number: 31144

Last save: 7/20/2022 3:36:

SAMPLE:
: IZ/AP
Vial number: 25
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.83	0.315	21.24	7.12	550.333	7.30
2	8.73	0.211	30.68	10.28	462.543	6.14
3	10.45	0.248	56.79	19.03	1023.856	13.58
4	13.41	0.304	33.08	11.09	673.886	8.94
5	15.25	0.341	41.63	13.95	982.706	13.04
6	20.57	0.508	52.51	17.60	2068.335	27.44
7	23.50	0.424	62.46	20.93	1775.812	23.56
7	27.00	0.336	298.41	100.00	7537.473	100.00

This report has been created by IC Net
METROHM LTD

Clear table

Instrument ID: IC-2 Analyst: IZ Method: 300.0 / 9056A


ident	concentra tion F-	concentratio n CL-	concentratio on NO2	concentratio on BR-	concentratio on NO3	concentratio on HPO4	concentratio on SO4	file name	date time	Initial wt/ Final Vol	Analyst
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2022-07-20_	7/20/22 11:28		IZ/AP
STD2	0.4440	0.6240	0.6780	2.2120	0.5410	0.9610	3.3840	2022-07-20_	7/20/22 11:58		IZ/AP
STD3	0.7650	1.1880	1.2040	3.9750	1.0160	1.8290	6.0140	2022-07-20_	7/20/22 12:34		IZ/AP
STD4	0.9920	1.4780	1.4820	5.0350	1.2340	2.4870	7.4130	2022-07-20_	7/20/22 13:04		IZ/AP
STD5	1.9830	3.0350	2.9220	9.9560	2.4600	5.3080	14.6610	2022-07-20_	7/20/22 13:37		IZ/AP
STD6	4.0330	5.9320	5.9430	19.8010	4.9360	10.0890	29.6160	2022-07-20_	7/20/22 14:07		IZ/AP
STD7	4.9830	7.5430	7.5710	25.1750	6.3120	12.3250	37.9120	2022-07-20_	7/20/22 14:37		IZ/AP

ident	True Value F-	True Value CL-	True Value NO2	True Value BR-	True Value NO3	True Value HPO4	True Value SO4
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000
STD2	0.4000	0.6000	0.6000	2.0000	0.5000	1.0000	3.0000
STD3	0.8000	1.2000	1.2000	4.0000	1.0000	2.0000	6.0000
STD4	1.0000	1.5000	1.5000	5.0000	1.2500	2.5000	7.5000
STD5	2.0000	3.0000	3.0000	10.0000	2.5000	5.0000	15.0000
STD6	4.0000	6.0000	6.0000	20.0000	5.0000	10.0000	30.0000
STD7	5.0000	7.5000	7.5000	25.0000	6.2500	12.5000	37.0000

ident	Relative Error F-	Relative Error CL-	Relative Error NO2	Relative Error BR-	Relative Error NO3	Relative Error HPO4	Relative Error SO4
STD1							
STD2	11.0000	4.0000	13.0000	10.6000	8.2000	-3.9000	12.8000
STD3	-4.3750	-1.0000	0.3333	-0.6250	1.6000	-8.5500	0.2333
STD4	-0.8000	-1.4667	-1.2000	0.7000	-1.2800	-0.5200	-1.1600
STD5	-0.8500	1.1667	-2.6000	-0.4400	-1.6000	6.1600	-2.2600
STD6	0.8250	-1.1333	-0.9500	-0.9950	-1.2800	0.8900	-1.2800
STD7	-0.3400	0.5733	0.9467	0.7000	0.9920	-1.4000	2.4649

CALIBRATION OF COMPONENT F-

Method: AnionsIC2-072022.mtw
 Equation: $Q = 0.0630706 \cdot A + 3.17342$
 RSD: 1.575 %
 Correlation coefficient: 0.999865



K3 = 0 K2 = 0 K1 = 0.0630706 K0 = 3.17342

Base: Area

Ref.channel: ch1

ISTD:


Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used File
1	0	0	0	0	0	0
2	3.658	90.35	0.4	20	5.653	
3	7.4	192.3	0.8	20	5.653	
4	9.63	264.3	1	20	5.653	
5	23.31	578.6	2	20	5.653	
6	48.79	1228	4	20	5.653	
7	62.13	1530	5	20	5.653	

CALIBRATION OF COMPONENT CL-

Method: AnionsIC2-072022.mtw
 Equation: $Q = 0.121487 \cdot A + 1.60479$
 RSD: 1.429 %
 Correlation coefficient: 0.999889



K3 = 0 K2 = 0 K1 = 0.121487 K0 = 1.60479

Base: Area

Ref.channel: chl

ISTD:


Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		
2	5.899	89.45	0.6	20	8.643		
3	12.1	182.3	1.2	20	8.643		
4	15.62	230	1.5	20	8.643		
5	32.79	486.5	3	20	8.643		
6	67.02	963.4	6	20	8.643		
7	87.15	1229	7.5	20	8.643		

CALIBRATION OF COMPONENT NO2

Method: AnionsIC2-072022.mtw
 Equation: $Q = 0.0526053 \cdot A + 4.04588$
 RSD: 2.187 %
 Correlation coefficient: 0.999739




K3 = 0 K2 = 0 K1 = 0.0526053 K0 = 4.04588

Base: Area
 Ref.channel: chl
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1		0	0	0	0		0
2	10.15	180.9	0.6	20	10.45		
3	21.7	380.8	1.2	20	10.45		
4	28.09	486.7	1.5	20	10.45		
5	59.36	1034	3	20	10.45		
6	124.6	2182	6	20	10.45		
7	159.9	2802	7.5	20	10.45		

CALIBRATION OF COMPONENT BR-

Method: AnionsIC2-072022.mtw
 Equation: $Q = 0.271885 \cdot A + 11.195$
 RSD: 1.708 %
 Correlation coefficient: 0.999841



K3 = 0 K2 = 0 K1 = 0.271885 K0 = 11.195

Base: Area

Ref.channel: ch1

ISTD:


Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used File
1	0	0	0	0	0	0
2	6.325	120.7	2	20	13.53	
3	12.87	240.8	4	20	13.53	
4	17.08	329.2	5	20	13.53	
5	35.56	691.2	10	20	13.53	
6	70.49	1415	20	20	13.53	
7	89.53	1811	25	20	13.53	

CALIBRATION OF COMPONENT NO3

Method: AnionsIC2-072022.mtw
 Equation: $Q = 0.0464083 \cdot A + 2.6414$
 RSD: 1.972 %
 Correlation coefficient: 0.999788



K3 = 0 K2 = 0 K1 = 0.0464083 K0 = 2.6414

Base: Area

Ref.channel: ch1

ISTD:


Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	7.674	176.1	0.5	20	15.58		
3	16.28	381.1	1	20	15.58		
4	21.38	474.8	1.25	20	15.58		
5	45.62	1003	2.5	20	15.58		
6	94.65	2070	5	20	15.58		
7	123.5	2664	6.25	20	15.58		

CALIBRATION OF COMPONENT HPO4

Method: AnionsIC2-072022.mtw
 Equation: $Q = 0.0550845 \cdot A - 4.96707$
 RSD: 3.682 %
 Correlation coefficient: 0.999261



K3 = 0 K2 = 0 K1 = 0.0550845 K0 = -4.96707

Base: Area

Ref.channel: chl

ISTD:


Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used File
1	0	0	0	0	0	0
2	9.996	439.3	1	20	20.18	
3	17.82	754.3	2	20	20.18	
4	23.86	993.1	2.5	20	20.18	
5	51.6	2017	5	20	20.18	
6	102	3753	10	20	20.18	
7	127.8	4565	12.5	20	20.18	

CALIBRATION OF COMPONENT SO4

Method: AnionsIC2-072022.mtw
 Equation: $Q = 0.151811 \cdot A + 15.8457$
 RSD: 2.323 %
 Correlation coefficient: 0.999706



K3 = 0 K2 = 0 K1 = 0.151811 K0 = 15.8457
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	11.68	341.5	3	20	23.09		
3	23.7	687.9	6	20	23.09		
4	30.92	872.3	7.5	20	23.09		
5	65.77	1827	15	20	23.09		
6	138	3797	30	20	23.09		
7	179.7	4890	37.5	20	23.09		

Report date: 7/21/2022 9:17:26 AM
Printed by: wet


Ident: STD1
Analysis from: 7/20/2022 11:28:42 AM
File: _2022-07-20_

Last save: 7/21/2022 9:16:36 AM

Method: AnionsIC2-072022.mtw
Run operator: wet
Analysis number: 31120

Last save: 7/20/2022 11:23

SAMPLE:
: IZ/AP
Vial number: 2
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 7/21/2022 9:17:35 AM
Printed by: wet


Ident: STD2
Analysis from: 7/20/2022 11:58:18 AM
File: _2022-07-20_

Last save: 7/21/2022 9:16:36 AM

Method: AnionsIC2-072022.mtw
Run operator: wet
Analysis number: 31121

Last save: 7/20/2022 11:23

SAMPLE:
: IZ/AP
Vial number: 3
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.86	0.317	3.66	6.60	90.348	6.28
2	8.71	0.210	5.90	10.65	89.447	6.22
3	10.41	0.242	10.15	18.32	180.874	12.57
4	13.28	0.291	6.34	11.45	121.548	8.45
5	15.10	0.336	7.67	13.85	176.126	12.24
6	20.63	0.562	9.99	18.04	439.265	30.52
7	23.54	0.434	11.68	21.08	341.493	23.73
7	27.00	0.342	55.40	99.99	1439.102	100.00

This report has been created by IC Net
METROHM LTD

Report date: 7/21/2022 9:17:44 AM
Printed by: wet


Ident: STD3
Analysis from: 7/20/2022 12:34:17 PM
File: _2022-07-20_

Last save: 7/21/2022 9:16:36 AM

Method: AnionsIC2-072022.mtw
Run operator: wet
Analysis number: 31123

Last save: 7/20/2022 11:23

SAMPLE:
: IZ/AP
Vial number: 4
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.82	0.323	7.40	6.60	192.272	6.79
2	8.69	0.213	12.10	10.79	182.330	6.44
3	10.40	0.240	21.69	19.35	380.799	13.46
4	13.25	0.289	13.10	11.69	251.225	8.88
5	15.03	0.328	16.28	14.52	381.134	13.47
6	20.68	0.542	17.82	15.89	754.341	26.66
7	23.57	0.432	23.69	21.14	687.902	24.31
7	27.00	0.338	112.08	99.99	2830.004	100.00

This report has been created by IC Net
METROHM LTD

Report date: 7/21/2022 9:17:52 AM
Printed by: wet


Ident: STD4
Analysis from: 7/20/2022 1:04:23 PM
File: _2022-07-20_

Last save: 7/21/2022 9:16:36 AM

Method: AnionsIC2-072022.mtw
Run operator: wet
Analysis number: 31124

Last save: 7/20/2022 11:23

SAMPLE:
: IZ/AP
Vial number: 5
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.76	0.319	9.63	6.57	264.302	7.24
2	8.55	0.207	15.62	10.66	230.043	6.30
3	10.21	0.238	28.09	19.16	486.693	13.33
4	13.02	0.289	17.08	11.65	329.172	9.02
5	14.78	0.325	21.38	14.59	474.849	13.01
6	20.45	0.531	23.86	16.28	993.065	27.20
7	23.31	0.420	30.91	21.09	872.277	23.90
7	27.00	0.333	146.57	99.99	3650.400	100.00

This report has been created by IC Net
METROHM LTD

Report date: 7/21/2022 9:18:00 AM
Printed by: wet


Ident: STD5
Analysis from: 7/20/2022 1:37:51 PM
File: _2022-07-20_

Last save: 7/21/2022 9:16:36 AM

Method: AnionsIC2-072022.mtw
Run operator: wet
Analysis number: 31125

Last save: 7/20/2022 11:23

SAMPLE:
: IZ/AP
Vial number: 6
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.78	0.298	23.31	7.42	578.618	7.58
2	8.61	0.208	32.79	10.44	486.491	6.37
3	10.29	0.240	59.36	18.90	1034.046	13.54
4	13.10	0.291	35.56	11.32	691.165	9.05
5	14.83	0.321	45.62	14.53	1003.404	13.14
6	20.56	0.501	51.60	16.43	2017.313	26.41
7	23.46	0.413	65.77	20.95	1827.099	23.92
7	27.00	0.325	314.03	100.00	7638.136	100.00

This report has been created by IC Net
METROHM LTD

Report date: 7/21/2022 9:18:07 AM
Printed by: wet


Ident: STD6
Analysis from: 7/20/2022 2:07:28 PM
File: _2022-07-20_

Last save: 7/21/2022 9:16:36 AM

Method: AnionsIC2-072022.mtw
Run operator: wet
Analysis number: 31126

Last save: 7/20/2022 11:23

SAMPLE:
: IZ/AP
Vial number: 7
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.83	0.320	48.79	7.56	1228.436	7.97
2	8.62	0.201	67.02	10.38	963.428	6.25
3	10.27	0.243	124.56	19.30	2182.372	14.16
4	13.11	0.301	70.49	10.92	1415.387	9.18
5	14.81	0.318	94.65	14.66	2070.288	13.43
6	20.56	0.477	102.01	15.80	3753.419	24.36
7	23.51	0.410	138.00	21.38	3797.275	24.64
7	27.00	0.324	645.51	100.00	15410.605	100.00

This report has been created by IC Net
METROHM LTD

Report date: 7/21/2022 9:18:14 AM
Printed by: wet


Ident: STD7
Analysis from: 7/20/2022 2:37:06 PM
File: _2022-07-20_

Last save: 7/21/2022 9:16:36 AM

Method: AnionsIC2-072022.mtw
Run operator: wet
Analysis number: 31127

Last save: 7/20/2022 11:23

SAMPLE:
: IZ/AP
Vial number: 8
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.82	0.313	62.12	7.49	1529.919	7.85
2	8.57	0.198	87.15	10.50	1228.620	6.30
3	10.20	0.244	159.90	19.27	2801.515	14.37
4	13.02	0.305	89.52	10.79	1810.721	9.29
5	14.69	0.314	123.52	14.89	2663.500	13.67
6	20.41	0.466	127.78	15.40	4565.211	23.42
7	23.36	0.406	179.65	21.65	4890.221	25.09
7	27.00	0.321	829.66	100.00	19489.707	100.00

This report has been created by IC Net
METROHM LTD

Report date: 7/21/2022 9:18:28 AM
Printed by: wet


Ident: ICV
Analysis from: 7/20/2022 3:45:02 PM
File: _2022-07-20_

Last save: 7/20/2022 4:12:02 PM

Method: AnionsIC2-072022.mtw
Run operator: wet
Analysis number: 31130

Last save: 7/20/2022 3:36:

SAMPLE:
: IZ/AP
Vial number: 26
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.82	0.317	22.00	7.24	562.852	7.60
2	8.60	0.205	31.47	10.35	459.115	6.20
3	10.25	0.238	58.40	19.21	1007.400	13.60
4	13.06	0.291	34.35	11.30	667.588	9.01
5	14.78	0.320	44.32	14.58	982.474	13.26
6	20.61	0.502	50.02	16.45	1960.610	26.46
7	23.54	0.415	63.45	20.87	1769.624	23.88
7	27.00	0.327	304.01	100.00	7409.663	100.00

This report has been created by IC Net
METROHM LTD

Report date: 7/21/2022 9:18:36 AM
Printed by: wet


Ident: ICB
Analysis from: 7/20/2022 4:15:11 PM
File: _2022-07-20_

Last save: 7/20/2022 4:42:11 PM

Method: AnionsIC2-072022.mtw
Run operator: wet
Analysis number: 31131

Last save: 7/20/2022 3:36:

SAMPLE:
: IZ/AP
Vial number: 10
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 7/22/2022 9:16:25 AM
Printed by: wet


Ident: CCB
Analysis from: 7/21/2022 9:57:52 AM
File: _2022-07-21_

Last save: 7/21/2022 10:24:52 AM

Method: AnionsIC2-072022.mtw
Run operator: wet
Analysis number: 31145

Last save: 7/20/2022 3:36:

SAMPLE:
: IZ/AP
Vial number: 12
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 7/22/2022 9:16:33 AM
Printed by: wet


Ident: LB121175BLW
Analysis from: 7/21/2022 10:27:30 AM
File: _2022-07-21_

Last save: 7/22/2022 9:15:21 AM

Method: AnionsIC2-072022.mtw
Run operator: wet
Analysis number: 31146

Last save: 7/20/2022 3:36:

SAMPLE:
: IZ/AP
Vial number: 13
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 7/22/2022 9:16:42 AM
Printed by: wet


Ident: LB121175BSW
Analysis from: 7/21/2022 10:57:08 AM
File: _2022-07-21_

Last save: 7/22/2022 9:15:21 AM

Method: AnionsIC2-072022.mtw
Run operator: wet
Analysis number: 31147

Last save: 7/20/2022 3:36:

SAMPLE:
: IZ/AP
Vial number: 14
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.80	0.320	19.98	6.95	525.082	7.31
2	8.61	0.210	29.34	10.21	439.645	6.12
3	10.29	0.245	54.97	19.13	982.279	13.68
4	13.14	0.299	32.15	11.19	642.544	8.95
5	14.91	0.332	40.93	14.24	930.337	12.95
6	20.50	0.504	50.30	17.50	1966.844	27.39
7	23.42	0.423	59.76	20.79	1695.449	23.61
7	27.00	0.333	287.42	100.00	7182.180	100.00

This report has been created by IC Net
METROHM LTD

Report date: 7/22/2022 9:17:28 AM
Printed by: wet


Ident: N3806-02
Analysis from: 7/21/2022 11:27:12 AM
File: _2022-07-21_

Last save: 7/21/2022 11:54:12 AM

Method: AnionsIC2-072022.mtw
Run operator: wet
Analysis number: 31148

Last save: 7/20/2022 3:36:

SAMPLE:
: IZ/AP
Vial number: 18
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.83	0.452	3.87	0.20	116.383	0.21
2	8.89	0.478	1584.27	83.64	47016.267	84.43
3	0.00	0.000	0.00	0.00	0.000	0.00
4	13.24	0.266	1.31	0.07	22.326	0.04
5	0.00	0.000	0.00	0.00	0.000	0.00
6	20.46	0.590	1.24	0.07	56.138	0.10
7	23.45	0.415	303.48	16.02	8472.398	15.22
7	27.00	0.314	1894.17	100.00	55683.513	100.00

This report has been created by IC Net
METROHM LTD

Report date: 7/22/2022 9:17:53 AM
Printed by: wet


Ident: N3806-02MS
Analysis from: 7/21/2022 11:56:58 AM
File: _2022-07-21_

Last save: 7/21/2022 12:23:58 PM

Method: AnionsIC2-072022.mtw
Run operator: wet
Analysis number: 31149

Last save: 7/20/2022 3:36:

SAMPLE:
: IZ/AP
Vial number: 19
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.84	0.429	28.08	1.27	931.401	1.46
2	8.90	0.465	1575.91	71.14	45563.957	71.62
3	10.31	0.245	60.48	2.73	1086.520	1.71
4	13.18	0.276	40.30	1.82	748.932	1.18
5	14.91	0.320	47.28	2.13	1037.863	1.63
6	20.64	0.515	94.36	4.26	3860.785	6.07
7	23.53	0.422	368.68	16.64	10390.373	16.33
7	27.00	0.382	2215.08	100.00	63619.832	100.00

This report has been created by IC Net
METROHM LTD

Report date: 7/22/2022 9:18:04 AM
Printed by: wet


Ident: N3806-02MSD
Analysis from: 7/21/2022 12:26:37 PM
File: _2022-07-21_

Last save: 7/21/2022 12:53:37 PM

Method: AnionsIC2-072022.mtw
Run operator: wet
Analysis number: 31150

Last save: 7/20/2022 3:36:

SAMPLE:
: IZ/AP
Vial number: 19
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.78	0.414	27.28	1.23	887.035	1.39
2	8.82	0.465	1577.44	70.92	45491.686	71.53
3	10.20	0.243	61.47	2.76	1089.550	1.71
4	13.03	0.274	40.96	1.84	754.431	1.19
5	14.74	0.316	48.27	2.17	1052.941	1.66
6	20.48	0.513	95.66	4.30	3909.375	6.15
7	23.35	0.418	373.26	16.78	10415.129	16.38
7	27.00	0.377	2224.34	100.00	63600.147	100.00

This report has been created by IC Net
METROHM LTD

Report date: 7/22/2022 9:18:22 AM
Printed by: wet


Ident: N3796-01X10
Analysis from: 7/21/2022 12:56:14 PM
File: _2022-07-21_

Last save: 7/21/2022 1:23:14 PM

Method: AnionsIC2-072022.mtw
Run operator: wet
Analysis number: 31151

Last save: 7/20/2022 3:36:

SAMPLE:
: IZ/AP
Vial number: 20
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.70	0.274	1383.40	99.54	25869.621	99.07
3	0.00	0.000	0.00	0.00	0.000	0.00
4	13.19	0.248	2.28	0.16	36.862	0.14
5	0.00	0.000	0.00	0.00	0.000	0.00
6	20.75	0.844	2.31	0.17	146.573	0.56
7	23.64	0.465	1.87	0.13	60.209	0.23
7	27.00	0.262	1389.86	100.00	26113.266	100.00

This report has been created by IC Net
METROHM LTD

Report date: 7/22/2022 9:19:46 AM
Printed by: wet


Ident: N3796-01DLX300
Analysis from: 7/21/2022 1:25:52 PM
File: _2022-07-21_

Last save: 7/22/2022 9:19:41 AM

Method: AnionsIC2-072022.mtw
Run operator: wet
Analysis number: 31152

Last save: 7/20/2022 3:36:

SAMPLE:
: IZ/AP
Vial number: 21
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.61	0.206	43.41	100.00	639.446	100.00
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	0.00	0.000	0.00	0.00	0.000	0.00
7	27.00	0.029	43.41	100.00	639.446	100.00

This report has been created by IC Net
METROHM LTD

Report date: 7/22/2022 9:20:15 AM
Printed by: wet


Ident: CCV
Analysis from: 7/21/2022 3:30:54 PM
File: _2022-07-21_

Last save: 7/21/2022 3:59:03 PM

Method: AnionsIC2-072022.mtw
Run operator: wet
Analysis number: 31156

Last save: 7/20/2022 3:36:

SAMPLE:
: IZ/AP
Vial number: 38
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.77	0.311	20.82	6.74	534.539	7.13
2	8.54	0.202	32.03	10.37	463.713	6.18
3	10.17	0.237	59.43	19.24	1020.178	13.61
4	12.95	0.289	34.39	11.13	664.306	8.86
5	14.67	0.318	44.46	14.39	970.223	12.94
6	20.42	0.496	54.27	17.57	2082.266	27.77
7	23.35	0.413	63.53	20.56	1762.711	23.51
7	27.00	0.324	308.94	100.00	7497.937	100.00

This report has been created by IC Net
METROHM LTD

Report date: 7/22/2022 9:20:24 AM
Printed by: wet


Ident: CCB
Analysis from: 7/21/2022 4:02:05 PM
File: _2022-07-21_

Last save: 7/21/2022 4:29:05 PM

Method: AnionsIC2-072022.mtw
Run operator: wet
Analysis number: 31157

Last save: 7/20/2022 3:36:

SAMPLE:
: IZ/AP
Vial number: 12
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD