



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Guerda

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB121179

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	07/21/2022	12:15
2	N3779-08		1	12	100.00	>212.0	>212.0	07/21/2022	12:35
3	N3779-08DUP		1	15	100.00	>212.0	>212.0	07/21/2022	01:20
4	N3779-09		1	18	100.00	>212.0	>212.0	07/21/2022	01:45
5	N3779-10		1	18	100.00	>212.0	>212.0	07/21/2022	02:10
6	N3779-11		1	11	100.00	>212.0	>212.0	07/21/2022	02:45
7	N3781-01		1	14	100.00	>212.0	>212.0	07/21/2022	03:15
8	N3781-02		1	15	100.00	>212.0	>212.0	07/21/2022	03:35
9	N3781-03		1	17	100.00	>212.0	>212.0	07/21/2022	04:15
10	N3781-04		1	16	100.00	>212.0	>212.0	07/21/2022	04:45
11	N3806-02		1	16	100.00	>212.0	>212.0	07/21/2022	05:10
12	N3813-06		1	16	100.00	>212.0	>212.0	07/21/2022	05:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 121179

Worklist Name : FLASHPOINT-07-21-22

Worklist ID : 161546

Department : Wet-Chemistry

Date : 07-21-2022 09:26:04

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/26/2022	Water	N3779-08	Flash Point	Cool 4 deg C	PSEG03	G12	070622-F	07/19/2022	1010B
07/26/2022	Water	N3779-09	Flash Point	Cool 4 deg C	PSEG03	G12	070622-G	07/19/2022	1010B
07/26/2022	Water	N3779-10	Flash Point	Cool 4 deg C	PSEG03	G12	070622-H	07/19/2022	1010B
07/26/2022	Water	N3779-11	Flash Point	Cool 4 deg C	PSEG03	G12	070622-I	07/19/2022	1010B
07/22/2022	Water	N3781-01	Flash Point	Cool 4 deg C	PSEG03	G12	ORA-0037	07/19/2022	1010B
07/22/2022	Water	N3781-02	Flash Point	Cool 4 deg C	PSEG03	G12	ORA-0019	07/19/2022	1010B
07/22/2022	Water	N3781-03	Flash Point	Cool 4 deg C	PSEG03	G12	ORA-0020	07/19/2022	1010B
07/22/2022	Water	N3781-04	Flash Point	Cool 4 deg C	PSEG03	G12	ORA-62322	07/19/2022	1010B
07/25/2022	Water	N3806-02	Flash Point	Cool 4 deg C	PSEG03		FRAC-TANK-N48964	07/20/2022	1010B
07/27/2022	Water	N3813-06	Flash Point	Cool 4 deg C	PARS02	N11	QUEENS54A-WC-20220720	07/20/2022	1010B

Date/Time 7/21/22 11:30
 Raw Sample Received by: BH (w)
 Raw Sample Relinquished by: OP Sm

Date/Time 7/21/22 5:40
 Raw Sample Received by: OP Sm
 Raw Sample Relinquished by: BH (w)