



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : ketankumar

Run Number: LB121182

Slope : 99.3

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	07/21/2022	14:20
2	CAL2	1	Water	NA	NA	20.7	7.00	07/21/2022	14:21
3	CAL3	1	Water	NA	NA	20.7	10.03	07/21/2022	14:22
4	ICV	1	Water	NA	NA	20.8	6.98	07/21/2022	14:25
5	CCV1	1	Water	NA	NA	20.8	2.01	07/21/2022	14:30
6	N3789-01	1	Solid	20.03	20	24.2	9.50	07/21/2022	15:30
7	N3830-06	1	Solid	20.02	20	24.7	8.60	07/21/2022	15:33
8	N3830-06DUP	1	Solid	20.03	20	24.8	8.61	07/21/2022	15:34
9	CCV2	1	Water	NA	NA	20.9	12.03	07/21/2022	15:36

WORKLIST(Hardcopy Internal Chain)

VB 121182

WorkList Name : corrosivity n3839

WorkList ID : 161559

Department : Wet-Chemistry

Date : 07-21-2022 14:04:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/19/2022	Solid	N3789-01	Corrosivity	Cool 4 deg C	JACO05	N11	IDW-VB1-071922	07/19/2022	9045D
07/21/2022	Solid	N3830-06	Corrosivity	Cool 4 deg C	PSEG03	N12	HS-SOIL-0721	07/21/2022	9045D

Date/Time 07/21/22 14:10
 Raw Sample Received by: JP (COC)
 Raw Sample Relinquished by: JP (COC)

Date/Time 07/21/22 14:10
 Raw Sample Received by: JP (COC)
 Raw Sample Relinquished by: JP (COC)