



PERCENT SOLID

Supervisor: mohammad

Analyst: JIGNESH

Date: 7/25/2022

OVENTEMP IN Celsius(°C): 107

Time IN: 15:20

In Date: 07/22/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:04

Out Date: 07/23/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB121192

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N3846-01	EXRC8	1	0.91	8.83	9.74	7.99	80.2	
N3846-02	EXRC9	2	0.92	9.03	9.95	8.53	84.3	
N3846-03	EXRD0	3	0.93	8.73	9.66	8.69	88.9	
N3846-04	EXRD1	4	0.91	8.60	9.51	8.23	85.1	
N3846-05	EXRD2	5	0.91	8.67	9.58	7.86	80.2	
N3846-06	EXRD3	6	0.95	8.89	9.84	8.51	85.0	
N3846-07	EXRD4	7	0.95	8.64	9.59	8.32	85.3	
N3846-08	EXRD5	8	0.91	8.72	9.63	8.44	86.4	
N3846-09	EXRD6	9	0.81	8.93	9.74	8.58	87.0	
N3846-10	EXRD7	10	0.92	8.60	9.52	8.04	82.8	
N3846-11	EXRD8	11	0.91	8.76	9.67	8.28	84.1	
N3846-12	EXRD9	12	0.9	8.94	9.84	8.3	82.8	
N3846-13	EXRE0	13	0.92	8.76	9.68	8.43	85.7	
N3846-14	EXRE1	14	0.91	9.07	9.98	8.68	85.7	
N3846-15	EXRE2	15	0.91	9.06	9.97	8.41	82.8	
N3846-16	EXRE3	16	0.93	9.04	9.97	8.68	85.7	
N3846-17	EXRE4	17	0.93	8.87	9.8	8.42	84.4	
N3846-18	EXRE5	18	0.89	9.00	9.89	8.2	81.2	
N3846-19	EXRE6	19	0.91	8.91	9.82	8.73	87.8	
N3846-20	EXRE9	20	0.95	8.60	9.55	8.4	86.6	
N3846-21	EXRE9D	21	0.95	8.60	9.55	8.4	86.6	
N3846-22	EXRE9S	22	0.95	8.60	9.55	8.4	86.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$