



PERCENT SOLID

Supervisor: mohammad

Analyst: JIGNESH

Date: 7/25/2022

OVENTEMP IN Celsius(°C): 107

Time IN: 15:40

In Date: 07/22/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:10

Out Date: 07/23/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB121193

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N3847-01	EXRE7	1	0.91	8.79	9.7	8.41	85.3	
N3847-02	EXRE8	2	0.93	8.68	9.61	8.12	82.8	
N3847-03	EXRF0	3	0.92	8.79	9.71	8.52	86.5	
N3847-04	EXRF1	4	0.95	9.02	9.97	8.29	81.4	
N3847-05	EXRF2	5	0.91	9.04	9.95	8.68	86.0	
N3847-06	EXRF3	6	0.98	8.92	9.9	8.71	86.7	
N3847-07	EXRF4	7	0.93	8.90	9.83	8.4	83.9	
N3847-08	EXRF5	8	0.96	8.73	9.69	8.4	85.2	
N3847-09	EXRF6	9	0.91	8.60	9.51	8.05	83.0	
N3847-10	EXRG8	10	0.95	9.01	9.96	9.1	90.5	
N3847-11	EXRG9	11	0.94	9.00	9.94	8.91	88.6	
N3847-12	EXRH0	12	0.91	8.71	9.62	8.46	86.7	
N3847-13	EXRH1	13	0.92	8.64	9.56	8.53	88.1	
N3847-14	EXRH2	14	0.93	8.73	9.66	8.75	89.6	
N3847-15	EXRH2D	15	0.93	8.73	9.66	8.75	89.6	
N3847-16	EXRH2S	16	0.93	8.73	9.66	8.75	89.6	
N3847-17	EXRH3	17	0.92	8.98	9.9	8.98	89.8	
N3847-18	EXRH4	18	0.91	9.07	9.98	8.94	88.5	
N3847-19	EXRH5	19	0.92	8.88	9.8	8.48	85.1	
N3847-20	EXRH6	20	0.91	8.75	9.66	8.64	88.3	
N3847-21	EXRH7	21	0.93	8.82	9.75	8.5	85.8	
N3847-22	EXRH8	22	0.91	9.02	9.93	9.32	93.2	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$