

LB121196


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

58718



A= 0.0000e+000

B= 1.7326e-004

C= -2.8716e-002

Rho= 0.9995954

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.064	0.064	536	0.000	536					2
0.2	0.200	0.205	0.005	1351	0.0 %	1351					-1
2.5	2.500	2.476	-0.024	14459	0.0 %	14459					0
5.0	5.000	5.010	0.010	29083	0.0 %	29083					-3
7.5	7.500	7.299	-0.201	42294	0.0 %	42294					1
10.0	10.000	10.145	0.145	58718	0.0 %	58718					

LB121196 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 22 Jul 2022 15:38:02					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	536	-	0.0000	7471B	PPB	22 Jul 2022 15:38:26	S	Std
0.2 - 1	50.2	1351	-	0.2000	7471B	PPB	22 Jul 2022 15:40:43	S	Std
2.5 - 1	52.5	14459	-	2.5000	7471B	PPB	22 Jul 2022 15:43:01	S	Std
5.0 - 1	55	29083	-	5.0000	7471B	PPB	22 Jul 2022 15:45:18	S	Std
7.5 - 1	57.5	42294	-	7.5000	7471B	PPB	22 Jul 2022 15:47:35	S	Std
10.0 - 1	510	58718	-	10.0000	7471B	PPB	22 Jul 2022 15:55:06	S	Std
ICV50 - 1	ICV50	23450	4.0342	-	7471B	PPB	22 Jul 2022 15:57:54	U	SMPL
ICB50 - 1	ICB50	498	0.0576	-	7471B	PPB	22 Jul 2022 16:00:08	U	SMPL
CCV72 - 1	CCV72	28567	4.9208	-	7471B	PPB	22 Jul 2022 16:02:24	U	SMPL
CCB72 - 1	CCB72	462	0.0513	-	7471B	PPB	22 Jul 2022 16:04:39	U	SMPL
CRA - 1	CRA	1331	0.2019	-	7471B	PPB	22 Jul 2022 16:06:56	U	SMPL
HighStd - 1	HighStd	57863	9.9966	-	7471B	PPB	22 Jul 2022 16:09:10	U	SMPL
ChkStd - 1	ChkStd	38288	6.6050	-	7471B	PPB	22 Jul 2022 16:11:26	U	SMPL
PB146397BL - 1	PBS	132	-0.0058	-	7471B	PPB	22 Jul 2022 16:13:44	U	SMPL
PB146397BS - 1	LCSS	21437	3.6854	-	7471B	PPB	22 Jul 2022 16:16:02	U	SMPL
N3814-01 - 1	ENV-DE-4(0-5)	38739	6.6832	-	7471B	PPB	22 Jul 2022 16:18:18	U	SMPL
N3814-02 - 1	ENV-DE-4(5-10)	26188	4.5086	-	7471B	PPB	22 Jul 2022 16:20:35	U	SMPL
N3814-03 - 1	ENV-DE-4(10-15)	7457	1.2633	-	7471B	PPB	22 Jul 2022 16:22:53	U	SMPL
N3814-04 - 1	ENV-DE-3(0-5)	34415	5.9340	-	7471B	PPB	22 Jul 2022 16:25:11	U	SMPL
N3814-05 - 1	ENV-DE-3(5-10)	22126	3.8048	-	7471B	PPB	22 Jul 2022 16:27:27	U	SMPL
CCV73 - 1	CCV73	29114	5.0156	-	7471B	PPB	22 Jul 2022 16:29:43	U	SMPL
CCB73 - 1	CCB73	504	0.0586	-	7471B	PPB	22 Jul 2022 16:31:59	U	SMPL
N3814-06 - 1	ENV-DE-3(10-15)	1685	0.2632	-	7471B	PPB	22 Jul 2022 16:34:16	U	SMPL
N3814-07 - 1	ENV-DE-2(0-5)	79421	13.7317	-	7471B	PPB	22 Jul 2022 16:36:31	U	SMPL
N3814-08 - 1	ENV-DE-2(5-10)	4667	0.7799	-	7471B	PPB	22 Jul 2022 16:38:46	U	SMPL
N3814-09 - 1	ENV-DE-2(10-15)	22794	3.9206	-	7471B	PPB	22 Jul 2022 16:41:04	U	SMPL
N3814-10 - 1	ENV-DE-1(0-5)	37439	6.4579	-	7471B	PPB	22 Jul 2022 16:43:20	U	SMPL
N3830-03 - 1	HS-SOIL-0721	7762	1.3161	-	7471B	PPB	22 Jul 2022 16:45:37	U	SMPL
N3833-01 - 1	OR-03-072122	2890	0.4720	-	7471B	PPB	22 Jul 2022 16:47:55	U	SMPL
N3834-01 - 1	SU-04-072122	3595	0.5942	-	7471B	PPB	22 Jul 2022 16:50:13	U	SMPL
N3834-01DUP - 1	SU-04-072122DUP	3620	0.5985	-	7471B	PPB	22 Jul 2022 16:52:30	U	SMPL
N3834-01MS - 1	SU-04-072122MS	23771	4.0898	-	7471B	PPB	22 Jul 2022 16:54:47	U	SMPL
CCV74 - 1	CCV74	28989	4.9939	-	7471B	PPB	22 Jul 2022 16:57:03	U	SMPL
CCB74 - 1	CCB74	395	0.0397	-	7471B	PPB	22 Jul 2022 16:59:19	U	SMPL
N3834-01MSD - 1	SU-04-072122MSD	23588	4.0581	-	7471B	PPB	22 Jul 2022 17:01:36	U	SMPL
N3835-01 - 1	MH-191E	22636	3.8932	-	7471B	PPB	22 Jul 2022 17:03:52	U	SMPL
N3834-01LX5 - 1		644	0.0829	-	7471B	PPB	22 Jul 2022 17:06:08	U	SMPL
N3834-01A - 1		26144	4.5010	-	7471B	PPB	22 Jul 2022 17:08:25	U	SMPL
N3814-07DLX5 - 1		18550	3.1852	-	7471B	PPB	22 Jul 2022 17:10:41	U	SMPL
CCV75 - 1	CCV75	28607	4.9277	-	7471B	PPB	22 Jul 2022 17:12:58	U	SMPL
CCB75 - 1	CCB75	118	-0.0083	-	7471B	PPB	22 Jul 2022 17:15:15	U	SMPL