



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB121203

Analyst By : JIGNESH

Supervisor Review By : ketankumar

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	07/22/2022	16:00
2	CAL2	1	Water	NA	NA	20.7	7.00	07/22/2022	16:01
3	CAL3	1	Water	NA	NA	20.7	10.03	07/22/2022	16:02
4	ICV	1	Water	NA	NA	20.8	6.99	07/22/2022	16:05
5	CCV1	1	Water	NA	NA	20.8	2.02	07/22/2022	16:35
6	N3845-01	1	Water	NA	NA	23.6	4.89	07/22/2022	16:40
7	N3845-01DUP	1	Water	NA	NA	23.7	4.90	07/22/2022	16:41
8	CCV2	1	Water	NA	NA	20.7	12.03	07/22/2022	16:45

13121203

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N3845

WorkList ID : 161600

Department : Wet-Chemistry

Date : 07-22-2022 15:43:34

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/18/2022	Water	N3845-01	pH	Cool 4 deg C	STER14	N11	JS10000	07/18/2022	9040C

Date/Time 07/22/22 15:50
Raw Sample Received by: CP sm
Raw Sample Relinquished by: CP sm

Date/Time 07/22/22 18:00
Raw Sample Received by: CP sm
Raw Sample Relinquished by: CP sm