



Analytical Summary Report

Analysis Method: 1010B
Parameter: Flash Point
Run Number: LB121208
Thermometer ID: flashpoint

Reviewed By: ketankumar
Supervisor Review By: mohammad
Ambient Barometric Pressure (mmHg): 765.00
Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	07/22/2022	09:15
2	N3789-01		1	12	100.00	>212.0	>212.0	07/22/2022	09:40
3	N3789-01DUP		1	12	100.00	>212.0	>212.0	07/22/2022	10:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB121208

WorkList Name : FLASHPOINT-07-22-22 WorkList ID : 161576 Department : Wet-Chemistry Date : 07-22-2022 08:25:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/02/2022	Solid	N3789-01	Flash Point	Cool 4 deg C	JACO05	N11	IDW-VB1-071922	07/19/2022	1010B

Date/Time 07/22/22 08:45
 Raw Sample Received by: KS(WC)
 Raw Sample Relinquished by: [Signature]

Date/Time 07/22/22 10:15
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]