


LB121209

SFAM01.1

INSTRUMENT ID: CVL

Linear



A= 0.0000e+000

B= 2.2648e-004 slope

C= 1.2104e-002 y-intercept

Rho= 0.9998427

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.040	0.040	125	0.000	125					=
0.05	0.050					185	0				2
0.20	0.200	0.203	0.003	842	0.0 %	842					-3
2.50	2.500	2.430	-0.070	10677	0.0 %	10677					-1
5.00	5.000	4.963	-0.037	21860	0.0 %	21860					2
7.50	7.500	7.621	0.121	33594	0.0 %	33594					-1
10.0	10.000	9.943	-0.057	43847	0.0 %	43847					

LB121209

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	125	-		0 SFAM01.1	PPB	7/25/2022 13:44	Std
	0.2 S01	842	-		0.2 SFAM01.1	PPB	7/25/2022 13:46	Std
	2.5 S02	10677	-		2.5 SFAM01.1	PPB	7/25/2022 13:50	Std
	5 S03	21860	-		5 SFAM01.1	PPB	7/25/2022 13:52	Std
	7.5 S04	33594	-		7.5 SFAM01.1	PPB	7/25/2022 13:55	Std
	10 S05	43847	-		10 SFAM01.1	PPB	7/25/2022 13:59	Std
ICV051	ICV051	17761	4.0347 -		SFAM01.1	PPB	7/25/2022 14:02	SMPL
ICB051	ICB051	116	0.0384 -		SFAM01.1	PPB	7/25/2022 14:04	SMPL
CCV076	CCV076	22045	5.005 -		SFAM01.1	PPB	7/25/2022 14:07	SMPL
CCB076	CCB076	128	0.0411 -		SFAM01.1	PPB	7/25/2022 14:09	SMPL
PB146429BL	PBW429	165	0.0495 -		SFAM01.1	PPB	7/25/2022 14:11	SMPL
N3723-07	X3B68	24574	5.5777 -		SFAM01.1	PPB	7/25/2022 14:16	SMPL
N3723-08	X3B68D	24246	5.5035 -		SFAM01.1	PPB	7/25/2022 14:18	SMPL
N3723-09	X3B68S	29298	6.6477 -		SFAM01.1	PPB	7/25/2022 14:25	SMPL
CCV077	CCV077	21893	4.9705 -		SFAM01.1	PPB	7/25/2022 14:27	SMPL
CCB077	CCB077	104	0.0357 -		SFAM01.1	PPB	7/25/2022 14:29	SMPL