


LB121212

SFAM01.1

INSTRUMENT ID: CV1

 μ Abs.:

45891



Linear

A= 0.0000e+000

B= 2.2090e-004 slope

C= -1.6533e-001 $y = 1.6533x$

Rho= 0.9994255

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/b
0.00	0.000	-0.071	-0.071	425	0.000	425					=
0.05	0.050					185	0				23
0.20	0.200	0.245	0.045	1859	0.0 %	1859					-3
2.50	2.500	2.418	-0.082	11696	0.0 %	11696					5
5.00	5.000	5.255	0.255	24537	0.0 %	24537					-2
7.50	7.500	7.381	-0.119	34160	0.0 %	34160					0
10.0	10.000	9.972	-0.028	45891	0.0 %	45891					

LB121212
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	425	-		0 SFAM01.1	PPB	7/25/2022 15:42	Std
	0.2 S01	1859	-		0.2 SFAM01.1	PPB	7/25/2022 15:44	Std
	2.5 S02	11696	-		2.5 SFAM01.1	PPB	7/25/2022 15:49	Std
	5 S03	24537	-		5 SFAM01.1	PPB	7/25/2022 15:53	Std
	7.5 S04	34160	-		7.5 SFAM01.1	PPB	7/25/2022 15:55	Std
	10 S05	45891	-		10 SFAM01.1	PPB	7/25/2022 15:57	Std
ICV052	ICV052	18654	3.9554 -		SFAM01.1	PPB	7/25/2022 16:01	SMPL
ICB052	ICB052	330	-0.0924 -		SFAM01.1	PPB	7/25/2022 16:05	SMPL
CCV078	CCV078	23560	5.0391 -		SFAM01.1	PPB	7/25/2022 16:08	SMPL
CCB078	CCB078	315	-0.0957 -		SFAM01.1	PPB	7/25/2022 16:10	SMPL
PB146430BL	PBS430	493	-0.0564 -		SFAM01.1	PPB	7/25/2022 16:12	SMPL
N3723-19	X3B72	33881	7.3191 -		SFAM01.1	PPB	7/25/2022 16:17	SMPL
N3723-20	X3B72D	32874	7.0966 -		SFAM01.1	PPB	7/25/2022 16:26	SMPL
N3723-21	X3B72S	43293	9.3982 -		SFAM01.1	PPB	7/25/2022 16:28	SMPL
CCV079	CCV079	23695	5.069 -		SFAM01.1	PPB	7/25/2022 16:35	SMPL
CCB079	CCB079	293	-0.1006 -		SFAM01.1	PPB	7/25/2022 16:40	SMPL