

LB121225


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

57551



A= 0.0000e+000

B= 1.7477e-004

C= -2.6705e-002

Rho= 0.9999621

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	90D
0.0	0.000	0.007	0.007	191	0.000	191					
0.2	0.200	0.185	-0.015	1213	0.0 %	1213					-8
2.5	2.500	2.516	0.016	14550	0.0 %	14550					1
5.0	5.000	5.024	0.024	28902	0.0 %	28902					0
7.5	7.500	7.436	-0.064	42701	0.0 %	42701					-1
10.0	10.000	10.031	0.031	57551	0.0 %	57551					0

LB121225 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 26 Jul 2022 07:55:02

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	191	-	0.00007471B	PPB	26 Jul 2022 08:08:04	S	Std
0.2 - 1	50.2	1213	-	0.20007471B	PPB	26 Jul 2022 08:10:21	S	Std
2.5 - 1	52.5	14550	-	2.50007471B	PPB	26 Jul 2022 08:12:38	S	Std
5.0 - 1	55	28902	-	5.00007471B	PPB	26 Jul 2022 08:14:55	S	Std
7.5 - 1	57.5	42701	-	7.50007471B	PPB	26 Jul 2022 08:17:12	S	Std
10.0 - 1	510	57551	-	10.00007471B	PPB	26 Jul 2022 08:19:29	S	Std
ICV53 - 1	ICV53	22357	3.8806	-7471B	PPB	26 Jul 2022 08:22:14	U	SMPL
ICB53 - 1	ICB53	72	-0.0141	-7471B	PPB	26 Jul 2022 08:24:29	U	SMPL
CCV80 - 1	CCV80	29111	5.0610	-7471B	PPB	26 Jul 2022 08:26:45	U	SMPL
CCB80 - 1	CCB80	54	-0.0173	-7471B	PPB	26 Jul 2022 08:29:00	U	SMPL
CRA - 1	CRA	1133	0.1713	-7471B	PPB	26 Jul 2022 08:31:17	U	SMPL
HighStd - 1	HighStd	57812	10.0770	-7471B	PPB	26 Jul 2022 08:33:32	U	SMPL
ChkStd - 1	ChkStd	40359	7.0268	-7471B	PPB	26 Jul 2022 08:35:47	U	SMPL
PB146449BL - 1	PBS	-5	-0.0276	-7471B	PPB	26 Jul 2022 08:38:05	U	SMPL
PB146449BS - 1	LCSS	20779	3.6048	-7471B	PPB	26 Jul 2022 08:40:23	U	SMPL
N3836-01 - 1	TP-1	7123	1.2182	-7471B	PPB	26 Jul 2022 08:42:39	U	SMPL
N3836-02 - 1	TP-2	11857	2.0455	-7471B	PPB	26 Jul 2022 08:44:56	U	SMPL
N3839-01 - 1	ENV-DE-1-(5-10)	19443	3.3713	-7471B	PPB	26 Jul 2022 08:47:13	U	SMPL
N3839-02 - 1	ENV-DE-1-(10-15)	7489	1.2821	-7471B	PPB	26 Jul 2022 08:49:31	U	SMPL
N3839-03 - 1	ENV-DE-5-(0-5)	1212	0.1851	-7471B	PPB	26 Jul 2022 08:51:47	U	SMPL
CCV81 - 1	CCV81	29022	5.0454	-7471B	PPB	26 Jul 2022 08:54:02	U	SMPL
CCB81 - 1	CCB81	50	-0.0180	-7471B	PPB	26 Jul 2022 08:56:18	U	SMPL
N3839-03DUP - 1	ENV-DE-5-(0-5)DUP	965	0.1419	-7471B	PPB	26 Jul 2022 08:58:34	U	SMPL
N3839-03MS - 1	ENV-DE-5-(0-5)MS	24833	4.3133	-7471B	PPB	26 Jul 2022 09:00:49	U	SMPL
N3839-03MSD - 1	ENV-DE-5-(0-5)MSD	22739	3.9473	-7471B	PPB	26 Jul 2022 09:03:04	U	SMPL
N3839-04 - 1	FIELD-DUPLICATE-01	1160	0.1760	-7471B	PPB	26 Jul 2022 09:05:21	U	SMPL
N3854-01 - 1	MH-189E	20956	3.6357	-7471B	PPB	26 Jul 2022 09:07:37	U	SMPL
N3855-01 - 1	EO-01-072222	8755	1.5034	-7471B	PPB	26 Jul 2022 09:09:54	U	SMPL
N3855-03 - 1	EO-03-072222	13378	2.3113	-7471B	PPB	26 Jul 2022 09:12:11	U	SMPL
N3857-01 - 1	TP-3	14327	2.4772	-7471B	PPB	26 Jul 2022 09:14:28	U	SMPL
N3857-02 - 1	TP-4	6320	1.0778	-7471B	PPB	26 Jul 2022 09:16:45	U	SMPL
N3858-01 - 1	SB-A	75686	13.2008	-7471B	PPB	26 Jul 2022 09:19:00	U	SMPL
CCV82 - 1	CCV82	28836	5.0129	-7471B	PPB	26 Jul 2022 09:21:16	U	SMPL
CCB82 - 1	CCB82	106	-0.0082	-7471B	PPB	26 Jul 2022 09:24:56	U	SMPL
N3858-03 - 1	SB-B	6579	1.1231	-7471B	PPB	26 Jul 2022 09:27:11	U	SMPL
N3839-03LX5 - 1		228	0.0131	-7471B	PPB	26 Jul 2022 09:29:27	U	SMPL
N3839-03A - 1		24423	4.2417	-7471B	PPB	26 Jul 2022 09:31:43	U	SMPL
N3858-01DLX5 - 1		16293	2.8208	-7471B	PPB	26 Jul 2022 09:33:59	U	SMPL
CCV83 - 1	CCV83	28678	4.9853	-7471B	PPB	26 Jul 2022 09:36:16	U	SMPL
CCB83 - 1	CCB83	50	-0.0180	-7471B	PPB	26 Jul 2022 09:38:33	U	SMPL