

LB 121231


7470A

INSTRUMENT ID: CVL

Linear

 μ Abs.:

4489E



A= 0.0000e+000

B= 2.2533e-004

C= 8.4846e-003

Rho= 0.9990632

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	1/01)
0.0	0.000	0.010	0.010	5	0.000	5					-7
0.2	0.200	0.186	-0.014	786	0.0 %	786					0
2.5	2.500	2.504	0.004	11075	0.0 %	11075					4
5.0	5.000	5.192	0.192	23003	0.0 %	23003					-41
7.5	7.500	7.184	-0.316	31842	0.0 %	31842					1
10.0	10.000	10.125	0.125	44898	0.0 %	44898					

LB121231 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 26 Jul 2022 14:52:09

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	5	-	0.0000	7470A	PPB	26 Jul 2022 14:52:27	S	Std
0.2 - 1	50.2	786	-	0.2000	7470A	PPB	26 Jul 2022 14:54:45	S	Std
2.5 - 1	52.5	11075	-	2.5000	7470A	PPB	26 Jul 2022 14:57:02	S	Std
5.0 - 1	55	23003	-	5.0000	7470A	PPB	26 Jul 2022 14:59:19	S	Std
7.5 - 1	57.5	31842	-	7.5000	7470A	PPB	26 Jul 2022 15:01:35	S	Std
10.0 - 1	51.5	44898	-	10.0000	7470A	PPB	26 Jul 2022 15:03:50	S	Std
ICV54 - 1	ICV54	16491	3.7244	-	7470A	PPB	26 Jul 2022 15:09:07	U	SMPL
ICB54 - 1	ICB54	-6	0.0071	-	7470A	PPB	26 Jul 2022 15:11:22	U	SMPL
CCV84 - 1	CCV84	20611	4.6528	-	7470A	PPB	26 Jul 2022 15:13:37	U	SMPL
CCB84 - 1	CCB84	-25	0.0029	-	7470A	PPB	26 Jul 2022 15:15:52	U	SMPL
CRA - 1	CRA	710	0.1685	-	7470A	PPB	26 Jul 2022 15:18:07	U	SMPL
HighStd - 1	HighStd	40247	9.0774	-	7470A	PPB	26 Jul 2022 15:20:22	U	SMPL
ChkStd - 1	ChkStd	27854	6.2849	-	7470A	PPB	26 Jul 2022 15:25:12	U	SMPL
PB146466BL - 1	PBW	-52	-0.0032	-	7470A	PPB	26 Jul 2022 15:27:29	U	SMPL
N3839-06 - 1	ENV-DE-WC-1	-67	-0.0066	-	7470A	PPB	26 Jul 2022 15:32:02	U	SMPL
N3854-03 - 1	MH-189E	16	0.0121	-	7470A	PPB	26 Jul 2022 15:34:19	U	SMPL
N3878-02 - 1	200-E-5TH-ST-DIRT-STONE-ASPHALT	30	0.0152	-	7470A	PPB	26 Jul 2022 15:36:35	U	SMPL
N3881-05 - 1	4299E	32	0.0157	-	7470A	PPB	26 Jul 2022 15:38:52	U	SMPL
N3881-06 - 1	TP2	29	0.0150	-	7470A	PPB	26 Jul 2022 15:41:07	U	SMPL
CCV85 - 1	CCV85	21494	4.8518	-	7470A	PPB	26 Jul 2022 15:43:22	U	SMPL
CCB85 - 1	CCB85	-27	0.0024	-	7470A	PPB	26 Jul 2022 15:45:38	U	SMPL
N3888-02 - 1	S-2ND-TP	55	0.0209	-	7470A	PPB	26 Jul 2022 15:47:53	U	SMPL
N3888-02DUP - 1	S-2ND-TPDUP	64	0.0229	-	7470A	PPB	26 Jul 2022 15:50:08	U	SMPL
N3888-02MS - 1	S-2ND-TPMS	16833	3.8015	-	7470A	PPB	26 Jul 2022 15:52:24	U	SMPL
N3888-02MSD - 1	S-2ND-TPMSD	17719	4.0011	-	7470A	PPB	26 Jul 2022 15:54:39	U	SMPL
PB146467BL - 1	PBW	-20	0.0040	-	7470A	PPB	26 Jul 2022 15:56:55	U	SMPL
PB146467BS - 1	LCSW	16233	3.6663	-	7470A	PPB	26 Jul 2022 16:01:59	U	SMPL
N3813-06 - 1	QUEENS54A-WC-20220720	-29	0.0020	-	7470A	PPB	26 Jul 2022 16:04:15	U	SMPL
N3813-06DUP - 1	QUEENS54A-WC-20220720DUP	25	0.0141	-	7470A	PPB	26 Jul 2022 16:06:32	U	SMPL
N3813-06MS - 1	QUEENS54A-WC-20220720MS	17394	3.9279	-	7470A	PPB	26 Jul 2022 16:13:59	U	SMPL
N3813-06MSD - 1	QUEENS54A-WC-20220720MSD	18318	4.1361	-	7470A	PPB	26 Jul 2022 16:16:14	U	SMPL
CCV86 - 1	CCV86	21731	4.9052	-	7470A	PPB	26 Jul 2022 16:18:31	U	SMPL
CCB86 - 1	CCB86	-22	0.0035	-	7470A	PPB	26 Jul 2022 16:20:46	U	SMPL
PB146468BL - 1	PBW	42	0.0179	-	7470A	PPB	26 Jul 2022 16:23:02	U	SMPL
PB146468BS - 1	LCSW	18235	4.1174	-	7470A	PPB	26 Jul 2022 16:25:17	U	SMPL
N3877-02 - 1	TT-049-IDWGW-20220722	40	0.0175	-	7470A	PPB	26 Jul 2022 16:27:33	U	SMPL
N3877-02DUP - 1	TT-049-IDWGW-20220722DUP	47	0.0191	-	7470A	PPB	26 Jul 2022 16:29:49	U	SMPL
N3877-02MS - 1	TT-049-IDWGW-20220722MS	19769	4.4631	-	7470A	PPB	26 Jul 2022 16:42:14	U	SMPL
N3877-02MSD - 1	TT-049-IDWGW-20220722MSD	16895	3.8155	-	7470A	PPB	26 Jul 2022 16:47:04	U	SMPL
PB146424TB - 1	PB146424TB	-33	0.0010	-	7470A	PPB	26 Jul 2022 16:49:21	U	SMPL
PB146426TB - 1	PB146426TB	11	0.0110	-	7470A	PPB	26 Jul 2022 16:51:38	U	SMPL
PB146466BS - 1	LCSW	16308	3.6832	-	7470A	PPB	26 Jul 2022 16:53:55	U	SMPL
CCV87 - 1	CCV87	20554	4.6400	-	7470A	PPB	26 Jul 2022 16:56:10	U	SMPL
CCB87 - 1	CCB87	-50	-0.0028	-	7470A	PPB	26 Jul 2022 16:58:26	U	SMPL
N3888-02LX5 - 1		4	0.0094	-	7470A	PPB	26 Jul 2022 17:00:43	U	SMPL
N3888-02A - 1		18957	4.2801	-	7470A	PPB	26 Jul 2022 17:05:53	U	SMPL
N3813-06LX5 - 1		-38	-0.0001	-	7470A	PPB	26 Jul 2022 17:08:09	U	SMPL
N3813-06A - 1		16332	3.6886	-	7470A	PPB	26 Jul 2022 17:10:25	U	SMPL
N3877-02LX5 - 1		-42	-0.0010	-	7470A	PPB	26 Jul 2022 17:12:40	U	SMPL
N3877-02A - 1		18191	4.1075	-	7470A	PPB	26 Jul 2022 17:14:56	U	SMPL
CCV88 - 1	CCV88	22820	5.1506	-	7470A	PPB	26 Jul 2022 17:17:12	U	SMPL
CCB88 - 1	CCB88	-38	-0.0001	-	7470A	PPB	26 Jul 2022 17:19:28	U	SMPL