



Analytical Summary Report

Analysis Method: 9040C
Parameter: pH
Run Number: LB121235

Analyst By : JIGNESH

Supervisor Review By : mohammad

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	07/26/2022	16:00
2	CAL2	1	Water	NA	NA	20.8	7.00	07/26/2022	16:01
3	CAL3	1	Water	NA	NA	20.9	10.03	07/26/2022	16:02
4	ICV	1	Water	NA	NA	20.9	6.98	07/26/2022	16:05
5	CCV1	1	Water	NA	NA	20.9	2.01	07/26/2022	16:10
6	N3877-02	1	Water	NA	NA	20.9	7.47	07/26/2022	16:15
7	N3898-02	1	Water	NA	NA	25.4	6.12	07/26/2022	16:19
8	N3898-03	1	Water	NA	NA	25.4	6.17	07/26/2022	16:22
9	N3898-04	1	Water	NA	NA	25.3	6.59	07/26/2022	16:25
10	N3898-05	1	Water	NA	NA	25.1	6.57	07/26/2022	16:30
11	N3898-06	1	Water	NA	NA	25.6	4.92	07/26/2022	16:32
12	N3898-07	1	Water	NA	NA	25.4	8.14	07/26/2022	16:35
13	N3907-01	1	Water	NA	NA	23.1	6.94	07/26/2022	16:36
14	N3916-03	1	Water	NA	NA	25.7	5.64	07/26/2022	16:40
15	N3916-03DUP	1	Water	NA	NA	25.6	5.65	07/26/2022	16:41
16	CCV2	1	Water	NA	NA	20.8	12.03	07/26/2022	16:45

WORKLIST(Hardcopy Internal Chain)

VB 121235

WorkList Name : PH N3898

WorkList ID : 161677

Department : Wet-Chemistry

Date : 07-26-2022 16:11:40

Due Date	Matrix	Sample	Test	Raw Sample			Collect Date	Method
				Preservative	Customer	Storage Location		
07/22/2022	Water	N3877-02	pH C	Cool 4 deg C	TETR06	N11 TT-049-IDWGW-20220722	07/22/2022	9040C
07/26/2022	Water	N3898-02	pH C	Cool 4 deg C	PSEG03	G11 61422-A	07/26/2022	9040C
07/26/2022	Water	N3898-03	pH C	Cool 4 deg C	PSEG03	G11 61422-C	07/26/2022	9040C
07/26/2022	Water	N3898-04	pH C	Cool 4 deg C	PSEG03	G11 61422-D	07/26/2022	9040C
07/26/2022	Water	N3898-05	pH C	Cool 4 deg C	PSEG03	G11 61422-B	07/26/2022	9040C
07/26/2022	Water	N3898-06	pH C	Cool 4 deg C	PSEG03	G11 61422-A	07/26/2022	9040C
07/26/2022	Water	N3898-07	pH C	Cool 4 deg C	PSEG03	G11 62422	07/26/2022	9040C
07/26/2022	Water	N3907-01	pH A	Cool 4 deg C	PSEG03	G11 CMS-OWS-OUTLET	07/26/2022	9040C
07/26/2022	Water	N3916-03	pH C	Cool 4 deg C	PSEG03	G11 998	07/26/2022	9040C

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by: