



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : mohammad

Run Number: LB121236

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	07/26/2022	16:00
2	CAL2	1	Water	NA	NA	20.8	7.00	07/26/2022	16:01
3	CAL3	1	Water	NA	NA	20.9	10.03	07/26/2022	16:02
4	ICV	1	Water	NA	NA	20.9	6.98	07/26/2022	16:05
5	CCV1	1	Water	NA	NA	20.2	2.02	07/26/2022	16:50
6	N3914-01	1	Solid	20.03	20	22.9	8.32	07/26/2022	17:00
7	N3914-01DUP	1	Solid	20.04	20	22.8	8.33	07/26/2022	17:01
8	CCV2	1	Water	NA	NA	20.8	12.03	07/26/2022	17:05

VP 12V36

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N3914 S WorkList ID : 161678 Department : Wet-Chemistry Date : 07-26-2022 16:12:29

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/23/2022	Solid	N3914-01	pH	E	Cool 4 deg C	AECO11	G11	METRO-DISP	9045D

Date/Time 07/26/22 15:50
Raw Sample Received by: 70600'
Raw Sample Relinquished by: CP SSM

Date/Time 07/26/22 18:00
Raw Sample Received by: CP SSM
Raw Sample Relinquished by: 70600'