

LB 121243


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

5611E



A= 0.0000e+000

B= 1.7973e-004

C= 2.7279e-002

Rho= 0.9997017

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.031	0.031	22	0.000	22					
0.2	0.200	0.203	0.003	978	0.0 %	978					
2.5	2.500	2.495	-0.005	13730	0.0 %	13730					
5.0	5.000	5.040	0.040	27892	0.0 %	27892					
7.5	7.500	7.317	-0.183	40563	0.0 %	40563					
10.0	10.000	10.113	0.113	56118	0.0 %	56118					

0/00
 2
 0
 1
 -2
 1

Method: 7471B

Operator: Admin

LB121243 INSTRUMENT ID : CV1

Date of Analysis: 27 Jul 2022 12:22:13

Sample ID	Extended ID	μ Abs	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	22	-	0.00007471B	PPB	27 Jul 2022 12:32:45	S	Std
0.2 - 1	50.2	978	-	0.20007471B	PPB	27 Jul 2022 12:35:02	S	Std
2.5 - 1	52.5	13730	-	2.50007471B	PPB	27 Jul 2022 12:37:20	S	Std
5.0 - 1	55	27892	-	5.00007471B	PPB	27 Jul 2022 12:39:36	S	Std
7.5 - 1	57	40563	-	7.50007471B	PPB	27 Jul 2022 12:41:53	S	Std
10.0 - 1	57P	56118	-	10.00007471B	PPB	27 Jul 2022 12:44:11	S	Std
ICV55 - 1	ICV55	22078	3.9953	-7471B	PPB	27 Jul 2022 12:47:02	U	SMPL
ICB55 - 1	ICB55	-73	0.0142	-7471B	PPB	27 Jul 2022 12:49:16	U	SMPL
CCV89 - 1	CCV89	27563	4.9811	-7471B	PPB	27 Jul 2022 12:51:32	U	SMPL
CCB89 - 1	CCB89	-37	0.0206	-7471B	PPB	27 Jul 2022 12:53:48	U	SMPL
CRA - 1	CRA	952	0.1984	-7471B	PPB	27 Jul 2022 12:56:03	U	SMPL
HighStd - 1	HighStd	55197	9.9476	-7471B	PPB	27 Jul 2022 12:58:18	U	SMPL
ChkStd - 1	ChkStd	37715	6.8056	-7471B	PPB	27 Jul 2022 13:00:33	U	SMPL
PB146520BL - 1	PBS	-145	0.0012	-7471B	PPB	27 Jul 2022 13:02:51	U	SMPL
PB146520BS - 1	LCSS	19302	3.4963	-7471B	PPB	27 Jul 2022 13:05:09	U	SMPL
N3878-01 - 1	200-E-5TH-ST-DIRT-STONE-ASPHALT	4299E	1.0097	-7471B	PPB	27 Jul 2022 13:07:25	U	SMPL
N3881-01 - 1	4299E	5466	1.0097	-7471B	PPB	27 Jul 2022 13:09:42	U	SMPL
N3881-03 - 1	TP2	4587	0.8517	-7471B	PPB	27 Jul 2022 13:11:59	U	SMPL
N3883-01 - 1	CL-01-072522	821	0.1748	-7471B	PPB	27 Jul 2022 13:14:16	U	SMPL
N3886-01 - 1	WH-1	5322	0.9838	-7471B	PPB	27 Jul 2022 13:16:31	U	SMPL
CCV90 - 1	CCV90	2966	0.5603	-7471B	PPB	27 Jul 2022 13:18:46	U	SMPL
CCB90 - 1	CCB90	27172	4.9108	-7471B	PPB	27 Jul 2022 13:21:01	U	SMPL
N3888-01 - 1	S-2ND-TP	-88	0.0115	-7471B	PPB	27 Jul 2022 13:23:17	U	SMPL
N3891-01 - 1	4553-E	5884	1.0848	-7471B	PPB	27 Jul 2022 13:25:33	U	SMPL
N3891-01DUP - 1	4553-EDUP	5036	0.9324	-7471B	PPB	27 Jul 2022 13:27:48	U	SMPL
N3891-01MS - 1	4553-EMS	4743	0.8797	-7471B	PPB	27 Jul 2022 13:30:03	U	SMPL
N3891-01MSD - 1	4553-EMSD	26699	4.8258	-7471B	PPB	27 Jul 2022 13:32:19	U	SMPL
N3891-04 - 1	TP-1	24996	4.5197	-7471B	PPB	27 Jul 2022 13:34:37	U	SMPL
N3891-07 - 1	7897-W	13418	2.4388	-7471B	PPB	27 Jul 2022 13:36:54	U	SMPL
N3892-01 - 1	BIN-4	4346	0.8084	-7471B	PPB	27 Jul 2022 13:39:12	U	SMPL
N3906-01 - 1	NB-7-7-26-22	1059	0.2176	-7471B	PPB	27 Jul 2022 13:41:29	U	SMPL
N3912-01 - 1	6601-S	1295	0.2600	-7471B	PPB	27 Jul 2022 13:43:45	U	SMPL
CCV91 - 1	CCV91	11553	2.1036	-7471B	PPB	27 Jul 2022 13:46:01	U	SMPL
CCB91 - 1	CCB91	27425	4.9563	-7471B	PPB	27 Jul 2022 13:48:17	U	SMPL
N3918-01 - 1	OR-02-07-26-2022	-41	0.0199	-7471B	PPB	27 Jul 2022 13:50:33	U	SMPL
N3891-01LX5 - 1		1271	0.2557	-7471B	PPB	27 Jul 2022 13:52:49	U	SMPL
N3891-01A - 1		829	0.1763	-7471B	PPB	27 Jul 2022 13:55:05	U	SMPL
CCV92 - 1	CCV92	22888	4.1408	-7471B	PPB	27 Jul 2022 13:57:20	U	SMPL
CCB92 - 1	CCB92	27344	4.9417	-7471B	PPB	27 Jul 2022 13:59:38	U	SMPL
		-96	0.0100	-7471B	PPB			