



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : mohammad

Run Number: LB121249

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	07/27/2022	15:30
2	CAL2	1	Water	NA	NA	20.8	7.00	07/27/2022	15:31
3	CAL3	1	Water	NA	NA	20.8	10.03	07/27/2022	15:32
4	ICV	1	Water	NA	NA	20.9	6.99	07/27/2022	15:35
5	CCV1	1	Water	NA	NA	20.8	2.01	07/27/2022	15:36
6	N3915-01	1	Solid	20.03	20	18.7	6.98	07/27/2022	15:40
7	N3915-02	1	Solid	20.05	20	21.2	7.46	07/27/2022	15:43
8	N3915-03	1	Solid	20.02	20	22.3	7.53	07/27/2022	15:45
9	N3915-04	1	Solid	20.03	20	21.6	7.88	07/27/2022	15:50
10	N3915-05	1	Solid	20.02	20	21.4	7.45	07/27/2022	16:00
11	N3915-06	1	Solid	20.03	20	21.8	7.50	07/27/2022	16:02
12	N3915-07	1	Solid	20.05	20	21.3	7.74	07/27/2022	16:10
13	N3915-08	1	Solid	20.02	20	21.3	7.76	07/27/2022	16:15
14	N3915-09	1	Solid	20.03	20	21.1	7.11	07/27/2022	16:20
15	N3915-10	1	Solid	20.05	20	21.9	5.25	07/27/2022	16:22
16	CCV2	1	Water	NA	NA	20.8	12.03	07/27/2022	16:25
17	N3915-11	1	Solid	20.03	20	21.7	3.71	07/27/2022	16:29
18	N3915-12	1	Solid	20.01	20	21.5	7.13	07/27/2022	16:33
19	N3915-13	1	Solid	20.02	20	22.1	6.89	07/27/2022	16:36
20	N3915-14	1	Solid	20.03	20	21.7	6.92	07/27/2022	16:30

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	N3915-15	1	Solid	20.03	20	21.7	6.92	07/27/2022	16:30
22	N3915-16	1	Solid	20.03	20	21.7	6.92	07/27/2022	16:30
23	CCV3	1	Water	NA	NA	20.9	2.01	07/27/2022	16:33

WORKLIST(Hardcopy Internal Chain)

12/22/22

WorkList Name : PH N3915 WorkList ID : 161729 Department : Wet-Chemistry Date : 07-27-2022 15:21:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/22/2022	Solid	N3915-01	pH	Cool 4 deg C	USEP01	YBB50		07/22/2022	9045D
07/22/2022	Solid	N3915-02	pH	Cool 4 deg C	USEP01	YBB52		07/22/2022	9045D
07/22/2022	Solid	N3915-03	pH	Cool 4 deg C	USEP01	YBB54		07/22/2022	9045D
07/22/2022	Solid	N3915-04	pH	Cool 4 deg C	USEP01	YBB57		07/22/2022	9045D
07/22/2022	Solid	N3915-05	pH	Cool 4 deg C	USEP01	YBB58		07/22/2022	9045D
07/22/2022	Solid	N3915-06	pH	Cool 4 deg C	USEP01	YBB59		07/22/2022	9045D
07/22/2022	Solid	N3915-07	pH	Cool 4 deg C	USEP01	YBB64		07/22/2022	9045D
07/22/2022	Solid	N3915-08	pH	Cool 4 deg C	USEP01	YBBC6		07/22/2022	9045D
07/22/2022	Solid	N3915-09	pH	Cool 4 deg C	USEP01	YBBC7		07/22/2022	9045D
07/20/2022	Solid	N3915-10	pH	Cool 4 deg C	USEP01	YBBC1		07/20/2022	9045D
07/20/2022	Solid	N3915-11	pH	Cool 4 deg C	USEP01	YBBC2		07/20/2022	9045D
07/20/2022	Solid	N3915-12	pH	Cool 4 deg C	USEP01	YBBC3		07/20/2022	9045D
07/20/2022	Solid	N3915-13	pH	Cool 4 deg C	USEP01	YBBC4		07/20/2022	9045D
07/20/2022	Solid	N3915-14	pH	Cool 4 deg C	USEP01	YBBC5		07/20/2022	9045D
07/20/2022	Solid	N3915-15	pH	Cool 4 deg C	USEP01	YBBC5D		07/20/2022	9045D
07/20/2022	Solid	N3915-16	pH	Cool 4 deg C	USEP01	YBBC5S		07/20/2022	9045D

Date/Time 07/27/22 15:25
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 07/27/22 17:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]