



Analytical Summary Report

Analysis Method: 9045D

Parameter: pH

Run Number: LB121250

BalanceID: WC SC-4

Analyst By : JIGNESH

Supervisor Review By : mohammad

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	07/27/2022	15:30
2	CAL2	1	Water	NA	NA	20.8	7.00	07/27/2022	15:31
3	CAL3	1	Water	NA	NA	20.8	10.03	07/27/2022	15:32
4	ICV	1	Water	NA	NA	20.9	6.99	07/27/2022	15:35
5	CCV1	1	Water	NA	NA	20.8	2.01	07/27/2022	16:35
6	N3917-01	1	Solid	20.02	20	22.7	6.82	07/27/2022	16:40
7	N3917-02	1	Solid	20.02	20	22.5	6.62	07/27/2022	16:42
8	N3917-03	1	Solid	20.03	20	22.3	6.69	07/27/2022	16:45
9	N3917-04	1	Solid	20.03	20	22.3	6.69	07/27/2022	16:45
10	N3917-05	1	Solid	20.03	20	22.3	6.69	07/27/2022	16:45
11	N3917-06	1	Solid	20.04	20	22.3	6.46	07/27/2022	16:50
12	N3917-07	1	Solid	20.01	20	21.7	5.79	07/27/2022	17:00
13	N3917-08	1	Solid	20.03	20	21.5	8.41	07/27/2022	17:02
14	N3917-09	1	Solid	20.04	20	21.2	6.44	07/27/2022	17:05
15	N3917-10	1	Solid	20.03	20	21.1	7.09	07/27/2022	17:10
16	CCV2	1	Water	NA	NA	20.9	12.03	07/27/2022	17:11
17	N3917-11	1	Solid	20.05	20	21.1	6.73	07/27/2022	17:15
18	CCV3	1	Water	NA	NA	20.9	2.01	07/27/2022	17:16

VB 12/250

WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-N3918 WorkList ID : 161730 Department : Wet-Chemistry Date : 07-27-2022 15:22:56

Due Date	Matrix	Sample	Test	Raw Sample			Customer Sample	Collect Date	Method
				Preservative	Customer	Storage Location			
07/22/2022	Solid	N3917-01	pH	Cool 4 deg C	USEP01	B63	YBB97	07/22/2022	9045D
07/22/2022	Solid	N3917-02	pH	Cool 4 deg C	USEP01	B63	YBB98	07/22/2022	9045D
07/22/2022	Solid	N3917-03	pH	Cool 4 deg C	USEP01	B63	YBB99	07/22/2022	9045D
07/22/2022	Solid	N3917-04	pH	Cool 4 deg C	USEP01	B63	YBB99D	07/22/2022	9045D
07/22/2022	Solid	N3917-05	pH	Cool 4 deg C	USEP01	B63	YBB99S	07/22/2022	9045D
07/22/2022	Solid	N3917-06	pH	Cool 4 deg C	USEP01	B63	YBBA0	07/22/2022	9045D
07/22/2022	Solid	N3917-07	pH	Cool 4 deg C	USEP01	B63	YBBA1	07/22/2022	9045D
07/22/2022	Solid	N3917-08	pH	Cool 4 deg C	USEP01	B63	YBBA2	07/22/2022	9045D
07/22/2022	Solid	N3917-09	pH	Cool 4 deg C	USEP01	B63	YBBA3	07/22/2022	9045D
07/22/2022	Solid	N3917-10	pH	Cool 4 deg C	USEP01	B63	YBBA4	07/22/2022	9045D
07/22/2022	Solid	N3917-11	pH	Cool 4 deg C	USEP01	B63	YBBA5	07/22/2022	9045D

Date/Time 07/27/22 15:25
 Raw Sample Received by: JP (w/c)
 Raw Sample Relinquished by: CR-SM

Date/Time 07/27/22 17:30
 Raw Sample Received by: CR-SM
 Raw Sample Relinquished by: JP (w/c)