



PERCENT SOLID

Supervisor: mohammad

Analyst: JIGNESH

Date: 7/29/2022

OVENTEMP IN Celsius(°C): 107

Time IN: 14:10

In Date: 07/28/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:17

Out Date: 07/29/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB121259

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N3926-01	EXS20	1	0.91	9.05	9.96	8.99	89.3	
N3926-02	EXS21	2	0.92	8.85	9.77	8.52	85.9	
N3926-03	EXS22	3	0.93	8.89	9.82	8.47	84.8	
N3926-04	EXS23	4	0.93	8.65	9.58	8.31	85.3	
N3926-05	EXS25	5	0.93	8.70	9.63	8.48	86.8	
N3926-06	EXS26	6	0.93	8.84	9.77	8.35	83.9	
N3926-07	EXS27	7	0.92	9.03	9.95	8.58	84.8	
N3926-08	EXS28	8	0.93	8.80	9.73	8.33	84.1	
N3926-09	EXS29	9	0.91	9.03	9.94	8.68	86.0	
N3926-10	EXS30	10	0.95	8.71	9.66	8.19	83.1	
N3926-11	EXS31	11	0.92	8.73	9.65	8.48	86.6	
N3926-12	EXS32	12	0.96	8.89	9.85	8.69	87.0	
N3926-13	EXS33	13	0.94	9.00	9.94	8.92	88.7	
N3926-14	EXS34	14	0.9	8.88	9.78	8.57	86.4	
N3926-15	EXS35	15	0.91	8.76	9.67	8.53	87.0	
N3926-16	EXS36	16	0.91	9.05	9.96	8.97	89.1	
N3926-17	EXS37	17	0.91	8.61	9.52	8.34	86.3	
N3926-18	EXS38	18	0.97	8.78	9.75	8.5	85.8	
N3926-19	EXS39	19	0.94	9.04	9.98	9.11	90.4	
N3926-20	EXS40	20	0.92	9.05	9.97	9.3	92.6	
N3926-21	EXS40D	21	0.92	9.05	9.97	9.3	92.6	
N3926-22	EXS40S	22	0.92	9.05	9.97	9.3	92.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$