

**PERCENT SOLID****Supervisor:** mohammad**Analyst:** JIGNESH**Date:** 7/29/2022**OVENTEMP IN Celsius(°C):** 107**Time IN:** 17:00**In Date:** 07/28/2022**Weight Check 1.0g:** 1.00**Weight Check 10g:** 10.00**OvenID:** M OVEN-1**OVENTEMP OUT Celsius(°C):** 103**Time OUT:** 08:30**Out Date:** 07/29/2022**Weight Check 1.0g:** 1.00**Weight Check 10g:** 10.00**BalanceID:** M SC-4**Thermometer ID:** %SOLIDS-OVEN**QC:**LB121262

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N3920-01	PB-07-27-2022	1	0.91	8.76	9.67	9.31	95.9	
N3927-01	RVE-75	2	0.91	9.05	9.96	9.59	95.9	
N3927-02	RVE-76	3	0.91	9.03	9.94	9.77	98.1	
N3927-03	RVE-77	4	0.9	9.08	9.98	9.55	95.3	
N3927-04	RVE-78	5	0.91	9.07	9.98	9.44	94.0	
N3927-05	RVE-79	6	0.9	9.01	9.91	9.8	98.8	
N3927-06	RVE-79D	7	0.92	8.90	9.82	9.8	99.8	
N3927-07	RVE-80	8	0.9	8.85	9.75	8.72	88.4	
N3927-08	RVE-81	9	0.91	9.04	9.95	9.14	91.0	
N3927-09	RVE-82	10	0.92	8.78	9.7	8.68	88.4	
N3927-10	RVE-83	11	0.91	8.81	9.72	8.89	90.6	
N3927-11	RVE-84	12	0.92	8.93	9.85	8.14	80.9	
N3927-12	RVE-85	13	0.93	9.05	9.98	8.89	88.0	
N3927-13	RVE-86	14	0.92	8.86	9.78	9.22	93.7	
N3927-14	RVE-87	15	0.91	9.05	9.96	8.52	84.1	
N3927-15	RVE-88	16	0.93	9.02	9.95	8.89	88.2	
N3927-16	RVE-89	17	0.91	9.07	9.98	8.58	84.6	
N3927-17	RVE-90	18	0.91	8.91	9.82	9.00	90.8	
N3927-18	RVE-91	19	0.91	8.89	9.8	8.47	85.0	
N3927-19	RVE-92	20	0.91	8.96	9.87	6.68	64.4	
N3927-20	RVE-93	21	0.92	8.97	9.89	8.9	89.0	
N3930-01	RVE-94	22	0.9	9.06	9.96	9.04	89.8	
N3930-02	RVE-95	23	0.91	9.01	9.92	9.06	90.5	
N3930-03	RVE-96	24	0.92	8.94	9.86	9.44	95.3	
N3930-04	RVE-97	25	0.92	8.98	9.9	9.15	91.6	
N3930-05	RVE-98	26	0.92	8.95	9.87	8.71	87.0	
N3930-06	RVE-99	27	0.92	9.06	9.98	8.89	88.0	
N3930-07	RVE-99D	28	0.91	9.03	9.94	8.9	88.5	



PERCENT SOLID

Supervisor: mohammad

Analyst: JIGNESH

Date: 7/29/2022

OVENTEMP IN Celsius(°C): 107

Time IN: 17:00

In Date: 07/28/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:30

Out Date: 07/29/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB121262

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N3930-08	RVE-100	29	0.92	9.05	9.97	8.92	88.4	
N3930-09	RVE-101	30	0.91	9.00	9.91	9.63	96.9	
N3930-10	RVE-102	31	0.92	8.88	9.8	8.87	89.5	
N3930-11	RVE-103	32	0.91	9.04	9.95	9.25	92.3	
N3930-12	RVE-104	33	0.92	8.90	9.82	8.97	90.4	
N3930-13	RVE-105	34	0.91	8.92	9.83	9.28	93.8	
N3930-14	RVE-106	35	0.91	9.07	9.98	9.07	90.0	
N3930-15	RVE-107	36	0.92	8.95	9.87	9.58	96.8	
N3930-16	RVE-108	37	0.91	9.06	9.97	9.68	96.8	
N3930-17	RVE-109	38	0.93	8.94	9.87	9.6	97.0	
N3930-18	RVE-110	39	0.9	9.08	9.98	9.06	89.9	
N3930-19	RVE-111	40	0.92	8.83	9.75	9.2	93.8	
N3930-20	RVE-112	41	0.93	9.02	9.95	9.24	92.1	
N3931-01	RVE-113	42	0.87	9.02	9.89	9.44	95.0	
N3931-02	RVE-114	43	0.88	9.09	9.97	9.1	90.4	
N3931-03	RVE-115	44	0.89	9.00	9.89	9.38	94.3	
N3931-04	RVE-116	45	0.85	8.93	9.78	9.28	94.4	
N3931-05	RVE-117	46	0.89	9.01	9.9	7.61	74.6	
N3931-06	RVE-118	47	0.85	9.12	9.97	8.98	89.1	
N3931-07	RVE-119	48	0.86	9.12	9.98	9.62	96.1	
N3931-08	RVE-119D	49	0.85	9.07	9.92	9.59	96.4	
N3931-09	RVE-120	50	0.89	9.06	9.95	9.58	95.9	
N3931-10	RVE-121	51	0.95	8.98	9.93	9.83	98.9	
N3931-11	RVE-122	52	0.86	8.94	9.8	9.39	95.4	
N3931-12	RVE-123	53	0.85	8.90	9.75	9.28	94.7	
N3931-13	RVE-124	54	0.86	9.00	9.86	9.06	91.1	
N3931-14	RVE-125	55	0.87	8.93	9.8	9.11	92.3	
N3931-15	RVE-126	56	0.89	8.93	9.82	9.1	91.9	

**PERCENT SOLID**

Supervisor: mohammad

Analyst: JIGNESH

Date: 7/29/2022

OVENTEMP IN Celsius(°C): 107

Time IN: 17:00

In Date: 07/28/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:30

Out Date: 07/29/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB121262

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N3931-16	RVE-127	57	0.92	9.06	9.98	9.13	90.6	
N3931-17	RVE-128	58	0.83	9.10	9.93	9.07	90.5	
N3931-18	RVE-129	59	0.88	9.10	9.98	9.47	94.4	
N3931-19	RVE-130	60	0.86	9.09	9.95	8.83	87.7	
N3931-20	RVE-131	61	0.87	8.95	9.82	9.09	91.8	
N3935-01	RVE-132	62	0.88	8.99	9.87	9.26	93.2	
N3935-02	RVE-133	63	0.86	9.09	9.95	8.93	88.8	
N3935-03	RVE-134	64	0.86	9.02	9.88	9.63	97.2	
N3935-04	RVE-135	65	0.83	9.13	9.96	9.49	94.9	
N3935-05	RVE-136	66	0.85	9.08	9.93	9.57	96.0	
N3935-06	RVE-137	67	0.88	9.11	9.99	9.52	94.8	
N3935-07	RVE-138	68	0.84	8.99	9.83	9.1	91.9	
N3935-08	RVE-139	69	0.85	9.12	9.97	9.22	91.8	
N3935-09	RVE-139D	70	0.85	9.02	9.87	9.21	92.7	
N3935-10	RVE-140	71	0.87	9.00	9.87	9.48	95.7	
N3935-11	RVE-141	72	0.85	9.13	9.98	9.33	92.9	
N3935-12	RVE-142	73	0.86	8.95	9.81	8.86	89.4	
N3935-13	RVE-143	74	0.87	8.93	9.8	8.67	87.3	
N3935-14	RVE-144	75	0.86	9.11	9.97	8.77	86.8	
N3935-15	RVE-145	76	0.86	9.12	9.98	9.67	96.6	
N3935-16	RVE-146	77	0.85	9.02	9.87	8.93	89.6	
N3935-17	RVE-147	78	0.88	8.97	9.85	9.11	91.8	
N3935-18	RVE-148	79	0.86	9.12	9.98	8.64	85.3	
N3945-01	S4-01-B	80	0.88	8.99	9.87	9.57	96.7	
N3945-02	S4-01-S	81	8.58	1.24	9.82	9.52	75.8	
N3945-03	DUP-4	82	0.85	9.12	9.97	9.77	97.8	
N3945-04	S4-02-S	83	2.96	7.02	9.98	9.22	89.2	
N3948-01	MHA	84	0.81	9.09	9.9	8.82	88.1	



PERCENT SOLID

Supervisor: mohammad

Analyst: JIGNESH

Date: 7/29/2022

OVENTEMP IN Celsius(°C): 107

Time IN: 17:00

In Date: 07/28/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:30

Out Date: 07/29/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB121262

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N3948-02	MHA-EPH-2	85	0.84	9.06	9.9	8.82	88.1	
N3964-01	HMGP-SOIL-A	86	0.91	8.95	9.86	7.45	73.1	
N3964-03	HMGP-SOIL-B	87	0.93	9.02	9.95	7.65	74.5	
N3964-05	HMGP-SOIL-C	88	0.91	9.06	9.97	7.85	76.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$