


LB121267

INSTRUMENT ID: CV4

7471B

 μ Abs.:

57350



Linear

A= 0.0000e+000

B= 1.7617e-004

C= -1.6807e-001

Rho= 0.9997734

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Yield
0.0	0.000	-0.005	-0.005	925	0.000	925					27
0.2	0.200	0.253	0.053	2389	0.0 %	2389					-5
2.5	2.500	2.365	-0.135	14377	0.0 %	14377					2
5.0	5.000	5.082	0.082	29800	0.0 %	29800					1
7.5	7.500	7.571	0.071	43929	0.0 %	43929					-1
10.0	10.000	9.935	-0.065	57350	0.0 %	57350					

LB121267 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 28 Jul 2022 14:06:33					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	925	-	0.0000	7471B	PPB	28 Jul 2022 14:28:32	S	Std
0.2 - 1	50.2	2389	-	0.2000	7471B	PPB	28 Jul 2022 14:30:49	S	Std
2.5 - 1	52.5	14377	-	2.5000	7471B	PPB	28 Jul 2022 14:33:06	S	Std
5.0 - 1	55	29800	-	5.0000	7471B	PPB	28 Jul 2022 14:35:23	S	Std
7.5 - 1	57.5	43929	-	7.5000	7471B	PPB	28 Jul 2022 14:40:19	S	Std
10.0 - 1	51.8	57350	-	10.0000	7471B	PPB	28 Jul 2022 14:47:55	S	Std
ICV58 - 1	ICV58	23673	4.0023	-	7471B	PPB	28 Jul 2022 14:53:21	U	SMPL
ICB58 - 1	ICB58	696	-0.0455	-	7471B	PPB	28 Jul 2022 14:55:36	U	SMPL
CCV01 - 1	CCV01	27117	4.6091	-	7471B	PPB	28 Jul 2022 14:57:51	U	SMPL
CCB01 - 1	CCB01	571	-0.0675	-	7471B	PPB	28 Jul 2022 15:00:07	U	SMPL
CRA - 1	CRA	2251	0.2285	-	7471B	PPB	28 Jul 2022 15:02:23	U	SMPL
HighStd - 1	HighStd	58641	10.1626	-	7471B	PPB	28 Jul 2022 15:04:38	U	SMPL
ChkStd - 1	ChkStd	39159	6.7305	-	7471B	PPB	28 Jul 2022 15:06:53	U	SMPL
PB146562BL - 1	PBS	654	-0.0529	-	7471B	PPB	28 Jul 2022 15:11:51	U	SMPL
PB146562BS - 1	LCSS	23597	3.9890	-	7471B	PPB	28 Jul 2022 15:16:41	U	SMPL
N3893-01 - 1	RVE-1	4823	0.6816	-	7471B	PPB	28 Jul 2022 15:18:56	U	SMPL
N3893-01DUP - 1	RVE-1DUP	5176	0.7438	-	7471B	PPB	28 Jul 2022 15:21:14	U	SMPL
N3893-01MS - 1	RVE-1MS	26410	4.4845	-	7471B	PPB	28 Jul 2022 15:23:31	U	SMPL
N3893-01MSD - 1	RVE-1MSD	26388	4.4806	-	7471B	PPB	28 Jul 2022 15:25:48	U	SMPL
N3893-02 - 1	RVE-2	2977	0.3564	-	7471B	PPB	28 Jul 2022 15:28:04	U	SMPL
CCV02 - 1	CCV02	27616	4.6970	-	7471B	PPB	28 Jul 2022 16:05:29	U	SMPL
CCB02 - 1	CCB02	760	-0.0342	-	7471B	PPB	28 Jul 2022 16:07:44	U	SMPL
N3893-03 - 1	RVE-3	2171	0.2144	-	7471B	PPB	28 Jul 2022 16:10:02	U	SMPL
N3893-04 - 1	RVE-4	4275	0.5850	-	7471B	PPB	28 Jul 2022 16:12:17	U	SMPL
N3893-05 - 1	RVE-5	2567	0.2842	-	7471B	PPB	28 Jul 2022 16:14:32	U	SMPL
N3893-06 - 1	RVE-6	3531	0.4540	-	7471B	PPB	28 Jul 2022 16:16:47	U	SMPL
N3893-07 - 1	RVE-7	2079	0.1982	-	7471B	PPB	28 Jul 2022 16:19:04	U	SMPL
N3893-08 - 1	RVE-8	1824	0.1533	-	7471B	PPB	28 Jul 2022 16:21:20	U	SMPL
N3893-09 - 1	RVE-9	2446	0.2628	-	7471B	PPB	28 Jul 2022 16:23:36	U	SMPL
N3893-10 - 1	RVE-10	3082	0.3749	-	7471B	PPB	28 Jul 2022 16:25:53	U	SMPL
N3893-11 - 1	RVE-11	2974	0.3559	-	7471B	PPB	28 Jul 2022 16:28:11	U	SMPL
N3893-12 - 1	RVE-12	2943	0.3504	-	7471B	PPB	28 Jul 2022 16:30:26	U	SMPL
CCV03 - 1	CCV03	30583	5.2197	-	7471B	PPB	28 Jul 2022 16:37:50	U	SMPL
CCB03 - 1	CCB03	705	-0.0439	-	7471B	PPB	28 Jul 2022 16:40:05	U	SMPL
N3893-13 - 1	RVE-13	2195	0.2186	-	7471B	PPB	28 Jul 2022 16:42:22	U	SMPL
N3893-14 - 1	RVE-14	2057	0.1943	-	7471B	PPB	28 Jul 2022 16:44:38	U	SMPL
N3893-15 - 1	RVE-15	2723	0.3116	-	7471B	PPB	28 Jul 2022 16:46:53	U	SMPL
N3893-16 - 1	RVE-16	2806	0.3263	-	7471B	PPB	28 Jul 2022 16:49:09	U	SMPL
N3893-17 - 1	RVE-17	3563	0.4596	-	7471B	PPB	28 Jul 2022 16:51:24	U	SMPL
N3893-18 - 1	RVE-18	6598	0.9943	-	7471B	PPB	28 Jul 2022 16:53:41	U	SMPL
N3893-19 - 1	RVE-19	2086	0.1994	-	7471B	PPB	28 Jul 2022 16:55:58	U	SMPL
N3893-20 - 1	RVE-19D	2678	0.3037	-	7471B	PPB	28 Jul 2022 16:58:15	U	SMPL
PB146563BL - 1	PBS	955	0.0002	-	7471B	PPB	28 Jul 2022 17:00:32	U	SMPL
PB146563BS - 1	LCSS	21023	3.5355	-	7471B	PPB	28 Jul 2022 17:02:47	U	SMPL
CCV04 - 1	CCV04	30235	5.1583	-	7471B	PPB	28 Jul 2022 17:09:06	U	SMPL
CCB04 - 1	CCB04	1020	0.0116	-	7471B	PPB	28 Jul 2022 17:11:21	U	SMPL
N3894-01 - 1	RVE-20	2627	0.2947	-	7471B	PPB	28 Jul 2022 17:21:11	U	SMPL
N3894-01DUP - 1	RVE-20DUP	2593	0.2887	-	7471B	PPB	28 Jul 2022 17:23:26	U	SMPL
N3894-01MS - 1	RVE-20MS	23535	3.9780	-	7471B	PPB	28 Jul 2022 17:25:41	U	SMPL
N3894-01MSD - 1	RVE-20MSD	24162	4.0885	-	7471B	PPB	28 Jul 2022 17:30:32	U	SMPL
N3894-02 - 1	RVE-21	4489	0.6227	-	7471B	PPB	28 Jul 2022 17:32:48	U	SMPL
N3894-03 - 1	RVE-22	2074	0.1973	-	7471B	PPB	28 Jul 2022 17:35:05	U	SMPL
N3894-04 - 1	RVE-23	3692	0.4823	-	7471B	PPB	28 Jul 2022 17:37:21	U	SMPL
N3894-05 - 1	RVE-24	2234	0.2255	-	7471B	PPB	28 Jul 2022 17:39:38	U	SMPL
N3894-06 - 1	RVE-25	3443	0.4385	-	7471B	PPB	28 Jul 2022 17:41:55	U	SMPL
N3894-07 - 1	RVE-26	1737	0.1379	-	7471B	PPB	28 Jul 2022 17:44:11	U	SMPL
CCV05 - 1	CCV05	29866	5.0933	-	7471B	PPB	28 Jul 2022 17:46:27	U	SMPL
CCB05 - 1	CCB05	940	-0.0025	-	7471B	PPB	28 Jul 2022 17:48:42	U	SMPL
N3894-08 - 1	RVE-27	3804	0.5021	-	7471B	PPB	28 Jul 2022 17:51:00	U	SMPL
N3894-09 - 1	RVE-28	7696	1.1877	-	7471B	PPB	28 Jul 2022 17:53:14	U	SMPL
N3894-10 - 1	RVE-29	5727	0.8408	-	7471B	PPB	28 Jul 2022 17:55:31	U	SMPL
N3894-11 - 1	RVE-30	4890	0.6934	-	7471B	PPB	28 Jul 2022 17:57:47	U	SMPL
N3894-12 - 1	RVE-31	3863	0.5125	-	7471B	PPB	28 Jul 2022 18:00:03	U	SMPL
N3894-13 - 1	RVE-32	3337	0.4198	-	7471B	PPB	28 Jul 2022 18:02:19	U	SMPL
N3894-14 - 1	RVE-33	2287	0.2348	-	7471B	PPB	28 Jul 2022 18:04:36	U	SMPL
N3894-15 - 1	RVE-34	5020	0.7163	-	7471B	PPB	28 Jul 2022 18:06:53	U	SMPL
N3894-16 - 1	RVE-35	5712	0.8382	-	7471B	PPB	28 Jul 2022 18:09:11	U	SMPL
N3894-17 - 1	RVE-36	1773	0.1443	-	7471B	PPB	28 Jul 2022 18:11:28	U	SMPL
CCV06 - 1	CCV06	28757	4.8980	-	7471B	PPB	28 Jul 2022 18:13:44	U	SMPL
CCB06 - 1	CCB06	929	-0.0044	-	7471B	PPB	28 Jul 2022 18:16:00	U	SMPL
N3894-18 - 1	RVE-37	4598	0.6420	-	7471B	PPB	28 Jul 2022 18:18:17	U	SMPL
N3894-19 - 1	RVE-38	2871	0.3377	-	7471B	PPB	28 Jul 2022 18:20:34	U	SMPL
N3894-20 - 1	RVE-39	4124	0.5584	-	7471B	PPB	28 Jul 2022 18:22:50	U	SMPL

LB121267 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 28 Jul 2022 14:06:33

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
N3893-01LX5 - 1		981	0.0048		-7471B	PPB	28 Jul 2022 18:25:05	U	SMPL
N3893-01A - 1		28620	4.8738		-7471B	PPB	28 Jul 2022 18:27:21	U	SMPL
N3894-01LX5 - 1		454	-0.0881		-7471B	PPB	28 Jul 2022 18:29:38	U	SMPL
N3894-01A - 1		26595	4.5171		-7471B	PPB	28 Jul 2022 18:31:56	U	SMPL
CCV07 - 1	CCV07	31373	5.3588		-7471B	PPB	28 Jul 2022 18:34:13	U	SMPL
CCB07 - 1	CCB07	1037	0.0146		-7471B	PPB	28 Jul 2022 18:36:31	U	SMPL