



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB121273

Analyst By : JIGNESH

Supervisor Review By : mohammad

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	07/28/2022	16:00
2	CAL2	1	Water	NA	NA	20.7	7.00	07/28/2022	16:01
3	CAL3	1	Water	NA	NA	20.7	10.03	07/28/2022	16:02
4	ICV	1	Water	NA	NA	20.8	6.98	07/28/2022	16:05
5	CCV1	1	Water	NA	NA	20.7	2.01	07/28/2022	16:10
6	N3953-01	1	Water	NA	NA	24.4	6.48	07/28/2022	16:11
7	N3953-01DUP	1	Water	NA	NA	24.5	6.49	07/28/2022	16:12
8	N3957-01	1	Water	NA	NA	23.4	5.12	07/28/2022	16:15
9	N3958-02	1	Water	NA	NA	22.1	6.90	07/28/2022	16:20
10	N3964-07	1	Water	NA	NA	25.4	5.63	07/28/2022	16:22
11	N3964-08	1	Water	NA	NA	25.7	9.16	07/28/2022	16:25
12	CCV2	1	Water	NA	NA	20.7	12.03	07/28/2022	16:30

WORKLIST(Hardcopy Internal Chain)

121273

WorkList Name : PH N3909

WorkList ID : 161768

Department : Wet-Chemistry

Date : 07-28-2022 13:16:45

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/28/2022	Water	N3953-01	pH	Cool 4 deg C	PSEG03	G11	IRV-OILY-WATER	07/28/2022	9040C
07/22/2022	Water	N3957-01	pH	Cool 4 deg C	STER14	N11	MINI-CELL	07/22/2022	9040C
07/27/2022	Water	N3958-02	pH	Cool 4 deg C	TETR06	N11	RW9-IDWGW-20220727	07/27/2022	9040C
07/28/2022	Water	N3964-07	pH	Cool 4 deg C	PSEG03	G11	HMGP-WATER-1	07/28/2022	9040C
07/28/2022	Water	N3964-08	pH	Cool 4 deg C	PSEG03	G11	HMGP-WATER-2	07/28/2022	9040C

Date/Time 07/28/22 15:15
 Raw Sample Received by: JD (CSCM)
 Raw Sample Relinquished by: JD (CSCM)

Date/Time 07/28/22 17:30
 Raw Sample Received by: JD (CSCM)
 Raw Sample Relinquished by: JD (CSCM)