



PERCENT SOLID

Supervisor: mohammad

Analyst: JIGNESH

Date: 7/30/2022

OVENTEMP IN Celsius(°C): 107

Time IN: 14:00

In Date: 07/29/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:00

Out Date: 07/30/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB121277

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N3932-01	EXS88	1	0.91	9.05	9.96	8.77	86.9	
N3932-02	EXS89	2	0.93	9.02	9.95	8.68	85.9	
N3932-03	EXS90	3	0.91	9.02	9.93	8.62	85.5	
N3932-04	EXS91	4	0.91	8.83	9.74	8.26	83.2	
N3932-05	EXS92	5	0.9	8.69	9.59	8.46	87.0	
N3932-06	EXS93	6	0.92	8.58	9.5	8.24	85.3	
N3932-07	EXS94	7	0.91	8.72	9.63	8.42	86.1	
N3932-08	EXS95	8	0.91	8.69	9.6	8.56	88.0	
N3932-09	EXS96	9	0.91	8.79	9.7	8.86	90.4	
N3932-10	EXS97	10	0.92	9.06	9.98	8.52	83.9	
N3932-11	EXS98	11	0.91	8.64	9.55	7.88	80.7	
N3932-12	EXS99	12	0.95	8.67	9.62	8.1	82.5	
N3932-13	EXSA0	13	0.91	8.75	9.66	8.17	83.0	
N3932-14	EXSA1	14	0.95	8.63	9.58	8.14	83.3	
N3932-15	EXSA2	15	0.93	8.79	9.72	8.33	84.2	
N3932-16	EXSA3	16	0.9	8.81	9.71	8.46	85.8	
N3932-17	EXSA4	17	0.95	8.87	9.82	8.05	80.0	
N3932-18	EXSA5	18	0.93	8.64	9.57	8.27	85.0	
N3932-19	EXSA6	19	0.94	8.99	9.93	8.52	84.3	
N3932-20	EXSB0	20	0.95	9.01	9.96	8.3	81.6	
N3932-21	EXSB0D	21	0.95	9.01	9.96	8.3	81.6	
N3932-22	EXSB0S	22	0.95	9.01	9.96	8.3	81.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$