


LB121283

7471B

INSTRUMENT ID: CVL

 μ Abs.:

53133



A= 0.0000e+000

B= 1.8550e-004

C= 3.2479e-002

Rho= 0.9997016

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% D
0.0	0.000	0.033	0.033	5	0.000	5					-8
0.2	0.200	0.184	-0.016	817	0.0 %	817					-5
2.5	2.500	2.387	-0.113	12695	0.0 %	12695					2
5.0	5.000	5.080	0.080	27209	0.0 %	27209					2
7.5	7.500	7.626	0.126	40937	0.0 %	40937					-1
10.0	10.000	9.889	-0.111	53133	0.0 %	53133					

LB121283 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 29 Jul 2022 14:23:54				Date	Type	Type
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units				
0.0 - 1	50	5	-	0.00007471B	PPB		29 Jul 2022	14:48:19	S	Std
0.2 - 1	50-2	817	-	0.20007471B	PPB		29 Jul 2022	14:50:36	S	Std
2.5 - 1	52-5	12695	-	2.50007471B	PPB		29 Jul 2022	14:52:53	S	Std
5.0 - 1	55	27209	-	5.00007471B	PPB		29 Jul 2022	14:55:10	S	Std
7.5 - 1	57-5	40937	-	7.50007471B	PPB		29 Jul 2022	14:57:26	S	Std
10.0 - 1	511	53133	-	10.00007471B	PPB		29 Jul 2022	15:02:18	S	Std
ICV60 - 1	ICV60	22046	4.1221	-7471B	PPB		29 Jul 2022	15:05:09	U	SMPL
ICB60 - 1	ICB60	-59	0.0215	-7471B	PPB		29 Jul 2022	15:07:23	U	SMPL
CCV11 - 1	CCV11	26691	4.9838	-7471B	PPB		29 Jul 2022	15:09:39	U	SMPL
CCB11 - 1	CCB11	-43	0.0245	-7471B	PPB		29 Jul 2022	15:11:54	U	SMPL
CRA - 1	CRA	1037	0.2248	-7471B	PPB		29 Jul 2022	15:14:10	U	SMPL
HighStd - 1	HighStd	53157	9.8933	-7471B	PPB		29 Jul 2022	15:16:25	U	SMPL
ChkStd - 1	ChkStd	35474	6.6130	-7471B	PPB		29 Jul 2022	15:18:40	U	SMPL
PB146599BL - 1	PBS	-80	0.0176	-7471B	PPB		29 Jul 2022	15:20:58	U	SMPL
PB146599BS - 1	LCSS	19612	3.6706	-7471B	PPB		29 Jul 2022	15:23:16	U	SMPL
N3895-01 - 1	RVE-39D	2813	0.5543	-7471B	PPB		29 Jul 2022	15:25:32	U	SMPL
N3895-01DUP - 1	RVE-39DDUP	3447	0.6719	-7471B	PPB		29 Jul 2022	15:27:49	U	SMPL
N3895-01MS - 1	RVE-39DMS	24106	4.5042	-7471B	PPB		29 Jul 2022	15:30:06	U	SMPL
N3895-01MSD - 1	RVE-39DMSD	23907	4.4673	-7471B	PPB		29 Jul 2022	15:32:23	U	SMPL
N3895-02 - 1	RVE-40	1766	0.3601	-7471B	PPB		29 Jul 2022	15:34:39	U	SMPL
CCV12 - 1	CCV12	26775	4.9993	-7471B	PPB		29 Jul 2022	15:36:55	U	SMPL
CCB12 - 1	CCB12	-38	0.0254	-7471B	PPB		29 Jul 2022	15:39:10	U	SMPL
N3895-03 - 1	RVE-41	3778	0.7333	-7471B	PPB		29 Jul 2022	15:41:26	U	SMPL
N3895-04 - 1	RVE-42	2783	0.5487	-7471B	PPB		29 Jul 2022	15:43:42	U	SMPL
N3895-05 - 1	RVE-43	3389	0.6612	-7471B	PPB		29 Jul 2022	15:45:58	U	SMPL
N3895-06 - 1	RVE-44	2730	0.5389	-7471B	PPB		29 Jul 2022	15:48:14	U	SMPL
N3895-07 - 1	RVE-45	1627	0.3343	-7471B	PPB		29 Jul 2022	15:50:29	U	SMPL
N3895-08 - 1	RVE-46	541	0.1328	-7471B	PPB		29 Jul 2022	15:52:46	U	SMPL
N3895-09 - 1	RVE-47	536	0.1319	-7471B	PPB		29 Jul 2022	15:55:03	U	SMPL
N3895-10 - 1	RVE-48	722	0.1664	-7471B	PPB		29 Jul 2022	15:57:20	U	SMPL
N3895-11 - 1	RVE-49	1485	0.3080	-7471B	PPB		29 Jul 2022	15:59:37	U	SMPL
N3895-12 - 1	RVE-50	2541	0.5038	-7471B	PPB		29 Jul 2022	16:01:52	U	SMPL
CCV13 - 1	CCV13	26092	4.8726	-7471B	PPB		29 Jul 2022	16:04:08	U	SMPL
CCB13 - 1	CCB13	-46	0.0239	-7471B	PPB		29 Jul 2022	16:06:24	U	SMPL
N3895-13 - 1	RVE-51	1342	0.2814	-7471B	PPB		29 Jul 2022	16:08:41	U	SMPL
N3895-14 - 1	RVE-52	3194	0.6250	-7471B	PPB		29 Jul 2022	16:10:57	U	SMPL
N3895-15 - 1	RVE-53	6450	1.2290	-7471B	PPB		29 Jul 2022	16:13:12	U	SMPL
N3895-16 - 1	RVE-54	5466	1.0464	-7471B	PPB		29 Jul 2022	16:15:28	U	SMPL
N3895-17 - 1	RVE-55	2042	0.4113	-7471B	PPB		29 Jul 2022	16:17:44	U	SMPL
N3895-18 - 1	RVE-56	835	0.1874	-7471B	PPB		29 Jul 2022	16:20:00	U	SMPL
N3895-19 - 1	RVE-57	5589	1.0693	-7471B	PPB		29 Jul 2022	16:22:17	U	SMPL
N3895-20 - 1	RVE-58	9050	1.7113	-7471B	PPB		29 Jul 2022	16:24:34	U	SMPL
PB146600BL - 1	PBS	0	0.0325	-7471B	PPB		29 Jul 2022	16:26:52	U	SMPL
PB146600BS - 1	LCSS	20672	3.8672	-7471B	PPB		29 Jul 2022	16:29:09	U	SMPL
CCV14 - 1	CCV14	26840	5.0114	-7471B	PPB		29 Jul 2022	16:31:25	U	SMPL
CCB14 - 1	CCB14	-48	0.0236	-7471B	PPB		29 Jul 2022	16:33:41	U	SMPL
N3896-01 - 1	RVE-59	1329	0.2790	-7471B	PPB		29 Jul 2022	16:35:58	U	SMPL
N3896-01DUP - 1	RVE-59DUP	1329	0.2790	-7471B	PPB		29 Jul 2022	16:46:14	U	SMPL
N3896-01MS - 1	RVE-59MS	20068	3.7552	-7471B	PPB		29 Jul 2022	16:51:08	U	SMPL
N3896-01MSD - 1	RVE-59MSD	22172	4.1455	-7471B	PPB		29 Jul 2022	16:57:27	U	SMPL
N3896-02 - 1	RVE-59D	948	0.2083	-7471B	PPB		29 Jul 2022	16:59:43	U	SMPL
N3896-03 - 1	RVE-60	2716	0.5363	-7471B	PPB		29 Jul 2022	17:02:00	U	SMPL
N3896-04 - 1	RVE-61	3631	0.7060	-7471B	PPB		29 Jul 2022	17:04:16	U	SMPL
N3896-05 - 1	RVE-62	1446	0.3007	-7471B	PPB		29 Jul 2022	17:06:33	U	SMPL
N3896-06 - 1	RVE-63	975	0.2133	-7471B	PPB		29 Jul 2022	17:08:51	U	SMPL
N3896-07 - 1	RVE-64	2648	0.5237	-7471B	PPB		29 Jul 2022	17:11:06	U	SMPL
CCV15 - 1	CCV15	27524	5.1383	-7471B	PPB		29 Jul 2022	17:13:22	U	SMPL
CCB15 - 1	CCB15	-39	0.0252	-7471B	PPB		29 Jul 2022	17:15:37	U	SMPL
N3896-08 - 1	RVE-65	2358	0.4699	-7471B	PPB		29 Jul 2022	17:17:54	U	SMPL
N3896-09 - 1	RVE-66	1825	0.3710	-7471B	PPB		29 Jul 2022	17:20:10	U	SMPL
N3896-10 - 1	RVE-67	978	0.2139	-7471B	PPB		29 Jul 2022	17:22:26	U	SMPL
N3896-11 - 1	RVE-68	4261	0.8229	-7471B	PPB		29 Jul 2022	17:24:41	U	SMPL
N3896-12 - 1	RVE-69	4005	0.7754	-7471B	PPB		29 Jul 2022	17:26:57	U	SMPL
N3896-13 - 1	RVE-70	830	0.1864	-7471B	PPB		29 Jul 2022	17:29:13	U	SMPL
N3896-14 - 1	RVE-71	2873	0.5654	-7471B	PPB		29 Jul 2022	17:31:30	U	SMPL
N3896-15 - 1	RVE-72	1075	0.2319	-7471B	PPB		29 Jul 2022	17:33:47	U	SMPL
N3896-16 - 1	RVE-73	1317	0.2768	-7471B	PPB		29 Jul 2022	17:36:04	U	SMPL
N3896-17 - 1	RVE-74	1361	0.2849	-7471B	PPB		29 Jul 2022	17:38:20	U	SMPL
CCV16 - 1	CCV16	27721	5.1748	-7471B	PPB		29 Jul 2022	17:40:36	U	SMPL
CCB16 - 1	CCB16	-32	0.0265	-7471B	PPB		29 Jul 2022	17:42:52	U	SMPL
PB146601BL - 1	PBS	13	0.0349	-7471B	PPB		29 Jul 2022	17:45:09	U	SMPL
PB146601BS - 1	LCSS	20245	3.7880	-7471B	PPB		29 Jul 2022	17:47:26	U	SMPL
N3920-01 - 1	PB-07-27-2022	1367	0.2861	-7471B	PPB		29 Jul 2022	17:49:42	U	SMPL
N3920-01DUP - 1	PB-07-27-2022DUP	1337	0.2805	-7471B	PPB		29 Jul 2022	17:51:58	U	SMPL
N3920-01MS - 1	PB-07-27-2022MS	21972	4.1084	-7471B	PPB		29 Jul 2022	17:54:14	U	SMPL
N3920-01MSD - 1	PB-07-27-2022MSD	21920	4.0987	-7471B	PPB		29 Jul 2022	17:56:30	U	SMPL
N3923-01 - 1	OK-03-072722	280	0.0844	-7471B	PPB		29 Jul 2022	17:58:47	U	SMPL
N3924-01 - 1	HD-1-072722	1968	0.3975	-7471B	PPB		29 Jul 2022	18:01:05	U	SMPL
N3925-01 - 1	JC-03-072722-A	8261	1.5649	-7471B	PPB		29 Jul 2022	18:03:22	U	SMPL
N3925-04 - 1	JC-03-072722-B	6048	1.1544	-7471B	PPB		29 Jul 2022	18:05:38	U	SMPL
CCV17 - 1	CCV17	27362	5.1082	-7471B	PPB		29 Jul 2022	18:07:55	U	SMPL

LB121283 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 29 Jul 2022 14:23:54

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
CCB17 - 1	CCB17	-50	0.0232		-7471B	PPB	29 Jul 2022 18:10:11	U	SMPL
N3942-02 - 1	72-11977	3255	0.6363		-7471B	PPB	29 Jul 2022 18:12:29	U	SMPL
N3944-01 - 1	RT2286	806	0.1820		-7471B	PPB	29 Jul 2022 18:14:44	U	SMPL
N3948-01 - 1	MHA	128420	23.8548		-7471B	PPB	29 Jul 2022 18:17:01	U	SMPL
N3964-01 - 1	HMGP-SOIL-A	3677	0.7146		-7471B	PPB	29 Jul 2022 18:19:17	U	SMPL
N3964-03 - 1	HMGP-SOIL-B	3847	0.7461		-7471B	PPB	29 Jul 2022 18:21:36	U	SMPL
N3964-05 - 1	HMGP-SOIL-C	604	0.1445		-7471B	PPB	29 Jul 2022 18:23:52	U	SMPL
CCV18 - 1	CCV18	27879	5.2041		-7471B	PPB	29 Jul 2022 18:26:09	U	SMPL
CCB18 - 1	CCB18	-44	0.0243		-7471B	PPB	29 Jul 2022 18:28:26	U	SMPL
N3895-01LX5 - 1		561	0.1365		-7471B	PPB	29 Jul 2022 18:30:44	U	SMPL
N3895-01A - 1		24898	4.6511		-7471B	PPB	29 Jul 2022 18:33:00	U	SMPL
N3896-01LX5 - 1		253	0.0794		-7471B	PPB	29 Jul 2022 18:35:16	U	SMPL
N3896-01A - 1		23462	4.3848		-7471B	PPB	29 Jul 2022 18:37:33	U	SMPL
N3920-01LX5 - 1		129	0.0564		-7471B	PPB	29 Jul 2022 18:39:49	U	SMPL
N3920-01A - 1		22793	4.2607		-7471B	PPB	29 Jul 2022 18:42:06	U	SMPL
N3948-01DLX5 - 1		23358	4.3655		-7471B	PPB	29 Jul 2022 18:50:02	U	SMPL
CCV19 - 1	CCV19	26940	5.0299		-7471B	PPB	29 Jul 2022 18:52:19	U	SMPL
CCB19 - 1	CCB19	-74	0.0188		-7471B	PPB	29 Jul 2022 18:54:37	U	SMPL