



PERCENT SOLID

Supervisor: mohammad

Analyst: JIGNESH

Date: 7/30/2022

OVENTEMP IN Celsius(°C): 107

Time IN: 16:10

In Date: 07/29/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:10

Out Date: 07/30/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB121284

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N3934-01	EXSC7	1	0.9	8.81	9.71	8.67	88.2	
N3934-02	EXSC8	2	0.93	8.65	9.58	8.38	86.1	
N3934-03	EXSC9	3	0.93	8.61	9.54	8.45	87.3	
N3934-04	EXSD0	4	0.9	8.68	9.58	7.88	80.4	
N3934-05	EXSD1	5	0.91	8.95	9.86	8.32	82.8	
N3934-06	EXSD2	6	0.96	8.68	9.64	8.29	84.4	
N3934-07	EXSD4	7	0.91	8.99	9.9	8.12	80.2	
N3934-08	EXSD5	8	0.98	8.78	9.76	8.04	80.4	
N3934-09	EXSD6	9	0.91	8.77	9.68	8.2	83.1	
N3934-10	EXSD7	10	0.93	8.65	9.58	7.75	78.8	
N3934-11	EXSD8	11	0.92	8.98	9.9	8.66	86.2	
N3934-12	EXSD9	12	0.91	8.67	9.58	8.15	83.5	
N3934-13	EXSE0	13	0.94	9.01	9.95	8.57	84.7	
N3934-14	EXSE1	14	0.93	9.01	9.94	8.79	87.2	
N3934-15	EXSE2	15	0.91	8.73	9.64	8.45	86.4	
N3934-16	EXSF2	16	0.95	9.01	9.96	8.94	88.7	
N3934-17	EXSF3	17	0.93	8.84	9.77	8.49	85.5	
N3934-18	EXSF4	18	0.91	8.94	9.85	8.7	87.1	
N3934-19	EXSF5	19	0.91	8.86	9.77	8.68	87.7	
N3934-20	EXSF5D	20	0.91	8.86	9.77	8.68	87.7	
N3934-21	EXSF5S	21	0.91	8.86	9.77	8.68	87.7	
N3934-22	EXSF6	22	0.93	8.65	9.58	8.42	86.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$