



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB121288

Analyst By : JIGNESH

Supervisor Review By : mohammad

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.9	4.01	07/29/2022	16:25
2	CAL2	1	Water	NA	NA	20.9	7.00	07/29/2022	16:26
3	CAL3	1	Water	NA	NA	20.9	10.03	07/29/2022	16:27
4	ICV	1	Water	NA	NA	20.8	6.99	07/29/2022	16:30
5	CCV1	1	Water	NA	NA	20.8	2.01	07/29/2022	16:32
6	N3977-01	1	Water	NA	NA	26.4	8.75	07/29/2022	16:35
7	N3977-01DUP	1	Water	NA	NA	26.5	8.76	07/29/2022	16:38
8	N3981-01	1	Water	NA	NA	25.2	9.42	07/29/2022	16:40
9	CCV2	1	Water	NA	NA	20.9	12.03	07/29/2022	16:42

121288

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N3977 WorkList ID : 161823 Department : Wet-Chemistry Date : 07-29-2022 16:22:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/29/2022	Water	N3977-01	pH ✓	Cool 4 deg C	PSEG03	G11	MS-TOTE-0729	07/29/2022	9040C
07/29/2022	Water	N3981-01	pH ✓	Cool 4 deg C	PSEG03	G11	PAT-TOTES-0729	07/29/2022	9040C

Date/Time 07/29/22 16:22
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 07/29/22 18:00
 Raw Sample Received by: F.C. Smith
 Raw Sample Relinquished by: [Signature]