



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : mohammad

Run Number: LB121289

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.9	4.01	07/29/2022	16:25
2	CAL2	1	Water	NA	NA	20.9	7.00	07/29/2022	16:26
3	CAL3	1	Water	NA	NA	20.9	10.03	07/29/2022	16:27
4	ICV	1	Water	NA	NA	20.8	6.99	07/29/2022	16:30
5	CCV1	1	Water	NA	NA	20.8	2.01	07/29/2022	16:32
6	N3976-02	1	Solid	20.03	20	22.2	5.37	07/29/2022	16:45
7	N3976-02DUP	1	Solid	20.04	20	22.3	5.38	07/29/2022	16:46
8	N3976-04	1	Solid	20.03	20	22.5	7.86	07/29/2022	16:50
9	N3976-06	1	Solid	20.04	20	21.9	7.01	07/29/2022	16:53
10	CCV2	1	Water	NA	NA	20.8	12.03	07/29/2022	17:00

LB121289

WORKLIST(Hardcopy Internal Chain)

WorkList Name: corrosivity n3976 WorkList ID: 161824 Department: Wet-Chemistry Date: 07-29-2022 16:23:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/29/2022	Solid	N3976-02	Corrosivity	Cool 4 deg C	PSEG03	G11	TP1-0729	07/29/2022	9045D
07/29/2022	Solid	N3976-04	Corrosivity	Cool 4 deg C	PSEG03	G11	TP2-0729	07/29/2022	9045D
07/29/2022	Solid	N3976-06	Corrosivity	Cool 4 deg C	PSEG03	G11	TP3-0729	07/29/2022	9045D

Date/Time 07/29/22 16:25
 Raw Sample Received by: J. C. Smith
 Raw Sample Relinquished by: J. C. Smith

Date/Time 07/29/22 18:00
 Raw Sample Received by: J. C. Smith
 Raw Sample Relinquished by: J. C. Smith