



PERCENT SOLID

Supervisor: mohammad

Analyst: JIGNESH

Date: 8/1/2022

OVENTEMP IN Celsius(°C): 107

Time IN: 14:30

In Date: 07/30/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:20

Out Date: 07/31/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB121296

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N3938-01	EXSK7	1	0.93	8.93	9.86	8.5	84.8	
N3938-02	EXSK8	2	0.93	8.97	9.9	8.58	85.3	
N3938-03	EXSK9	3	0.9	8.63	9.53	8.7	90.4	
N3938-04	EXSL0	4	0.91	8.83	9.74	8.94	90.9	
N3938-05	EXSL1	5	0.92	8.83	9.75	8.99	91.4	
N3938-06	EXSL2	6	0.93	8.69	9.62	8.87	91.4	
N3938-07	EXSL2D	7	0.93	8.69	9.62	8.87	91.4	
N3938-08	EXSL2S	8	0.93	8.69	9.62	8.87	91.4	
N3938-09	EXSL3	9	0.91	9.05	9.96	8.11	79.6	
N3938-10	EXSL4	10	0.91	8.65	9.56	8.00	82.0	
N3938-11	EXSL5	11	0.92	8.68	9.6	8.42	86.4	
N3938-12	EXSL6	12	0.93	8.73	9.66	8.48	86.5	
N3938-13	EXSL7	13	0.91	8.86	9.77	8.8	89.1	
N3938-14	EXSL8	14	0.87	8.90	9.77	8.84	89.6	
N3938-15	EXSL9	15	0.92	9.06	9.98	8.47	83.3	
N3938-16	EXSM0	16	0.94	8.99	9.93	8.65	85.8	
N3938-17	EXSM1	17	0.97	8.69	9.66	8.5	86.7	
N3938-18	EXSM2	18	0.81	8.90	9.71	8.33	84.5	
N3938-19	EXSM3	19	0.92	8.64	9.56	8.29	85.3	
N3938-20	EXSM4	20	0.94	8.83	9.77	8.48	85.4	
N3938-21	EXSM5	21	0.95	8.70	9.65	8.62	88.2	
N3938-22	EXSM6	22	0.91	9.07	9.98	8.59	84.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$