

LB121304


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

43018



A= 0.0000e+000

B= 2.3200e-004

C= 4.9286e-002

Rho= 0.9994166

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	610D
0.0	0.000	0.061	0.061	50	0.000	50					—
0.2	0.200	0.220	0.020	737	0.0 %	737					10
2.5	2.500	2.550	0.050	10778	0.0 %	10778					2
5.0	5.000	4.726	-0.274	20159	0.0 %	20159					-5
7.5	7.500	7.613	0.113	32604	0.0 %	32604					2
10.0	10.000	10.029	0.029	43018	0.0 %	43018					0

LB121304 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 01 Aug 2022 10:20:25

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	50	-	0.00007470A	PPB		01 Aug 2022 10:49:53	S	Std
0.2 - 1	50.2	737	-	0.20007470A	PPB		01 Aug 2022 10:52:11	S	Std
2.5 - 1	52.5	10778	-	2.50007470A	PPB		01 Aug 2022 10:54:28	S	Std
5.0 - 1	55	20159	-	5.00007470A	PPB		01 Aug 2022 10:56:45	S	Std
7.5 - 1	57.5	32604	-	7.50007470A	PPB		01 Aug 2022 10:59:02	S	Std
10.0 - 1	518	43018	-	10.00007470A	PPB		01 Aug 2022 11:01:18	S	Std
ICV61 - 1	ICV61	17363	4.0775	-7470A	PPB		01 Aug 2022 11:04:32	U	SMPL
ICB61 - 1	ICB61	-4	0.0484	-7470A	PPB		01 Aug 2022 11:06:47	U	SMPL
CCV20 - 1	CCV20	19951	4.6779	-7470A	PPB		01 Aug 2022 11:09:03	U	SMPL
CCB20 - 1	CCB20	7	0.0509	-7470A	PPB		01 Aug 2022 11:11:17	U	SMPL
CRA - 1	CRA	750	0.2233	-7470A	PPB		01 Aug 2022 11:13:32	U	SMPL
HighStd - 1	HighStd	43341	10.1044	-7470A	PPB		01 Aug 2022 11:15:47	U	SMPL
ChkStd - 1	ChkStd	30379	7.0972	-7470A	PPB		01 Aug 2022 11:18:03	U	SMPL
PB146604BL - 1	PBW	-20	0.0446	-7470A	PPB		01 Aug 2022 11:20:20	U	SMPL
PB146604BS - 1	LCSW	15527	3.6516	-7470A	PPB		01 Aug 2022 11:22:38	U	SMPL
N3858-05 - 1	SB-A	31	0.0565	-7470A	PPB		01 Aug 2022 11:24:54	U	SMPL
N3858-06 - 1	SB-B	53	0.0616	-7470A	PPB		01 Aug 2022 11:27:10	U	SMPL
N3920-02 - 1	PB-07-27-2022	76	0.0669	-7470A	PPB		01 Aug 2022 11:29:27	U	SMPL
N3948-03 - 1	MHA	47	0.0602	-7470A	PPB		01 Aug 2022 11:31:44	U	SMPL
N3964-02 - 1	HMGP-SOIL-A	56	0.0623	-7470A	PPB		01 Aug 2022 11:34:00	U	SMPL
CCV21 - 1	CCV21	20422	4.7872	-7470A	PPB		01 Aug 2022 11:36:16	U	SMPL
CCB21 - 1	CCB21	5	0.0504	-7470A	PPB		01 Aug 2022 11:38:31	U	SMPL
N3964-04 - 1	HMGP-SOIL-B	72	0.0660	-7470A	PPB		01 Aug 2022 11:40:46	U	SMPL
N3964-06 - 1	HMGP-SOIL-C	61	0.0634	-7470A	PPB		01 Aug 2022 11:45:35	U	SMPL
N3970-03 - 1	MHB-UPPER	41	0.0588	-7470A	PPB		01 Aug 2022 11:47:50	U	SMPL
N3970-06 - 1	MHB-LOWER	29	0.0560	-7470A	PPB		01 Aug 2022 11:50:06	U	SMPL
N3970-06DUP - 1	MHB-LOWERDUP	0	0.0493	-7470A	PPB		01 Aug 2022 11:52:20	U	SMPL
N3970-06DUP - 1	MHB-LOWERDUP	83	0.0685	-7470A	PPB		01 Aug 2022 11:52:58	U	SMPL
N3970-06MS - 1	MHB-LOWERMS	15612	3.6713	-7470A	PPB		01 Aug 2022 11:55:14	U	SMPL
N3970-06MSD - 1	MHB-LOWERMSD	14626	3.4425	-7470A	PPB		01 Aug 2022 12:00:04	U	SMPL
PB146544TB - 1	PB146544TB	3	0.0500	-7470A	PPB		01 Aug 2022 12:02:20	U	SMPL
N3970-06LX5 - 1		2	0.0498	-7470A	PPB		01 Aug 2022 12:04:38	U	SMPL
N3970-06A - 1		17708	4.1575	-7470A	PPB		01 Aug 2022 12:06:53	U	SMPL
CCV22 - 1	CCV22	20360	4.7728	-7470A	PPB		01 Aug 2022 12:09:08	U	SMPL
CCB22 - 1	CCB22	6	0.0507	-7470A	PPB		01 Aug 2022 12:11:25	U	SMPL