



PERCENT SOLID

Supervisor: mohammad

Analyst: JIGNESH

Date: 8/2/2022

OVENTEMP IN Celsius(°C): 107

Time IN: 17:00

In Date: 08/01/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:15

Out Date: 08/02/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB121315

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N3974-15	FLORISIL-CHECK	1	1.00	1.00	2.00	2.00	100.0	
N3987-01	080122	9	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
N3988-01	63022-B	10	0.87	9.01	9.88	9.83	99.4	
N3988-02	63022-C	11	0.89	9.02	9.91	9.9	99.9	
N3988-03	63022-D	12	0.89	9.07	9.96	9.94	99.8	
N3988-04	63022-E	13	0.89	9.09	9.98	9.94	99.6	
N3989-01	OK-04-080122	2	0.91	8.71	9.62	8.64	88.7	
N3989-02	OK-04-080122-EPH-2	3	0.91	8.71	9.62	8.64	88.7	
N3990-01	HR-01-080122	14	0.91	8.86	9.77	8.94	90.6	
N3990-02	HR-01-080122-EPH-2	15	0.92	8.88	9.8	9.09	92.0	
N3990-03	HR-04-080122	16	0.86	8.99	9.85	8.93	89.8	
N3990-04	HR-04-080122-EPH-2	17	0.89	8.88	9.77	9.27	94.4	
N3991-01	CL-02-080122	4	0.93	8.84	9.77	8.87	89.8	
N3991-02	CL-02-080122-EPH-2	5	0.93	8.84	9.77	8.87	89.8	
N3993-01	PLA-154	18	1.00	1.00	2.00	2.00	100.0	ASPHALT SAMPLE
N3993-02	PLA-155	19	1.00	1.00	2.00	2.00	100.0	ASPHALT SAMPLE
N3993-03	PLA-156	20	1.00	1.00	2.00	2.00	100.0	ASPHALT SAMPLE
N3993-04	PLA-157	21	1.00	1.00	2.00	2.00	100.0	ASPHALT SAMPLE
N3995-01	COMP-1	6	0.9	8.85	9.75	7.71	76.9	
N3995-02	COMP-2	7	0.98	8.83	9.81	8.46	84.7	
N3995-03	COMP-3	8	0.96	8.83	9.79	7.24	71.1	
N3996-01	MH-CS	22	0.9	8.80	9.7	8.44	85.7	
N3996-02	MH-CS-EPH-2	23	0.9	8.80	9.7	8.44	85.7	
N3996-04	MH-CD	24	0.93	9.01	9.94	8.54	84.5	
N3996-05	MH-CD-EPH-2	25	0.93	9.01	9.94	8.54	84.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$