



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB121316

Analyst By : JIGNESH

Supervisor Review By : mohammad

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	08/01/2022	16:20
2	CAL2	1	Water	NA	NA	20.8	7.00	08/01/2022	16:21
3	CAL3	1	Water	NA	NA	20.8	10.03	08/01/2022	16:22
4	ICV	1	Water	NA	NA	20.7	6.98	08/01/2022	16:25
5	CCV1	1	Water	NA	NA	20.7	2.01	08/01/2022	16:30
6	N3988-05	1	Water	NA	NA	23.6	7.02	08/01/2022	16:33
7	N3994-01	1	Water	NA	NA	21.0	6.21	08/01/2022	16:35
8	N3994-01DUP	1	Water	NA	NA	21.1	6.22	08/01/2022	16:36
9	CCV2	1	Water	NA	NA	20.7	12.03	08/01/2022	16:37

VB 121316

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph n3994

WorkList ID : 161897

Department : Wet-Chemistry

Date : 08-01-2022 16:27:04

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/01/2022	Water	N3988-05	pH	Cool 4 deg C	PSEG03	G11	63022-A	08/01/2022	9040C
07/29/2022	Water	N3994-01	pH	Cool 4 deg C	STER01	G11	POOL-WATER	07/29/2022	9040C

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by: