



PERCENT SOLID

Supervisor: mohammad

Analyst: jignesh

Date: 8/3/2022

OVENTEMP IN Celsius(°C): 107

Time IN: 17:10

In Date: 08/02/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:25

Out Date: 08/03/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB121323

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N3966-07	S3-03-S	1	0.91	8.84	9.75	9.11	92.8	
N3968-12	S2-02-S2	2	0.92	8.60	9.52	9.13	95.5	
N3968-13	S2-02-B2	3	0.93	8.82	9.75	9.27	94.6	
N3997-11	COMP1-0802	15	1.00	1.00	2.00	2.00	100.0	debris
N3997-12	COMP2-0802	16	0.91	8.64	9.55	8.39	86.6	
N3997-13	COMP3-0802	17	0.92	9.01	9.93	7.65	74.7	
N3997-14	1987	18	1.00	1.00	2.00	2.00	100.0	debris
N3998-01	SB-1	4	0.93	8.59	9.52	9.35	98.0	
N3998-02	SB-2	5	0.93	8.68	9.61	9.43	97.9	
N3998-03	SB-3	6	0.88	8.65	9.53	9.15	95.6	
N3998-04	SB-4	7	0.91	8.84	9.75	9.44	96.5	
N3998-05	SB-5	8	0.88	8.78	9.66	8.92	91.6	
N3998-06	SB-6	9	0.91	9.04	9.95	8.78	87.1	
N3998-07	SB-7	10	0.92	9.02	9.94	7.96	78.0	
N3998-08	SS-1	11	0.95	9.02	9.97	8.34	81.9	
N3998-09	SS-2	12	0.94	9.00	9.94	8.87	88.1	
N3998-10	ST-1	13	0.8	9.02	9.82	8.22	82.3	
N3998-11	DW-1	14	0.91	8.72	9.63	8.48	86.8	
N4009-01	TR-05-080222	19	0.95	8.79	9.74	9.03	91.9	
N4009-02	TR-05-080222-EPH-2	20	0.91	8.98	9.89	8.93	89.3	
N4015-03	MDL-SOIL-QT3-2022-05	21	1.00	1.00	2.00	2.00	100.0	
N4015-04	MDL-WATER-QT3-2022-06	22	1.00	1.00	2.00	2.00	100.0	
N4015-05	MDL-MED-SOIL-QT3-2022-05	23	1.00	1.00	2.00	2.00	100.0	
N4015-06	MDL-MED-SOIL-QT3-2022-06	24	1.00	1.00	2.00	2.00	100.0	
N4016-01	MH-DS	25	0.85	8.75	9.6	8.45	86.9	
N4016-02	MH-DS-EPH-2	26	0.85	8.75	9.6	8.45	86.9	
N4016-04	MH-ES	27	0.93	9.00	9.93	8.9	88.6	



PERCENT SOLID

Supervisor: mohammad

Analyst: jignesh

Date: 8/3/2022

OVENTEMP IN Celsius(°C): 107

Time IN: 17:10

In Date: 08/02/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:25

Out Date: 08/03/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB121323

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N4016-05	MH-ES-EPH-2	28	0.93	9.00	9.93	8.9	88.6	
N4016-07	MH-ED	29	0.98	8.65	9.63	8.74	89.7	
N4016-08	MH-ED-EPH-2	30	0.98	8.65	9.63	8.74	89.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$