

**PERCENT SOLID**

Supervisor: ketankumar

Analyst: jignesh

Date: 8/4/2022

OVENTEMP IN Celsius(°C): 107

Time IN: 17:00

In Date: 08/03/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 108

Time OUT: 08:00

Out Date: 08/04/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB121333

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N4012-01	WASTE	1	0.9	8.82	9.72	8.18	82.5	
N4013-01	SB-1	2	0.91	9.01	9.92	8.92	88.9	
N4013-03	SB-2	3	0.94	9.03	9.97	9.59	95.8	
N4013-05	SB-3	4	0.95	8.67	9.62	8.99	92.7	
N4013-07	SB-4	5	0.93	8.64	9.57	8.58	88.5	
N4013-09	SB-5	6	0.9	8.92	9.82	8.85	89.1	
N4013-11	SB-6	7	0.91	8.74	9.65	9.02	92.8	
N4013-13	SB-7	8	0.91	8.83	9.74	8.13	81.8	
N4013-15	SB-8	9	0.92	9.04	9.96	9.19	91.5	
N4013-17	SB-9	10	0.9	8.80	9.7	8.87	90.6	
N4014-01	SB-10	11	0.88	8.67	9.55	8.91	92.6	
N4014-03	SB-11	12	0.91	8.80	9.71	8.94	91.2	
N4014-05	SB-12	13	0.91	8.62	9.53	8.45	87.5	
N4014-07	SB-13	14	0.96	9.01	9.97	9.42	93.9	
N4014-09	SB-14	15	0.89	8.88	9.77	7.99	80.0	
N4014-11	SB-15	16	0.91	9.03	9.94	8.85	87.9	
N4014-13	SB-16	17	0.9	9.07	9.97	8.88	88.0	
N4017-01	SB08	18	0.92	8.82	9.74	8.44	85.3	
N4017-02	SB09	19	0.91	8.83	9.74	6.84	67.2	
N4017-03	SB10	20	0.9	9.01	9.91	8.74	87.0	
N4017-04	SB11	21	0.92	8.93	9.85	8.34	83.1	
N4017-05	SB12	22	0.92	8.70	9.62	8.52	87.4	
N4017-06	SB07	23	0.91	8.81	9.72	8.29	83.8	
N4017-07	SB06	24	0.89	8.98	9.87	9.46	95.4	
N4017-08	SB05	25	0.91	9.02	9.93	9.29	92.9	
N4017-09	SB03	26	0.91	8.71	9.62	9.22	95.4	
N4017-10	SB04	27	0.91	8.72	9.63	9.27	95.9	
N4023-01	TP1-1-6	28	0.91	8.84	9.75	9.2	93.8	



PERCENT SOLID

Supervisor: ketankumar

Analyst: jignesh

Date: 8/4/2022

OVENTEMP IN Celsius(°C): 107

Time IN: 17:00

In Date: 08/03/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 108

Time OUT: 08:00

Out Date: 08/04/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB121333

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N4023-03	TP1-1-6-EPH-2	29	0.91	8.84	9.75	9.2	93.8	
N4023-04	TP2-1-6	30	0.93	8.76	9.69	9.01	92.2	
N4023-06	TP2-1-6-EPH-2	31	0.93	8.76	9.69	9.01	92.2	
N4023-07	TP3-1-6	32	0.91	8.70	9.61	8.87	91.5	
N4023-09	TP3-1-6-EPH-2	33	0.91	8.70	9.61	8.87	91.5	
N4038-01	S-1	34	0.9	9.08	9.98	9.41	93.7	
N4038-02	S-2	35	0.9	8.84	9.74	9.18	93.7	
N4038-03	S-3	36	0.89	9.05	9.94	9.25	92.4	
N4038-04	S-4	37	0.91	8.90	9.81	9.14	92.5	
N4038-05	S-5	38	0.93	9.05	9.98	9.62	96.0	
N4038-06	S-6	39	0.9	8.79	9.69	9.27	95.2	
N4038-07	S-7	40	0.91	9.07	9.98	9.72	97.1	
N4038-08	S-8	41	0.92	8.85	9.77	9.4	95.8	
N4038-09	S-9	42	0.91	9.07	9.98	8.7	85.9	
N4038-10	S-10	43	0.92	9.06	9.98	8.86	87.6	
N4038-11	S-11	44	0.9	8.87	9.77	8.32	83.7	
N4038-12	S-12	45	0.91	8.84	9.75	8.32	83.8	
N4038-13	S-13	46	0.9	9.09	9.99	8.8	86.9	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$