


LB121335

INSTRUMENT ID: 112

7470A

Linear



A= 0.0000e+000

B= 2.3581e-004

C= 4.6409e-002

Rho= 0.9997768

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	4/50
0.0	0.000	0.051	0.051	20	0.000	20					14
0.2	0.200	0.227	0.027	766	0.0 %	766					14
2.5	2.500	2.403	-0.097	9993	0.0 %	9993					14
5.0	5.000	4.932	-0.068	20720	0.0 %	20720					14
7.5	7.500	7.631	0.131	32164	0.0 %	32164					14
10.0	10.000	9.956	-0.044	42023	0.0 %	42023					14

LB121335 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 03 Aug 2022 10:18:36

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	20	-	0.0000	7470A	PPB	03 Aug 2022 10:30:04	S	Std
0.2 - 1	512	766	-	0.2000	7470A	PPB	03 Aug 2022 10:32:22	S	Std
2.5 - 1	52.5	9993	-	2.5000	7470A	PPB	03 Aug 2022 10:34:39	S	Std
5.0 - 1	55	20720	-	5.0000	7470A	PPB	03 Aug 2022 10:36:55	S	Std
7.5 - 1	57.5	32164	-	7.5000	7470A	PPB	03 Aug 2022 10:39:11	S	Std
10.0 - 1	510	42023	-	10.0000	7470A	PPB	03 Aug 2022 10:44:04	S	Std
ICV63 - 1	ICV63	15993	3.8177	-7470A	PPB	PPB	03 Aug 2022 10:47:03	U	SMPL
ICB63 - 1	ICB63	-70	0.0299	-7470A	PPB	PPB	03 Aug 2022 10:49:18	U	SMPL
CCV32 - 1	CCV32	20355	4.8463	-7470A	PPB	PPB	03 Aug 2022 10:51:33	U	SMPL
CCB32 - 1	CCB32	-64	0.0313	-7470A	PPB	PPB	03 Aug 2022 10:53:48	U	SMPL
CRA - 1	CRA	798	0.2346	-7470A	PPB	PPB	03 Aug 2022 10:56:04	U	SMPL
HighStd - 1	HighStd	38436	9.1099	-7470A	PPB	PPB	03 Aug 2022 11:00:55	U	SMPL
ChkStd - 1	ChkStd	28579	6.7856	-7470A	PPB	PPB	03 Aug 2022 11:03:12	U	SMPL
PB146687BL - 1	PBW	-204	-0.0017	-7470A	PPB	PPB	03 Aug 2022 11:05:30	U	SMPL
PB146687BS - 1	LCSW	15440	3.6873	-7470A	PPB	PPB	03 Aug 2022 11:10:23	U	SMPL
N3976-02 - 1	TP1-0729	-94	0.0242	-7470A	PPB	PPB	03 Aug 2022 11:12:39	U	SMPL
N3976-04 - 1	TP2-0729	-11	0.0438	-7470A	PPB	PPB	03 Aug 2022 11:14:56	U	SMPL
N3976-06 - 1	TP3-0729	70	0.0629	-7470A	PPB	PPB	03 Aug 2022 11:17:12	U	SMPL
N3988-01 - 1	63022-B	51	0.0584	-7470A	PPB	PPB	03 Aug 2022 11:19:30	U	SMPL
N3988-02 - 1	63022-C	18	0.0507	-7470A	PPB	PPB	03 Aug 2022 11:21:45	U	SMPL
CCV33 - 1	CCV33	20976	4.9927	-7470A	PPB	PPB	03 Aug 2022 11:24:00	U	SMPL
CCB33 - 1	CCB33	-31	0.0391	-7470A	PPB	PPB	03 Aug 2022 11:26:16	U	SMPL
N3988-03 - 1	63022-D	23	0.0518	-7470A	PPB	PPB	03 Aug 2022 11:28:31	U	SMPL
N3988-04 - 1	63022-E	37	0.0551	-7470A	PPB	PPB	03 Aug 2022 11:30:46	U	SMPL
N3996-03 - 1	MH-CS	37	0.0551	-7470A	PPB	PPB	03 Aug 2022 11:33:05	U	SMPL
N3996-06 - 1	MH-CD	23	0.0518	-7470A	PPB	PPB	03 Aug 2022 11:35:21	U	SMPL
N3996-06DUP - 1	MH-CDDUP	24	0.0521	-7470A	PPB	PPB	03 Aug 2022 11:37:37	U	SMPL
N3996-06MS - 1	MH-CDMS	14162	3.3859	-7470A	PPB	PPB	03 Aug 2022 11:39:53	U	SMPL
N3996-06MSD - 1	MH-CDMSD	14607	3.4908	-7470A	PPB	PPB	03 Aug 2022 11:45:03	U	SMPL
PB146692BL - 1	PBW	-32	0.0389	-7470A	PPB	PPB	03 Aug 2022 11:47:19	U	SMPL
PB146692BS - 1	LCSW	16971	4.0483	-7470A	PPB	PPB	03 Aug 2022 11:49:37	U	SMPL
N3977-01 - 1	MS-TOTE-0729	65	0.0617	-7470A	PPB	PPB	03 Aug 2022 11:51:52	U	SMPL
CCV34 - 1	CCV34	21616	5.1436	-7470A	PPB	PPB	03 Aug 2022 11:54:08	U	SMPL
CCB34 - 1	CCB34	-24	0.0407	-7470A	PPB	PPB	03 Aug 2022 11:56:23	U	SMPL
N3981-01 - 1	PAT-TOTES-0729	1996	0.5171	-7470A	PPB	PPB	03 Aug 2022 11:58:39	U	SMPL
N3981-01DUP - 1	PAT-TOTES-0729DUP	1835	0.4791	-7470A	PPB	PPB	03 Aug 2022 12:00:55	U	SMPL
N3981-01MS - 1	PAT-TOTES-0729MS	18724	4.4617	-7470A	PPB	PPB	03 Aug 2022 12:03:10	U	SMPL
N3981-01MSD - 1	PAT-TOTES-0729MSD	18514	4.4121	-7470A	PPB	PPB	03 Aug 2022 12:05:26	U	SMPL
N3996-06LX5 - 1		-43	0.0363	-7470A	PPB	PPB	03 Aug 2022 12:07:42	U	SMPL
N3996-06A - 1		16033	3.8271	-7470A	PPB	PPB	03 Aug 2022 12:09:58	U	SMPL
N3981-01LX5 - 1		95	0.0688	-7470A	PPB	PPB	03 Aug 2022 12:12:15	U	SMPL
N3981-01A - 1		16943	4.0417	-7470A	PPB	PPB	03 Aug 2022 12:14:33	U	SMPL
PB146633TB - 1	PB146633TB	-150	0.0110	-7470A	PPB	PPB	03 Aug 2022 12:16:50	U	SMPL
CCV35 - 1	CCV35	21586	5.1365	-7470A	PPB	PPB	03 Aug 2022 12:19:05	U	SMPL
CCB35 - 1	CCB35	-41	0.0367	-7470A	PPB	PPB	03 Aug 2022 12:21:21	U	SMPL