



Analytical Summary Report

Analysis Method: 1010B
Parameter: Flash Point
Run Number: LB121338
Thermometer ID: flashpoint

Reviewed By: ketankumar
Supervisor Review By: Iwona
Ambient Barometric Pressure (mmHg): 765.00
Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	28.00	82.4	82.1	08/03/2022	12:45
2	N3977-01		1	12	100.00	>212.0	>212.0	08/03/2022	13:05
3	N3977-01DUP		1	12	100.00	>212.0	>212.0	08/03/2022	13:35
4	N3981-01		1	15	100.00	>212.0	>212.0	08/03/2022	14:05
5	N3988-05		1	18	100.00	>212.0	>212.0	08/03/2022	14:32
6	N3997-15		1	17	100.00	>212.0	>212.0	08/03/2022	14:58

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 121338

WorkList Name : fp8322 WorkList ID : 161954 Department : Wet-Chemistry Date : 08-03-2022 12:17:41

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/05/2022	Water	N3977-01	Flash Point	Cool 4 deg C	PSEG03	G11	MS-TOTE-0729	07/29/2022	1010B
08/05/2022	Water	N3981-01	Flash Point	Cool 4 deg C	PSEG03	G11	PAT-TOTES-0729	07/29/2022	1010B
08/04/2022	Water	N3988-05	Flash Point	Cool 4 deg C	PSEG03	G11	63022-A	08/01/2022	1010B
08/05/2022	Water	N3997-15	Flash Point	Cool 4 deg C	PSEG03	J11	1986	08/01/2022	1010B

Date/Time 08/03/22 12:25
 Raw Sample Received by: KS(wc)
 Raw Sample Relinquished by: RS CECT-1666

Date/Time 08/03/22 15:05
 Raw Sample Received by: RS CECT-1666
 Raw Sample Relinquished by: KS(wc)