


LB121399

7471B

INSTRUMENT ID: CV2

 μ Abs.:

57707



A= 0.0000e+000

B= 1.7378e-004

C= -8.5645e-003

Rho= 0.9999860

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.002	-0.002	39	0.000	39					
0.2	0.200	0.195	-0.005	1174	0.0 %	1174					
2.5	2.500	2.530	0.030	14608	0.0 %	14608					
5.0	5.000	4.979	-0.021	28702	0.0 %	28702					
7.5	7.500	7.477	-0.023	43074	0.0 %	43074					
10.0	10.000	10.020	0.020	57707	0.0 %	57707					

-3
1
0
0
0

LB121339 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 03 Aug 2022 13:26:49					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	39	-	0.0000	7471B	PPB	03 Aug 2022 13:27:52	S	Std
0.2 - 1	50.2	1174	-	0.2000	7471B	PPB	03 Aug 2022 13:30:09	S	Std
2.5 - 1	52.5	14608	-	2.5000	7471B	PPB	03 Aug 2022 13:32:26	S	Std
5.0 - 1	55	28702	-	5.0000	7471B	PPB	03 Aug 2022 13:34:43	S	Std
7.5 - 1	57.5	43074	-	7.5000	7471B	PPB	03 Aug 2022 13:36:59	S	Std
10.0 - 1	510	57707	-	10.0000	7471B	PPB	03 Aug 2022 13:39:17	S	Std
ICV64 - 1	ICV64	23250	4.0319	-	7471B	PPB	03 Aug 2022 13:42:07	U	SMPL
ICB64 - 1	ICB64	-98	-0.0256	-	7471B	PPB	03 Aug 2022 13:44:22	U	SMPL
CCV36 - 1	CCV36	28554	4.9536	-	7471B	PPB	03 Aug 2022 13:46:38	U	SMPL
CCB36 - 1	CCB36	-61	-0.0192	-	7471B	PPB	03 Aug 2022 13:48:53	U	SMPL
CRA - 1	CRA	1128	0.1875	-	7471B	PPB	03 Aug 2022 13:51:10	U	SMPL
HighStd - 1	HighStd	53251	9.2456	-	7471B	PPB	03 Aug 2022 13:53:24	U	SMPL
ChkStd - 1	ChkStd	40135	6.9662	-	7471B	PPB	03 Aug 2022 13:55:40	U	SMPL
PB146684BL - 1	PBS	-125	-0.0303	-	7471B	PPB	03 Aug 2022 13:57:58	U	SMPL
PB146684BS - 1	LCSS	20589	3.5695	-	7471B	PPB	03 Aug 2022 14:00:15	U	SMPL
N3931-01 - 1	RVE-113	1008	0.1666	-	7471B	PPB	03 Aug 2022 14:02:31	U	SMPL
N3931-01DUP - 1	RVE-113DUP	1054	0.1746	-	7471B	PPB	03 Aug 2022 14:04:49	U	SMPL
N3931-01MS - 1	RVE-113MS	22861	3.9643	-	7471B	PPB	03 Aug 2022 14:07:06	U	SMPL
N3931-01MSD - 1	RVE-113MSD	22799	3.9535	-	7471B	PPB	03 Aug 2022 14:09:23	U	SMPL
N3931-02 - 1	RVE-114	3047	0.5210	-	7471B	PPB	03 Aug 2022 14:11:39	U	SMPL
CCV37 - 1	CCV37	28723	4.9830	-	7471B	PPB	03 Aug 2022 14:13:55	U	SMPL
CCB37 - 1	CCB37	-63	-0.0195	-	7471B	PPB	03 Aug 2022 14:16:10	U	SMPL
N3931-03 - 1	RVE-115	17551	3.0415	-	7471B	PPB	03 Aug 2022 14:18:27	U	SMPL
N3931-04 - 1	RVE-116	1596	0.2688	-	7471B	PPB	03 Aug 2022 14:20:42	U	SMPL
N3931-05 - 1	RVE-117	1738	0.2935	-	7471B	PPB	03 Aug 2022 14:22:58	U	SMPL
N3931-06 - 1	RVE-118	4600	0.7908	-	7471B	PPB	03 Aug 2022 14:25:14	U	SMPL
N3931-07 - 1	RVE-119	318732	HIGH	-	7471B	PPB	03 Aug 2022 14:27:30	X	
N3931-08 - 1	RVE-119D	338446	HIGH	-	7471B	PPB	03 Aug 2022 14:29:23	X	
N3931-09 - 1	RVE-120	847	0.1386	-	7471B	PPB	03 Aug 2022 14:31:19	U	SMPL
N3931-10 - 1	RVE-121	876	0.1437	-	7471B	PPB	03 Aug 2022 14:33:40	U	SMPL
N3931-11 - 1	RVE-122	2244	0.3814	-	7471B	PPB	03 Aug 2022 14:35:57	U	SMPL
N3931-12 - 1	RVE-123	1495	0.2512	-	7471B	PPB	03 Aug 2022 14:38:13	U	SMPL
CCV38 - 1	CCV38	29149	5.0570	-	7471B	PPB	03 Aug 2022 14:40:28	U	SMPL
CCB38 - 1	CCB38	-64	-0.0197	-	7471B	PPB	03 Aug 2022 14:42:44	U	SMPL
N3931-13 - 1	RVE-124	3316	0.5677	-	7471B	PPB	03 Aug 2022 14:45:00	U	SMPL
N3931-14 - 1	RVE-125	1771	0.2992	-	7471B	PPB	03 Aug 2022 14:47:16	U	SMPL
N3931-15 - 1	RVE-126	1227	0.2047	-	7471B	PPB	03 Aug 2022 14:49:31	U	SMPL
N3931-16 - 1	RVE-127	3687	0.6322	-	7471B	PPB	03 Aug 2022 14:51:47	U	SMPL
N3931-17 - 1	RVE-128	2502	0.4262	-	7471B	PPB	03 Aug 2022 14:54:02	U	SMPL
N3931-18 - 1	RVE-129	2673	0.4560	-	7471B	PPB	03 Aug 2022 14:56:19	U	SMPL
N3931-19 - 1	RVE-130	4433	0.7618	-	7471B	PPB	03 Aug 2022 14:58:36	U	SMPL
N3931-20 - 1	RVE-131	2699	0.4605	-	7471B	PPB	03 Aug 2022 15:00:52	U	SMPL
PB146685BL - 1	PBS	54	0.0008	-	7471B	PPB	03 Aug 2022 15:03:10	U	SMPL
PB146685BS - 1	LCSS	19560	3.3906	-	7471B	PPB	03 Aug 2022 15:05:25	U	SMPL
CCV39 - 1	CCV39	28423	4.9309	-	7471B	PPB	03 Aug 2022 15:07:41	U	SMPL
CCB39 - 1	CCB39	-63	-0.0195	-	7471B	PPB	03 Aug 2022 15:09:57	U	SMPL
N3935-01 - 1	RVE-132	4143	0.7114	-	7471B	PPB	03 Aug 2022 15:12:13	U	SMPL
N3935-01DUP - 1	RVE-132DUP	4205	0.7222	-	7471B	PPB	03 Aug 2022 15:14:29	U	SMPL
N3935-01MS - 1	RVE-132MS	26197	4.5440	-	7471B	PPB	03 Aug 2022 15:16:44	U	SMPL
N3935-01MSD - 1	RVE-132MSD	24954	4.3280	-	7471B	PPB	03 Aug 2022 15:19:00	U	SMPL
N3935-02 - 1	RVE-133	2929	0.5004	-	7471B	PPB	03 Aug 2022 15:21:16	U	SMPL
N3935-03 - 1	RVE-134	3236	0.5538	-	7471B	PPB	03 Aug 2022 15:23:34	U	SMPL
N3935-04 - 1	RVE-135	2117	0.3593	-	7471B	PPB	03 Aug 2022 15:25:51	U	SMPL
N3935-05 - 1	RVE-136	4317	0.7417	-	7471B	PPB	03 Aug 2022 15:28:08	U	SMPL
N3935-06 - 1	RVE-137	3574	0.6125	-	7471B	PPB	03 Aug 2022 15:30:25	U	SMPL
N3935-07 - 1	RVE-138	1826	0.3088	-	7471B	PPB	03 Aug 2022 15:32:41	U	SMPL
CCV40 - 1	CCV40	27232	4.7239	-	7471B	PPB	03 Aug 2022 15:34:57	U	SMPL
CCB40 - 1	CCB40	-59	-0.0188	-	7471B	PPB	03 Aug 2022 15:37:12	U	SMPL
N3935-08 - 1	RVE-139	2699	0.4605	-	7471B	PPB	03 Aug 2022 15:39:29	U	SMPL
N3935-09 - 1	RVE-139D	3670	0.6292	-	7471B	PPB	03 Aug 2022 15:41:45	U	SMPL
N3935-10 - 1	RVE-140	2166	0.3678	-	7471B	PPB	03 Aug 2022 15:44:01	U	SMPL
N3935-11 - 1	RVE-141	2054	0.3484	-	7471B	PPB	03 Aug 2022 15:46:16	U	SMPL
N3935-12 - 1	RVE-142	3080	0.5267	-	7471B	PPB	03 Aug 2022 15:48:32	U	SMPL
N3935-13 - 1	RVE-143	3643	0.6245	-	7471B	PPB	03 Aug 2022 15:50:48	U	SMPL
N3935-14 - 1	RVE-144	2130	0.3616	-	7471B	PPB	03 Aug 2022 15:53:05	U	SMPL
N3935-15 - 1	RVE-145	5483	0.9443	-	7471B	PPB	03 Aug 2022 15:55:21	U	SMPL
N3935-16 - 1	RVE-146	2108	0.3578	-	7471B	PPB	03 Aug 2022 15:57:38	U	SMPL
N3935-17 - 1	RVE-147	2810	0.4798	-	7471B	PPB	03 Aug 2022 15:59:54	U	SMPL
CCV41 - 1	CCV41	29350	5.0920	-	7471B	PPB	03 Aug 2022 16:02:10	U	SMPL
CCB41 - 1	CCB41	-201	-0.0435	-	7471B	PPB	03 Aug 2022 16:04:26	U	SMPL
N3935-18 - 1	RVE-148	3288	0.5628	-	7471B	PPB	03 Aug 2022 16:06:44	U	SMPL
PB146709BL - 1	PBS	5	-0.0077	-	7471B	PPB	03 Aug 2022 16:09:00	U	SMPL
PB146709BS - 1	LCSS	22567	3.9132	-	7471B	PPB	03 Aug 2022 16:11:16	U	SMPL
N3989-01 - 1	OK-04-080122	1148	0.1909	-	7471B	PPB	03 Aug 2022 16:13:32	U	SMPL
N3989-01DUP - 1	OK-04-080122DUP	1203	0.2005	-	7471B	PPB	03 Aug 2022 16:15:49	U	SMPL
N3989-01MS - 1	OK-04-080122MS	22694	3.9353	-	7471B	PPB	03 Aug 2022 16:18:05	U	SMPL
N3989-01MSD - 1	OK-04-080122MSD	22759	3.9466	-	7471B	PPB	03 Aug 2022 16:20:22	U	SMPL
N3990-01 - 1	HR-01-080122	7375	1.2731	-	7471B	PPB	03 Aug 2022 16:22:40	U	SMPL
N3990-03 - 1	HR-04-080122	6465	1.1149	-	7471B	PPB	03 Aug 2022 16:24:58	U	SMPL
N3991-01 - 1	CL-02-080122	4631	0.7962	-	7471B	PPB	03 Aug 2022 16:27:14	U	SMPL
CCV42 - 1	CCV42	28139	4.8815	-	7471B	PPB	03 Aug 2022 16:29:30	U	SMPL

LB121339 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 03 Aug 2022 13:26:49

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
CCB42 - 1	CCB42	-47	-0.0167	-7471B	PPB		03 Aug 2022 16:31:46	U	SMPL
N3995-01 - 1	COMP-1	29033	5.0369	-7471B	PPB		03 Aug 2022 16:34:04	U	SMPL
N3995-02 - 1	COMP-2	40043	6.9502	-7471B	PPB		03 Aug 2022 16:36:20	U	SMPL
N3995-03 - 1	COMP-3	97311	16.9024	-7471B	PPB		03 Aug 2022 16:38:37	U	SMPL
N3996-01 - 1	MH-CS	26320	4.5654	-7471B	PPB		03 Aug 2022 16:40:55	U	SMPL
N3996-04 - 1	MH-CD	320	0.0470	-7471B	PPB		03 Aug 2022 16:43:13	U	SMPL
N3998-07 - 1	SB-7	21	-0.0049	-7471B	PPB		03 Aug 2022 16:45:31	U	SMPL
N3998-08 - 1	SS-1	5176	0.8909	-7471B	PPB		03 Aug 2022 16:47:48	U	SMPL
N3998-09 - 1	SS-2	6527	1.1257	-7471B	PPB		03 Aug 2022 16:50:05	U	SMPL
N3998-10 - 1	ST-1	246	0.0342	-7471B	PPB		03 Aug 2022 16:52:21	U	SMPL
N3998-11 - 1	DW-1	128	0.0137	-7471B	PPB		03 Aug 2022 16:54:38	U	SMPL
CCV43 - 1	CCV43	27181	4.7150	-7471B	PPB		03 Aug 2022 16:56:54	U	SMPL
CCB43 - 1	CCB43	-46	-0.0166	-7471B	PPB		03 Aug 2022 16:59:10	U	SMPL
N4009-01 - 1	TR-05-080222	6324	1.0904	-7471B	PPB		03 Aug 2022 17:01:28	U	SMPL
N4016-01 - 1	MH-DS	4298	0.7384	-7471B	PPB		03 Aug 2022 17:03:44	U	SMPL
N4016-04 - 1	MH-ES	37024	6.4256	-7471B	PPB		03 Aug 2022 17:06:00	U	SMPL
N4016-07 - 1	MH-ED	3410	0.5840	-7471B	PPB		03 Aug 2022 17:08:16	U	SMPL
N3931-01LX5 - 1		140	0.0158	-7471B	PPB		03 Aug 2022 17:10:36	U	SMPL
N3931-01A - 1		20087	3.4822	-7471B	PPB		03 Aug 2022 17:12:52	U	SMPL
N3935-01LX5 - 1		602	0.0961	-7471B	PPB		03 Aug 2022 17:15:10	U	SMPL
N3935-01A - 1		26840	4.6558	-7471B	PPB		03 Aug 2022 17:17:29	U	SMPL
N3989-01LX5 - 1		199	0.0260	-7471B	PPB		03 Aug 2022 17:19:46	U	SMPL
N3989-01A - 1		23551	4.0842	-7471B	PPB		03 Aug 2022 17:22:04	U	SMPL
CCV44 - 1	CCV44	28474	4.9397	-7471B	PPB		03 Aug 2022 17:29:12	U	SMPL
CCB44 - 1	CCB44	-50	-0.0173	-7471B	PPB		03 Aug 2022 17:31:28	U	SMPL
N3931-07DLX10 - 1		48501	8.4201	-7471B	PPB		03 Aug 2022 17:33:47	U	SMPL
N3931-08DLX10 - 1		49242	8.5489	-7471B	PPB		03 Aug 2022 17:36:03	U	SMPL
N3995-03DLX5 - 1		20422	3.5404	-7471B	PPB		03 Aug 2022 17:38:22	U	SMPL
N3931-07DL2X25 - 1		17392	3.0139	-7471B	PPB		03 Aug 2022 17:44:39	U	SMPL
N3931-08DL2X25 - 1		17792	3.0834	-7471B	PPB		03 Aug 2022 17:49:37	U	SMPL
CCV45 - 1	CCV45	30955	5.3709	-7471B	PPB		03 Aug 2022 17:54:43	U	SMPL
CCB45 - 1	CCB45	-290	-0.0590	-7471B	PPB		03 Aug 2022 17:57:00	U	SMPL